# **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



## Locations covered by the organisation and their relevant activities

## Laboratory locations:

Location details		Activity	Location code
Address 5 Cecil Pashley Way Shoreham Airport Shoreham-by-Sea West Sussex BN43 5FF	Local contact Dr Mark Hindle	Pressure Electrical Temperature indicators - Electrical simulation Air velocity Air flow Water flow Hydrocarbon flow Temperature Humidity	Lab

## Site activities performed away from the locations listed above:

Location details		Activity	Location code
The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Local contact Dr Mark Hindle	Pressure Electrical Temperature indicators - Electrical simulation Temperature Humidity	Site

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
OG04 Accredited to ISO/IEC 17025:2017	Young Calibration Limited Issue No: 044 Issue date: 01 May 2024
	Calibration performed by the Organisation at the locations specified
	Calibration and Measurement Capability (CMC)

Calibration and Measurement Capability (CMC)				
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
FLOW Hydrocarbon oils Calibration of flow meters using gravimetric and reference meter methods Volume flow rate Quantity of fluid passed	0.5 l/min to 440 l/min 0.33 l to 361.7 l 361.7 l to 4 400 l (at flow rates of 0.5 l/min to 440	0.40 % 0.40 % 1.0 %	Calibrations are carried out with fluids within the viscosity range 5 to 20cSt at fluid temperatures of up to 60 °C	Lab
Mass flow rate Mass of fluid passed	0.5 kg/min to 367.4 kg/min 0.4 kg to 302 kg 302 kg to 3 674 kg (at flow rates of 0.5 kg/min to 367.4 kg/min)	0.40 % 0.40 % 1.0 %	Calibrations are carried out at pressures of up to 10 bar	
Water				
Calibration of flow meters using gravimetric and reference meter methods				
Volume flow rate	1.0 ml/hr to 2400 ml/hr 1.0 ml/hr to 2400 ml/hr 0.04 l/min to 1 000 l/min	Q[0.10 %,0.0040 ml/hr] Q[0.55 %, 0.36 ml/hr] 0.15 %	Gravimetric Reference meter Gravimetric	Lab
Quantity of fluid passed	0.25 I to 801.6 I 801.6 I to 7 000 I (at flow rates of 0.04 I/min to 1 000 I/min)	0.15 % 1.0 %	Using both flying start/stop and standing start/stop methods	
Mass flow rate	1 g/hr to 2400 g/hr 1 g/hr to 2400 g/hr	Q[0.10 %, 0.0040 g/hr] Q[0.55 %, 0.36 g/hr]	Gravimetric Reference meter	
Mass of fluid passed	0.04 kg/min to 998 kg/min 0.25 g to 800 kg 800 kg to 9 980 kg (at flow rates of 0.04 kg/min to 998 kg/min)	0.15 % 0.15% 1.0 %	Gravimetric Gravimetric Reference meter	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
0604	Issue	No: 044 Issue date	: 01 May 2024	
Accredited to ISO/IEC 17025:2017				
	Calibration performed by the Org	anisation at the locations	specified	
		Expanded		
Measured Quantity Instrument or Gauge	Range	Measurement Uncertainty $(k = 2)$	Remarks	Location Code
Gas			Calibration of flow meters with an electrical or pressure output can be undertaken	
Calibration of flow meters using Sonic nozzle method				
Volume flow rate	1 ml/min to 50 ml/min 50 ml/min to 2500 l/min	Q[0.49 %,0.059 ml/min] 0.49 %	Calibrations are carried out at pressures of up to 8 bar	Lab
Quantity of gas passed	300 ml to 50000 l (at flow rates of 1 ml/min to 50 ml/min 50 ml/min to 2600 l/min)	Q[0.50 %,0.059 ml/min] 0.50 %		
Mass flow rate	0.001 2 g/min to 0.0060 g/min 0.0060 g/min to 3.02 kg/min	Q[0.48 %,0.071 mg/min] 0.48 %		
Mass of gas passed	0.36 g to 60.4 kg (at flow rates of 0.001 2 g/min to 0.0060 g/min 0.0060 g/min to 3.02 kg/min)	Q[0.50 %,0.071 mg/min] 0.50 %		
Calibration of flow meters using Turbine meter method				
Volume flow rate	10 l/s to 450 l/s	0.90 %	Calibrations are carried out at ambient conditions	
Quantity of gas passed	300 l to 540 kl (at flow rates of 10 l/s to 450 l/s)	0.90 %		
Mass flow rate	12 g/s to 0.54 kg/s	0.90 %		
Mass of gas passed	0.36 kg to 643 kg (at flow rates of 12 g/s to 0.54 kg/s)	0.90 %		
Calibration of flow meters using LDA method				
Volume flow rate	40 l/s to 1 250 l/s	Q[0.70 %, 0.030 l/s]	Calibrations are carried out	Lab
Quantity of gas passed	12 kl to 375 kl (at flow rates of 40 l/s to 1 250 l/s)	Q[0.70 %, 9.0 l]	a laser doppler anemometer	

Q[0.70 %, 0.036 g/s]

Q[0.70 %, 11 g]

48 g/s to 1.488 kg/s

14.28 kg to 1 785 kg (at flow rates of 48 g/s to 1.488 kg/s)

	United 2 Pine Trees, Cf	Schedule of Accr issued by Kingdom Accrec nertsey Lane, Staines-u	editation ditation Service	R, UK	
UKAS CALIBRATION 0604 Accredited to ISO/IEC 17025:2017	Is	Young Calibration ssue No: 044 Issue date	Limited e: 01 May 2024		
Calibration performed by the Organisation at the locations specified					
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
AIR VELOCITY					
Pitot tubes Thermal and ultrasonic anemometers	1.0 m/s to 80 m/s 0.05 m/s to 80 m/s	0.15 % + 0.000 20 m/s 0.16 % + 0.000 30 m/s	1) Calibrations are performed against a laser doppler anemometer, or a secondary standard where	Lab	

Thermal and ultrasonic anemometers	0.05 m/s to 80 m/s	0.16 % + 0.000 30 m/s	doppler anemometer, or a secondary standard where requested	
Vane anemometers	0.1 m/s to 40 m/s	0.20 % + 0.0010 m/s	2) Pitot tube uncertainty dependent on pitot	
Rotating cup anemometers	1.0 m/s to 21 m/s	0.23 %	differential pressure range 3) Air velocity instruments up to 480 x 120 mm diameter (working area): uncertainty is dependent on design of instrument under test.	
PRESSURE			Methods consistent with EURAMET CG17	
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	5 kPa to 130 kPa	22 Pa	Devices with an electrical output can also be calibrated.	Lab
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges	-90 kPa to 0 Pa 0 Pa to 2.5 kPa 2.5 kPa to 5 kPa 5 kPa to 6.9 kPa 6.9 kPa to 20 kPa 20 kPa to 270 kPa 270 kPa to 2.1 MPa 2.1 MPa to 3.5 MPa	38 Pa 0.045 % + 0.060 Pa 0.046 % + 0.31 Pa 9 Pa 16 Pa 0.029 % 0.023 % 1.1 kPa 0.029 % + 0.84 kPa	Absolute pressure calibrations can be undertaken using gauge pressure generation and the associated barometric pressure with the additional absolute pressure uncertainty as listed Oxygen-duty pressure	Lab Lab & Site
	270 kPa to 1.6 MPa	0.023 % + 0.90 kPa	instruments	
Hydraulic pressure (gauge)				
Calibration of pressure indicating instruments and	0.55 MPa to 55 MPa 55 MPa to 110 MPa	0.019 % + 90 Pa 0.018 %		Lab
gauge	0 kPa to 5.5 MPa 5.5 MPa to 40 MPa 40 MPa to 70 MPa	0.023 % + 11 kPa 0.018 % + 12 kPa 72 kPa	Oxygen-duty pressure instruments	Lab & Site
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	5.0 kPa to 130 kPa	22 Pa		Site

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
CALIBRATION	Young Calibration Limited				
0604 Accredited to	Issue No: 044 Issue date: 01 May 2024				
100/120 17023.2017	Calibration performed by the	Organisation at the locations	specified		
		Expanded			
Measured Quantity Instrument or Gauge	Range	Measurement Uncertainty $(k = 2)$	Remarks	Location Code	
PRESSURE (continued)					
Gas pressure (gauge)					
Calibration of pressure indicating instruments and gauges	-90 kPa to 0 Pa 0 Pa to 2.5 kPa 2.5 kPa to 5.0 kPa 5.0 kPa to 6.9 kPa 6.9 kPa to 34.5 kPa 34.5 kPa to 250 kPa 2.1 MPa to 3.5 Mpa	38 Pa 0.045 % + 0.060 Pa 0.046 % + 0.31 Pa 9 Pa 16 Pa 82 Pa 11 kPa		Site	
ELECTRICAL	2.1 101 a to 5.5 mpa	1.1 K a		Lab and	
DC Voltage Generation	0 V to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 1000 V	0.012 % + 6.0 µV 0.0070 % + 33 µV 0.0070 % + 370 µV 0.0073 % + 29 mV	Using multifunction calibrator.	5110	
Measurement	0 V to 100 mV 100 mV to1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	$\begin{array}{c} 0.0047 \ \% + 5.5 \ \mu V \\ 0.0034 \ \% + 34 \ \mu V \\ 0.0033 \ \% + 330 \ \mu V \\ 0.0048 \ \% + 3.3 \ m V \\ 0.0051 \ \% + 31 \ m V \end{array}$	Using digital multimeter.		
Mains Voltage Measurement Mains Voltage Frequency	200 V to 300 V Nominal 50 Hz	1.0 % 2.5 %			
DC Current Generation	0 mA to 33 mA 33 mA to 330 mA 330 mA to 2.2 A 2.2 A to 11 A	0.016 % + 0.35 μA 0.017 % + 4.5 μA 0.041 % + 190 μA 0.079 % + 460 μA	Using multifunction calibrator.		
Measurement	0 mA to 30 mA 0 mA to 100 mA 0 mA to 10A	0.017 % + 6.0 μA 0.060 % + 7.5 μA 0.21 % + 1.2 mA	Using digital multimeter.		
DC Resistance	$\begin{array}{c} 0 \ \Omega \ \text{to} \ 33 \ \Omega \\ 33 \ \Omega \ \text{to} \ 330 \ \Omega \\ 330 \ \Omega \ \text{to} \ 1.1 \ \text{k}\Omega \\ 1.1 \ \text{k}\Omega \ \text{to} \ 3.3 \ \text{k}\Omega \\ 1.1 \ \text{k}\Omega \ \text{to} \ 3.3 \ \text{k}\Omega \\ 3.3 \ \text{k}\Omega \ \text{to} \ 11 \ \text{k}\Omega \\ 33 \ \text{k}\Omega \ \text{to} \ 110 \ \text{k}\Omega \\ 110 \ \text{k}\Omega \ \text{to} \ 330 \ \text{k}\Omega \\ 330 \ \text{k}\Omega \ \text{to} \ 1.1 \ \text{M}\Omega \\ 1.1 \ \text{M}\Omega \ \text{to} \ 3.3 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 1.1 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 110 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 100 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 100 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 100 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 100 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 100 \ \text{M}\Omega \\ 3.3 \ \text{M}\Omega \ \text{to} \ 100 \ \text{M}\Omega \\ 0 \ \text{matrix} \ \text{matrix}$	$\begin{array}{c} 25 \ m\Omega \\ 60 \ m\Omega \\ 210 \ m\Omega \\ 500 \ m\Omega \\ 1.6 \ \Omega \\ 5.0 \ \Omega \\ 25 \ \Omega \\ 60 \ \Omega \\ 200 \ \Omega \\ 800 \ \Omega \\ 10 \ k\Omega \\ 43 \ k\Omega \\ 800 \ k\Omega \end{array}$	Using multifunction calibrator.		
Measurement	0 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 10 MΩ 100 MΩ to 10 Ω	20 mΩ 150 mΩ 1.5 Ω 15 Ω 140 Ω 5.0 kΩ 1.1 % + 23 kΩ 3.2 % + 7.4 MΩ	Using digital multimeter.		

(E)

	Sch United Ki 2 Pine Trees, Cherts	nedule of Accr issued by ngdom Accred sey Lane, Staines-up	editation litation Service pon-Thames, TW18 3H	R, UK	
CALIBRATION 0604 Accredited to ISO/IEC 17025:2017	Young Calibration Limited Issue No: 044 Issue date: 01 May 2024				
	Calibration performed by the Org	anisation at the locations	specified		
	1	Expanded		1	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code	

Instrument or Gauge		Uncertainty $(k = 2)$		Code
ELECTRICAL (continued) Frequency			Using GPS disciplined oscillator and frequency	Lab and site
Specific values	1 MHz, 5 MHz and 10 MHz	0.12 parts in 10 <sup>6</sup>	counter.	
Ranges source and measure	0.5 Hz to 250 kHz	3.7 parts in 10 <sup>6</sup>	Frequency can be	
Source only	250 MHz to 1 GHz	0.13 parts in 10 <sup>6</sup>	for example RPM, at equivalent uncertainties.	
Elapsed time	10 s to 24 hours	0.13 parts in 10 <sup>6</sup> + 54 ms	For calibration of timers and stopwatches.	
Electrical calibration of temperature	e indicators			
Ambient	15 °C to 30 °C	0.41 °C	For reporting reference junction temperature.	
Base Metal Thermocouples	-270 °C to 0 °C 0 °C to 1370 °C	0.67 °C 0.55 °C	Including reference junction compensation	
Noble Metal Thermocouples	-50 °C to +399 °C 400 °C to 1760 °C	0.96 °C 0.70 °C		
PRTs Generate resistance (Pt 100)	-200 °C to +850 °C	0.43 °C		
Measure resistance (Pt 100)	-200 °C to +850 °C	0.27 °C		
TEMPERATURE			By comparison with Reference thermometers	
Resistance thermometers and electronic probes with indicators	-40 °C to +20 °C 20 °C to 200 °C 0 °C	0.35 °C 0.059 °C 0.043 °C	In dry block calibrator In liquid bath In dry block calibrator	Lab
Base metal thermocouples	-40 °C to +20 °C 20 °C to 200 °C	0.55 °C 0.70 °C	In dry block calibrator In liquid bath	
Block Calibrators	-40 °C to +200 °C	0.045 °C		Lab and site
Liquid baths	-40 °C to +200 °C	0.045 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	
Air temperature	0 °C to 60 °C 0 °C to 60 °C	0.19 °C 0.21 °C	In air chamber In air chamber	Lab Site

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Young Calibration Limited Issue No: 044 Issue date: 01 May 2024			
CALIBRATION 0604 Accredited to ISO/IEC 17025:2017				
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
HUMIDITY			By comparison with chilled mirror hygrometer and reference thermometers	
Dew point	-20 °C to +60 °C	0.31 °C		Lab
Relative humidity	Relative Humidity derived from dew point and temperature		Achievable ranges:- 0 °C to 10 °C 20 %rh to 95 %rh	Lab

Site

END

0.22 °C

0.20 %rh

0.31 %rh 0.70 %rh

1.4 %rh 1.9 %rh

0.22 °C

1.3 %rh

1.4 %rh 1.7 %rh 2.2 %rh

2.5 %rh

10 °C to 18 °C

18 °C to 60 °C

5 %rh to 95 %rh

By comparison with reference hygrometer

10 %rh to 95 %rh

Temperature range 0 °C to 60 °C 5 %rh to 20 %rh

20 %rh to 40 %rh

40 %rh to 60 %rh

60 %rh to 80 %rh 80 %rh to 95 %rh

Temperature range 0 °C to 60 °C

5 %rh to 20 %rh

20 %rh to 40 %rh 40 %rh to 60 %rh

60 %rh to 80 %rh 80 %rh to 95 %rh



## Schedule of Accreditation issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

## Young Calibration Limited

Issue No: 044 Issue date: 01 May 2024

Accredited to ISO/IEC 17025:2017

Calibration performed by the Organisation at the locations specified

#### Appendix - Calibration and Measurement Capabilities

#### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### **Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where *q* is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$