

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0632

Accredited to
ISO/IEC 17025:2017

Gedore Torque Limited

Issue No: 026 Issue date: 05 December 2024

Gosden Common
Bramley
Guildford
Surrey
GU5 0AJ

Contact: Mr M Twumasi
Tel: +44 (0)1483-892772
Fax: +44 (0)1483-898536
E-Mail: salesandrepairs@gedore-torque.com
Website: www.gedore-torque.com

Calibration performed at the above address only

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k=2)	Remarks
TORQUE Torque Wrenches Torque Drivers Static Torque Transducers in clockwise and/or anti-clockwise direction in increasing Torque only	0.14 N·m to 2000 N·m to BS EN ISO 6789-2:2017 0.1 N·m to 3000 N·m to BS 7882:2017	1.0 % of reading See Notes 1 to 3 0.25 % of reading See Notes 1 to 3	NOTES 1 The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated. 2 Calibration results may also be given in other units of torque, such as lbf·in and lbf·ft. 3 Calibrated statically using un-supported beam and masses or torque transducer.
END			



0632
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Gedore Torque Limited
Issue No: 026 **Issue date:** 05 December 2024

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$