

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|---|
|  0636 Accredited to ISO/IEC 17025:2017 | Hilton Instruments Limited | |
| | Issue No: 013 Issue date: 20 August 2021 | |
| | Howe Moss Avenue Kirkhill Industrial Estate Dyce Aberdeen AB21 0GP | Contact: Mr S Bryce Tel: +44 (0)1224 729990 E-Mail: service@hiltoninstruments.com Website: www.hiltoninstruments.com |
| Calibration performed by the Organisations at the locations specified below | | |

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|--|---|---------------|
| Customer Premises*, e.g. Hospitals, Laboratories, and Manufacturing Plants etc (* and those of TSSS) | Temperature (Sterilizers etc, and Thermal Products) Time Rotational Speed | S |



0636
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Hilton Instruments Limited
Issue No: 013 **Issue date:** 20 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---------------------------------------|--|--|------------------|
| TEMPERATURE Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, fridges/refrigerators, freezers, environmental cabinets & enclosures, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges) | - 50 °C to 150 °C 150 °C to 200 °C | 0.60 °C 0.70 °C | Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping | S |
| TIME INTERVAL Timers | 60 s to 36000 s | 2.0 s | Calibration with a reference timer | S |
| ROTATIONAL SPEED Centrifuges Orbital shakers | 60 rpm to 30000 rpm | 6 rpm | Calibration with a reference tachometer | S |
| END | | | | |



0636
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Hilton Instruments Limited
Issue No: 013 **Issue date:** 20 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$