

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0654</b> Accredited to ISO/IEC 17025:2017	<b>Calibration, Maintenance and Repair Ltd</b>	
	Issue No: 039    Issue date: 02 September 2021	
	11 Frensham Road Norwich NR3 2BT	Contact: Mr P K Clark Tel: +44 (0)1603 279557 Fax: +44 (0)1603 278008 E-Mail: support@cmrcalibrate.co.uk Website: www.cmrcalibrate.co.uk

**Calibration performed by the Organisation at the locations specified**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 11 Frensham Road Norwich NR3 2BT	<b>Local contact:</b> Mr P K Clark Tel: +44 (0)1603 279557	Electrical Calibration Accelerometer calibration  A
<b>Address</b> 5 Octavian Way, Team Valley Trading Est Gateshead NE11 0HZ	<b>Local contact:</b> Mr John Fryer Tel: +44 (0)191 4875951	Electrical Calibration Pressure Calibration  B



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
<b>ELECTRICAL MEASUREMENTS</b>			Calibrations are performed as a direct comparison against a reference standard	
<b>DC VOLTAGE</b>				A
Generation	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1100 V	0.55 $\mu$ V 1.4 $\mu$ V 6.6 $\mu$ V 44 $\mu$ V 630 $\mu$ V 8.0 mV	For the calibration of measuring instruments	
Measurement	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV	1.0 $\mu$ V 2.0 $\mu$ V 10 $\mu$ V 98 $\mu$ V 1.3 mV 14 mV	For instruments with a voltage output	
<b>DC RESISTANCE</b>				A
Sourcing - Spot Values	1 $\Omega$  1.9 $\Omega$ 10 $\Omega$ 19 $\Omega$ 100 $\Omega$  190 $\Omega$ 1 k $\Omega$ 1.9 k $\Omega$ 10 k $\Omega$ 19 k $\Omega$  100 k $\Omega$ 190 k $\Omega$ 1 M $\Omega$ 1.9 M $\Omega$  10 M $\Omega$ 19 M $\Omega$ 100 M $\Omega$	110 $\mu$ $\Omega$  210 $\mu$ $\Omega$ 270 $\mu$ $\Omega$ 540 $\mu$ $\Omega$ 1.2 m $\Omega$  2.5 m $\Omega$ 8.7 m $\Omega$ 17 m $\Omega$ 81 m $\Omega$ 170 m $\Omega$  1.1 $\Omega$ 2.4 $\Omega$ 16 $\Omega$ 48 $\Omega$  460 $\Omega$ 1.1 k $\Omega$ 12 k $\Omega$	For the calibration of measuring instruments	



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
<b>DC RESISTANCE</b> (continued)				A
Sourcing - Range values	0 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$	2.0 m $\Omega$ 5.0 m $\Omega$ 31 m $\Omega$	For the calibration of measuring instruments	A
	1 k $\Omega$ to 10 k $\Omega$ 10 k $\Omega$ to 100 k $\Omega$	310 m $\Omega$ 3.0 $\Omega$		
	100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 33 M $\Omega$	37 $\Omega$ 2.0 k $\Omega$ 11 k $\Omega$		
	33 M $\Omega$ to 110 M $\Omega$ 110 M $\Omega$ to 330 M $\Omega$ 330 M $\Omega$ to 1.1 G $\Omega$	60 k $\Omega$ 1.0 M $\Omega$ 17 M $\Omega$		
Measurement	0 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 1 k $\Omega$ to 10 k $\Omega$ 10 k $\Omega$ to 100 k $\Omega$	250 $\mu\Omega$ 2.0 m $\Omega$ 14 m $\Omega$ 140 m $\Omega$ 1.0 $\Omega$	For instruments with a resistance output	
	100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 G $\Omega$ 10 G $\Omega$ to 1 T $\Omega$	22 $\Omega$ 0.71 % 1.0 %		
<b>DC CURRENT</b>				A
Generation	0 $\mu$ A to 10 $\mu$ A 10 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA	7.4 nA 12 nA 49 nA 450 nA 6.0 $\mu$ A	For the calibration of measuring instruments	
	100 mA to 1 A 1 A to 3 A 3 A to 5 A 5 A to 10 A 10 A to 20 A	110 $\mu$ A 1.0 mA 4.0 mA 6.0 mA 12 mA		
	16 A to 160 A 160 A to 1000 A	25 mA 580 mA	Current clamp calibration using a multi turn coil	



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
--	-------	--	---------	------------------

Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.

<b>DC CURRENT</b> (continued)				A
Measurement	0 $\mu$ A to 10 $\mu$ A 10 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1 mA to 10 mA	2.0 nA 4.0 nA 34 nA 340 nA	For instruments which generate current	
	10 mA to 100 mA 100 mA to 1 A 1 A to 2 A 2 A to 10 A	5.0 $\mu$ A 150 $\mu$ A 2.4 mA 25 mA		
<b>AC VOLTAGE</b>				A
Generation	1 Hz to 10 Hz 10 $\mu$ V to 10 mV	60 $\mu$ V	For the calibration of measuring instruments	
	10 mV to 100 mV 100 mV to 1 V 1 V to 7 V	270 $\mu$ V 1.9 mV 9.2 mV		
	10 $\mu$ V to 2.2 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	5.3 $\mu$ V 4.9 $\mu$ V 4.9 $\mu$ V 4.9 $\mu$ V 4.9 $\mu$ V 4.9 $\mu$ V 5.2 $\mu$ V 7.1 $\mu$ V 14 $\mu$ V 27 $\mu$ V 30 $\mu$ V		
	2.2 mV to 22 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	11 $\mu$ V 7.1 $\mu$ V 6.9 $\mu$ V 6.9 $\mu$ V 6.9 $\mu$ V 6.9 $\mu$ V 10 $\mu$ V 19 $\mu$ V 38 $\mu$ V 59 $\mu$ V 92 $\mu$ V		



0654

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Calibration, Maintenance and Repair Ltd

Issue No: 039 Issue date: 02 September 2021

#### Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
<b>AC VOLTAGE</b> (continued)				A
Generation (continued)	22 mV to 220 mV 10 Hz to 20 Hz	76 $\mu$ V	For measuring instruments	
	20 Hz to 40 Hz	31 $\mu$ V		
	40 Hz to 500 Hz	23 $\mu$ V		
	500 Hz to 1 kHz	23 $\mu$ V		
	1 kHz to 10 kHz	23 $\mu$ V		
	10 kHz to 20 kHz	23 $\mu$ V		
	20 kHz to 50 kHz	39 $\mu$ V		
	50 kHz to 100 kHz	99 $\mu$ V		
	100 kHz to 300 kHz	190 $\mu$ V		
	300 kHz to 500 kHz	380 $\mu$ V		
	500 kHz to 1 MHz	740 $\mu$ V		
	220 mV to 2.2 V			
	10 Hz to 20 Hz	660 $\mu$ V		
	20 Hz to 40 Hz	250 $\mu$ V		
	40 Hz to 500 Hz	120 $\mu$ V		
	500 Hz to 1 kHz	120 $\mu$ V		
	1 kHz to 10 kHz	120 $\mu$ V		
	10 kHz to 20 kHz	120 $\mu$ V		
	20 kHz to 50 kHz	180 $\mu$ V		
	50 kHz to 100 kHz	250 $\mu$ V		
	100 kHz to 300 kHz	950 $\mu$ V		
	300 kHz to 500 kHz	2.8 mV		
	500 kHz to 1 MHz	4.7 mV		
	2.2 V to 22 V			
	10 Hz to 20 Hz	6.6 mV		
	20 Hz to 40 Hz	2.5 mV		
	40 Hz to 500 Hz	1.1 mV		
	500 Hz to 1 kHz	1.1 mV		
	1 kHz to 10 kHz	1.1 mV		
	10 kHz to 20 kHz	1.1 mV		
	20 kHz to 50 kHz	1.8 mV		
	50 kHz to 100 kHz	2.3 mV		
	100 kHz to 300 kHz	7.2 mV		
	300 kHz to 500 kHz	28 mV		
	500 kHz to 1 MHz	42 mV		
	22 V to 220 V			
	10 Hz to 20 Hz	66 mV		
	20 Hz to 40 Hz	25 mV		
	40 Hz to 500 Hz	14 mV		
	500 Hz to 1 kHz	14 mV		
	1 kHz to 10 kHz	14 mV		
	10 kHz to 20 kHz	14 mV		
	20 kHz to 50 kHz	22 mV		
	50 kHz to 100 kHz	41 mV		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
<b>AC VOLTAGE (continued)</b>				A
Generation (continued)	220 V to 1100 V		For measuring instruments	
	55 Hz to 1 kHz	110 mV		
	220 V to 500 V 50 Hz to 10 kHz	180 mV		
	500 V to 750 V 50 Hz to 10 kHz	260 mV		
	750 V to 1 kV 50 Hz to 10 kHz	350 mV	For instruments with a voltage output	
Measurement	30 $\mu$ V to 10 mV	8.0 $\mu$ V		
	1 Hz to 40 Hz	5.0 $\mu$ V		
	40 Hz to 1kHz	6.0 $\mu$ V		
	1 kHz to 20 kHz	13 $\mu$ V		
	20 kHz to 50 kHz	60 $\mu$ V		
	50 kHz to 100 kHz	160 $\mu$ V		
	100 kHz to 1 MHz			
	10 mV to 100 mV	15 $\mu$ V		
	1 Hz to 40 Hz	13 $\mu$ V		
	40 Hz to 1kHz	20 $\mu$ V		
	1kHz to 20 kHz	38 $\mu$ V		
	20 kHz to 50 kHz	100 $\mu$ V		
50 kHz to 100 kHz	360 $\mu$ V			
100 kHz to 300 kHz	2.3 mV			
300 kHz to 1 MHz				
100mV to 1V	140 $\mu$ V			
1 Hz to 40 Hz	120 $\mu$ V			
40 Hz to 1kHz	200 $\mu$ V			
1kHz to 20 kHz	390 $\mu$ V			
20 kHz to 50 kHz	970 $\mu$ V			
50 kHz to 100 kHz	3.6 mV			
100 kHz to 300 kHz	24 mV			
300 kHz to 1 MHz				
1 V to 10 V	2.0 mV			
1 Hz to 40 Hz	1.0 mV			
40 Hz to 1 kHz	2.0 mV			
1kHz to 20 kHz	4.0 mV			
20 kHz to 50 kHz	10 mV			
50 kHz to 100 kHz	36 mV			
100 kHz to 300 kHz	240 mV			
300 kHz to 1 MHz				



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
--	-------	--	---------	------------------

Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.

<b>AC VOLTAGE (continued)</b> Measurement (continued)	10V to 100V 1 Hz to 40 Hz 40 Hz to 1kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100kHz to 300 kHz	38 mV 36 mV 37 mV 51 mV 150 mV 480 mV	For instruments with a voltage output.	A
	100V to 1 kV 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	520 mV 490 mV 720 mV 1.4 V 3.5 V		
<b>AC CURRENT</b> Generation	10 µA to 220 µA		For the calibration measuring instruments	A
	55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	38 nA 38 nA 88 nA 360 nA		
	220 µA to 2.2 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	320 nA 320 nA 660 nA 3.6 µA		
	2.2 mA to 22 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	29 µA 29 µA 55 µA 290 µA		
	22 mA to 220 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	3.0 µA 3.0 µA 5.7 µA 34 µA		
	220 mA to 2.2 A 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	660 µA 660 µA 1.3 mA 18 mA		
	10 Hz to 1 kHz 2 A to 3 A 3 A to 5 A 5 A to 10 A	7.0 mA 9.0 mA 14 mA		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
<b>AC CURRENT</b> (continued)				A
Generation (continued)	10 Hz to 100 Hz 10 A to 32 A 16 A to 160 A 160 A to 1000 A	9.0 mA 35 mA 580 mA	Calibration of current clamps using a multi turn coil	
Measurement	20 nA to 10 $\mu$ A 20 Hz to 45 kHz 45 Hz to 1 kHz	23 nA 14 nA	For instruments which generate current	
	10 $\mu$ A to 100 $\mu$ A 20 Hz to 45 kHz 45 Hz to 1 kHz	210 nA 110 nA		
	100 $\mu$ A to 1 mA 20 Hz to 45 kHz 45 Hz to 20 kHz 20 kHz to 50 kHz	2.0 $\mu$ A 1.0 $\mu$ A 5.0 $\mu$ A		
	1 mA to 10 mA 20 Hz to 45 kHz 45 Hz to 20 kHz 20 kHz to 50 kHz	20 $\mu$ A 9.0 $\mu$ A 51 $\mu$ A		
	10 mA to 100 mA 20Hz to 45Hz 45Hz to 20kHz 20kHz to 50kHz	200 $\mu$ A 95 $\mu$ A 510 $\mu$ A		
	100 mA to 1 A 20Hz to 45Hz 45Hz to 20kHz 20kHz to 50kHz	2.0 mA 1.0 mA 12 mA		
<b>FREQUENCY</b>				A
Generation	0.5 Hz to 5.4 GHz	2.1 in $10^9$	For the calibration of measuring instruments May be reported as events per unit time For instruments outputting frequency	
	1 Hz to 100 Hz	35 $\mu$ Hz		
Generation	100 Hz to 1 kHz 1 kHz to 27 GHz	0.40 Hz 0.24 ppm	For the calibration of measuring instruments May be reported as events per unit time	
	0.5 Hz to 5.4 GHz 5.4 GHz to 20 GHz	2.1 in $10^9$ 4.0 MHz		





0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
--	-------	--	---------	------------------

Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.

<b>TIME and FREQUENCY</b>				A
Elapsed time			Mechanical timers / stop watches	
Single event	10 s to 12 hrs	0.23 s		
Revolutions Per Minute	60 RPM to 1000 RPM 1000 RPM to 30000 RPM 30000 RPM to 96000 RPM	0.37 RPM 1.6 RPM 23 RPM		
<b>CAPACITANCE</b>				A
Measurement For capacitance sources	<i>1 kHz</i> 0.1 pF to 1 pF 1 pF to 10 pF 10 pF to 100 pF 100 pF to 1000 pF	0.030 pF 0.035 pF 0.10 pF 1.2 pF	For capacitance sources	
	<i>1 MHz</i> 0.1 pF to 1 pF 1 pF to 10 pF 10 pF to 100 pF 100 pF to 1000 pF	0.0050 pF 0.030 pF 0.90 pF 8.0 pF		
Simulated For capacitance meters	<i>10 Hz to 10 kHz</i> 220 pF to 400 pF 400 pF to 1.1 nF	13 pF 16 pF		
	<i>10 Hz to 3 kHz</i> 1.1 nF to 3.3 nF	27 pF		
	<i>10 Hz to 1 kHz</i> 3.3 nF to 11 nF 11 nF to 33 nF	39 pF 180 pF		
	<i>10 Hz to 1 kHz</i> 33 nF to 110 nF 110 nF to 330 nF	380 pF 1.0 nF		
	<i>10 Hz to 600 Hz</i> 330 nF to 1.1 µF	5.0 nF		
	<i>10 Hz to 300 Hz</i> 1.1 µF to 3.3 µF	12 nF		
	<i>10 Hz to 150 Hz</i> 3.3 µF to 11 µF	39 nF		
	<i>10 Hz to 120 Hz</i> 11 µF to 33 µF	140 nF		
	<i>10 Hz to 80 Hz</i> 33 µF to 110 µF	610 nF		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
<b>CAPACITANCE</b> (continued) For capacitance meters	<i>0 Hz to 50 Hz</i> 110 $\mu$ F to 330 $\mu$ F	2.0 $\mu$ F		A
	<i>0 Hz to 20 Hz</i> 330 $\mu$ F to 1.1 mF	6.0 $\mu$ F		
	<i>0 Hz to 6 Hz</i> 1.1 mF to 3.3 mF	2.0 $\mu$ F		
	<i>0 Hz to 2 Hz</i> 3.3 mF to 11 mF	60 $\mu$ F		
	<i>0 Hz to 0.6 Hz</i> 11 mF to 33 mF	280 $\mu$ F		
	<i>0 Hz to 0.2 Hz</i> 33 mF to 110 mF	1.0 mF		
<b>CALIBRATION OF INSULATION TESTERS</b>				A
Insulation Resistance	100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$	2.0 k $\Omega$ 32 k $\Omega$ 510 k $\Omega$ 7.0 M $\Omega$	Up to 1350 V	
Insulation test voltage Nominal	50 V 100 V 250 V 500 V 1 kV	1.0 V 2.0 V 3.0 V 4.0 V 7.0 V	The test voltage will normally be measured with a 1 mA load.	
Continuity Resistance	0 $\Omega$ to 1 $\Omega$ 1 $\Omega$ to 50 $\Omega$ 50 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 200 $\Omega$	58 m $\Omega$ 100 m $\Omega$ 120 m $\Omega$ 140 m $\Omega$		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
<b>Temperature indicators, calibration by electrical simulation</b>				A
Type k thermocouple	-200 °C to +1372 °C	0.73 °C	Including reference junction compensation	
Type T thermocouple	-250 °C to +400 °C	0.65 °C	Including reference junction compensation	
Type N thermocouple	-200 °C to +1300 °C	0.46 °C	Including reference junction compensation	
Type S thermocouple	0 °C to 1767 °C	0.77 °C	Including reference junction compensation	
Type J thermocouple	-210 °C to +1200 °C	0.23 °C	Including reference junction compensation	
Type E thermocouple	-250 °C to +1000 °C	0.26 °C	Including reference junction compensation	
Resistance thermometer (Pt 100)	-200°C to +800°C	0.05°C		
<b>Calibration of Oscilloscopes</b>				A
Voltage deflection	0 V to 6 V 11 V to 130 V	0.33 % 0.22 %	Into 50 Ω Into 1 MΩ	
Edge verification - Amplitude	4.5 mV to 2.27 V 1 kHz, 10 kHz & 100 kHz	2.7 %		
Time markers	100 ns to 20 ms 20 ms to 50 ms	83 ns 250 μs		
Rise time	1 kHz to 2 MHz 2 MHz to 10 MHz	300 ps 350 ps		
Bandwidth Level flatness 50 kHz ref Nominal 3 V	To 100 MHz To 300 MHz To 600 MHz To 1.1 GHz	3.3 % 5.4 % 8.5 % 9.5 %	CMC is with respect to relative amplitude of level flatness this will be quoted in frequency terms when stating bandwidth uncertainty as it will vary from device to device.	



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
<b>ACCELEROMETRY</b>			Uncertainties for accelerometry sensitivity calibrations at ambient temperature applies for masses up to 40 grams.	A
Accelerometer calibration Charge	Sensitivity range: 0.3 pC/g to 1000 pC/g (0.03 pC/ms <sup>2</sup> to 100 pC/m <sup>2</sup> )  1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	3.0 % 1.2 % 0.80 % 0.90 %	Nominal acceleration ms <sup>-2</sup> : 2 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.31 to 1.0	
Devices with integral electronics	Sensitivity range: 0.04 pC/g to 0.3 pC/g (0.004 pC/ms <sup>2</sup> to 0.03 pC/ms <sup>2</sup> )  20 Hz to 10 kHz	3.0 %	Nominal acceleration 0.31 ms <sup>-2</sup> to 1.0 ms <sup>-2</sup> :	
Voltage (mV/ms <sup>2</sup> )	0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup> 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	2.0 % 0.83 % 0.86 % 0.90 %	Nominal acceleration ms <sup>-2</sup> : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.31 to 1.0	
Resistive/Capacitive	0.004 mV/ms <sup>2</sup> to 0.03 mV/ms <sup>2</sup> 20 Hz to 10 kHz  0.01 mV/ms <sup>2</sup> to 100 mV/ms <sup>2</sup> 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	3.0 %  2.0 % 0.83 % 0.86 % 0.90 %	Nominal acceleration ms <sup>-2</sup> : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.31 to 1.0	
Transducer at any temperature from -60 °C to +180 °C:	0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup>  20 Hz to 630 Hz	3.0 %	At a nominal 1 g <sub>n</sub>  This uncertainty applies for masses up to 60 grams, larger masses can be calibrated but at greater uncertainties.	



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
<p><b>ACCELEROMETRY</b> (continued)</p> <p>Transducer at any temperature from -60 °C to +180 °C:</p> <p>Portable calibrators</p> <p>Bench Calibrators</p> <p>Acceleration</p> <p>Velocity</p> <p>Displacement</p> <p><b>ACCELERATION TRANSDUCERS - SHOCK CALIBRATION</b></p> <p><u>Piezoelectric and Piezoresistive and voltage types</u></p> <p>Transducer at nominal 23 °C</p> <p>Vibration meters and analysers</p> <p>Shock and Impulse hammers</p> <p>Spring hammers</p> <p>Impact Energy Imparted from Spring Operated Impact Test Apparatus - as specified in BS EN 60068-2-75:1997 and IEC 60068-2-75 1997</p>	<p>0.03 mV/ms<sup>2</sup> to 1000 mV/ms<sup>2</sup> 20 Hz to 630 Hz</p> <p>Frequency 160 Hz 1 ms<sup>2</sup> nominal</p> <p>10mV to 10V, 5 Hz to 10KHz Up to 25g</p> <p>10mV to 10V, 10 Hz to 1KHz Up to 1000 m/sec pk</p> <p>10 mV to 10V, 10 Hz to 250 Hz Up to 10 mm pk to pk</p> <p>Sensitivity 0.001 pC/ms<sup>2</sup> to 100 pC/ms<sup>2</sup> 0.001 mV/ms<sup>2</sup> to 10 mV/ms<sup>2</sup> Within the range of: 4 ms<sup>2</sup> to 1000 ms<sup>2</sup></p> <p>Frequency 10 Hz to 1 kHz Range 0.1 ms<sup>2</sup> to 1 ms<sup>2</sup></p> <p>Half Sine Wave Frequency 500 Hz to 20 kHz Applied Force 0.2 N to 450 N</p> <p>0.1 Joule to 1.0 Joule</p>	<p>3.0 %</p> <p>1.5 %</p> <p>2.4 %</p> <p>2.4 %</p> <p>2.4 %</p> <p>3.0 %</p> <p>5.0 %</p> <p>4.0 %</p> <p>0.015 Joule</p>	<p>At a nominal 1 g<sub>n</sub> This uncertainty applies for masses up to 60 grams, larger masses can be calibrated but at greater uncertainties</p> <p>Note indicated outputs on Bench Calibrators for acceleration are normally indicated in "g".</p> <p>Calibration of charge sensitivity by comparison with a reference (precision grade) transducer The transducer to be calibrated must have a mass of no more than 40 grams.</p> <p>Calibrations can be given in Joule or Newton Metre units.</p>	<p>A</p> <p>A</p>



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
--	-------	--	---------	------------------

Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.

**Gateshead Laboratory**

<b>ELECTRICAL</b>				B
DC VOLTAGE Generation	0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V	9.3 $\mu$ V 47 $\mu$ V 510 $\mu$ V 7.4 mV 24 mV		B
Measurement	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV	1.7 $\mu$ V 5.8 $\mu$ V 54 $\mu$ V 0.82 mV 9.1 mV		B
DC RESISTANCE Generation	0 $\Omega$ to 11 $\Omega$ 11 $\Omega$ to 33 $\Omega$ 33 $\Omega$ to 110 $\Omega$ 110 $\Omega$ to 330 $\Omega$ 330 $\Omega$ to 1.1 k $\Omega$  1.1 k $\Omega$ to 3.3 k $\Omega$ 3.3 k $\Omega$ to 11 k $\Omega$ 11 k $\Omega$ to 33 k $\Omega$ 33 k $\Omega$ to 110 k $\Omega$  110 k $\Omega$ to 330 k $\Omega$ 330 k $\Omega$ to 1.1 M $\Omega$ 1.1 M $\Omega$ to 3.3 M $\Omega$ 3.3 M $\Omega$ to 11 M $\Omega$ 11 M $\Omega$ to 33 M $\Omega$  33 M $\Omega$ to 110 M $\Omega$ 110 M $\Omega$ to 330 M $\Omega$ 330 M $\Omega$ to 1.1 G $\Omega$	0.79 m $\Omega$ 1.5 m $\Omega$ 3.9 m $\Omega$ 11 m $\Omega$ 36 m $\Omega$  110 m $\Omega$ 130 m $\Omega$ 1.1 $\Omega$ 3.6 $\Omega$  13 $\Omega$ 43 $\Omega$ 240 $\Omega$ 1.7 k $\Omega$ 9.6 k $\Omega$  65 k $\Omega$ 1.1 M $\Omega$ 19 M $\Omega$		B
Measurement	0 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 1 k $\Omega$ to 10 k $\Omega$ 10 k $\Omega$ to 100 k $\Omega$  100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$	270 $\mu$ $\Omega$ 2.5 m $\Omega$ 14 m $\Omega$ 140 m $\Omega$ 1.4 $\Omega$  22 $\Omega$ 0.74 k $\Omega$ 64 k $\Omega$		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
DC CURRENT				B
Generation	0 $\mu$ A to 330 $\mu$ A 330 $\mu$ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA  330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20.5 A  11 A to 30 A 30 A to 110 A 110 A to 150 A 150 A to 550 A 550 A to 1025 A	81 nA 450 nA 4.2 $\mu$ A 42 $\mu$ A  300 $\mu$ A 1.4 mA 7.0 mA 25 mA  110 mA 420 mA 560 mA 2.0 A 3.9 A	Current clamp calibration using a multi turn coil	
Measurement	0 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	3.6 nA 31 nA 320 nA 5.2 $\mu$ A 150 $\mu$ A		
AC VOLTAGE				B
Generation	1 mV to 33 mV 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz  33 mV to 330 mV 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz	38 $\mu$ V 13 $\mu$ V 13 $\mu$ V 15 $\mu$ V 45 $\mu$ V 150 $\mu$ V 360 $\mu$ V  130 $\mu$ V 65 $\mu$ V 65 $\mu$ V 71 $\mu$ V 140 $\mu$ V 340 $\mu$ V 850 $\mu$ V		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
AC VOLTAGE (cont'd)				B
Generation (cont'd)	330 mV to 3.3 V 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz  3.3 V to 33 V 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz  33 V to 330 V 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz  330 V to 1020 V 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	1.2 mV 640 μV 640 μV 800 μV 1.2 mV 2.8 mV 9.8 mV  12 mV 6.4 mV 6.4 mV 9.9 mV 14 mV 36 mV  75 mV 83 mV 100 mV 120 mV 820 mV  380 mV 130 mV 380 mV		
Measurement	10 mV to 100 mV 40 Hz to 1kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	14 μV 21 μV 39 μV 100 μV 360 μV		





0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE (cont'd) Measurement (cont'd)	100 mV to 1 V 40 Hz to 1kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	110 $\mu$ V 200 $\mu$ V 390 $\mu$ V 970 $\mu$ V 3.6 mV		B
	1 V to 10 V 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	1.2 mV 2.0 mV 3.9 mV 9.6 mV 36 mV		
	10 V to 100 V 40 Hz to 1kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	26 mV 27 mV 43 mV 140 mV		
	100 V to 700 V 40 Hz to 1 kHz	490 mV		
AC CURRENT Generation	29 $\mu$ A to 330 $\mu$ A 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.88 $\mu$ A 0.69 $\mu$ A 0.50 $\mu$ A 1.3 $\mu$ A 3.3 $\mu$ A		B
	330 $\mu$ A to 3.3 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	7.8 $\mu$ A 5.0 $\mu$ A 4.0 $\mu$ A 7.9 $\mu$ A 19 $\mu$ A		
	3.3 mA to 33 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	71 $\mu$ A 37 $\mu$ A 18 $\mu$ A 33 $\mu$ A 80 $\mu$ A		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
**Issue No: 039 Issue date: 02 September 2021**

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
--	-------	--	---------	------------------

Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.

AC CURRENT (cont'd) Generation (cont'd)	33 mA to 330 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz  330 mA to 1.1 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz  1.1 A to 3 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz  3 A to 11 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz  11 A to 20.5 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz  45 Hz to 1 kHz 11 A to 30 A  30 A to 110 A 45 Hz to 100 Hz 100 Hz to 1 kHz  110 A to 150 A 45 Hz to 1 kHz  150 A to 550 A 45 Hz to 100 Hz 100 Hz to 1 kHz  550 A to 1025 A 45 Hz to 100 Hz 100 Hz to 1 kHz	710 $\mu$ A 370 $\mu$ A 180 $\mu$ A 440 $\mu$ A 880 $\mu$ A  2.5 mA 0.81 mA 8.8 mA 38 mA  6.4 mA 2.3 mA 22 mA 92 mA  11 mA 16 mA 0.38 mA  36 mA 42 mA 720 mA  110 mA  430 mA 450 mA  610 mA  2.1 A 2.3 A  4.1 A 4.6 A	Calibration of current clamps using a multi turn coil	
--	--	--	---	--



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT Measurement	100 Hz to 5 kHz 0 □A to 100 □A 100 □A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	0.11 μA 0.60 μA 5.9 μA 59 μA 0.33 mA		
FREQUENCY				B
Generation	10 MHz	3.2 in 10 <sup>9</sup>	Reference	
Measurement	10 Hz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 50 MHz 50 MHz to 600 MHz 600 MHz to 6 GHz	16 mHz 170 mHz 1.7 Hz 17 Hz 4.8 Hz 35 Hz		
Generation	1 Hz to 10 Hz  10 Hz to 100 Hz 100 Hz to 20 MHz	2.3 ppm  1.2 ppm 0.012 ppm	May be reported as events per unit time	
Elapsed time			Mechanical timers / stop watches	
Single event	10 s to 12 hrs	0.23 s		
Revolutions Per Minute Simulation	6 RPM to 600 RPM 600 RPM to 6000 RPM 6000 RPM to 120000 RPM	0.064 RPM 0.065 RPM 0.066 RPM	Optical	
CAPACITANCE				B
Simulated generation	1 kHz 220 pF to 400 pF 400 pF to 1.1 nF 1.1 nF to 3.3 nF 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF	14 pF 18 pF 31 pF 44 pF 110 pF 330 pF 990 pF		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
--	-------	--	---------	------------------

Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.

CAPACITANCE (cont'd) Simulated generation (cont'd)	100 Hz 0.33 $\mu$ F to 1.1 $\mu$ F 1.1 $\mu$ F to 3.3 $\mu$ F 3.3 $\mu$ F to 11 $\mu$ F 11 $\mu$ F to 33 $\mu$ F 33 $\mu$ F to 110 $\mu$ F 110 $\mu$ F to 330 $\mu$ F 330 $\mu$ F to 1.1 mF 1.1 mF to 3.3 mF 3.3 mF to 11 mF 11 mF to 33 mF 33 mF to 110 mF	4.4 nF 13 nF 44 nF 190 nF 700 nF 2.1 $\mu$ F 7.0 $\mu$ F 21 $\mu$ F 70 $\mu$ F 320 $\mu$ F 1.5 mF		B
---	--	---	--	---



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
CALIBRATION OF INSULATION TESTERS				B
Insulation Resistance	100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$	2.0 k $\Omega$ 32 k $\Omega$ 510 k $\Omega$ 7.0 M $\Omega$	Up to 1350 V	
Insulation test voltage Nominal	50 V 100 V 250 V 500 V 1 kV	1.0 V 2.0 V 3.0 V 4.0 V 7.0 V	The test voltage will normally be measured with a 1 mA load.	
Continuity Resistance	0 $\Omega$ to 1 $\Omega$ 1 $\Omega$ to 50 $\Omega$ 50 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 200 $\Omega$	58 m $\Omega$ 100 m $\Omega$ 120 m $\Omega$ 140 m $\Omega$		
Temperature indicators, calibration by electrical simulation				B
Type E thermocouple	-250 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -25 $^{\circ}$ C -25 $^{\circ}$ C to +350 $^{\circ}$ C 350 $^{\circ}$ C to 650 $^{\circ}$ C 650 $^{\circ}$ C to 1000 $^{\circ}$ C	0.62 $^{\circ}$ C 0.20 $^{\circ}$ C 0.18 $^{\circ}$ C 0.20 $^{\circ}$ C 0.26 $^{\circ}$ C	Including reference junction compensation	
Type J thermocouple	-210 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -30 $^{\circ}$ C -30 $^{\circ}$ C to +150 $^{\circ}$ C 150 $^{\circ}$ C to 760 $^{\circ}$ C 760 $^{\circ}$ C to 1200 $^{\circ}$ C	0.34 $^{\circ}$ C 0.23 $^{\circ}$ C 0.18 $^{\circ}$ C 0.21 $^{\circ}$ C 0.28 $^{\circ}$ C	Including reference junction compensation	
Type k thermocouple	-200 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -25 $^{\circ}$ C -25 $^{\circ}$ C to +120 $^{\circ}$ C 120 $^{\circ}$ C to 1000 $^{\circ}$ C 1000 $^{\circ}$ C to 1370 $^{\circ}$ C	0.41 $^{\circ}$ C 0.23 $^{\circ}$ C 0.21 $^{\circ}$ C 0.31 $^{\circ}$ C 0.47 $^{\circ}$ C	Including reference junction compensation	



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
Temperature indicators, calibration by electrical simulation (cont'd)				
Type N thermocouple	-200 °C to -100 °C -100 °C to -25 °C -25 °C to +120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.48 °C 0.28 °C 0.25 °C 0.23 °C 0.33 °C	Including reference junction compensation	
Type R thermocouple	0 °C to 250 °C 250 °C to 400 °C 400 °C to 1000 °C 1000 °C to 1767 °C	0.76 °C 0.46 °C 0.43 °C 0.50 °C	Including reference junction compensation	
Type S thermocouple	0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1400 °C 1400 °C to 1767 °C	0.66 °C 0.48 °C 0.47 °C 0.57 °C	Including reference junction compensation	
Type T thermocouple	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.80 °C 0.30 °C 0.21 °C 0.18 °C	Including reference junction compensation	
Resistance thermometer Pt 100 Generation: simulation	-200 °C to -80 °C -80 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 400 °C 400 °C to 630 °C 630 °C to 800 °C	0.058 °C 0.058 °C 0.081 °C 0.10 °C 0.12 °C 0.14 °C 0.27 °C		
Pt100 Measure: simulation	-200 °C to 0 °C 0 °C to 850 °C	0.0069 °C 0.020 °C		



0654

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Calibration, Maintenance and Repair Ltd**  
Issue No: 039 Issue date: 02 September 2021

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
Calibration of Oscilloscopes				B
Voltage deflection	1 mV to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 6 V	0.29 % + 76 $\mu$ V 0.29 % + 250 $\mu$ V 0.29 % + 3.6 mV 0.29 % + 9.0 mV	Into 50 $\Omega$	
	1 mV to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 11 V 11 V to 130 V	0.12 % + 61 $\mu$ V 0.12 % + 68 $\mu$ V 0.12 % + 800 $\mu$ V 0.12 % + 4.0 mV 0.12 % + 40 mV	Into 1 M $\Omega$	
Time markers	2 ns to 500 ns 500 ns to 500 $\mu$ s 500 $\mu$ s to 50 ms 50 ms to 5 s	0.082 % + 0.20 ps 0.082 % + 0.50 ns 0.058 % + 50 ns 0.59 % + 0.25 ms		
<b>PRESSURE</b>				
Pneumatic Pressure (gauge)			Methods consistent with EURAMET CG17. Absolute pressures can be generated within these gauge pressure ranges. This will attract an additional uncertainty of 10 Pa.	
Calibration of pressure indicating instruments and gauges	- 95 kPa to -10 kPa -10 kPa to -1.5 kPa 1.5 kPa to 10 kPa 10 kPa to 20 kPa 20 kPa to 2.5 MPa 2.5 MPa to 14 MPa	84 ppm 87 ppm + 0.77 Pa 61 ppm + 0.77 Pa 65 ppm 55 ppm 73 ppm	Sensors with an electrical output can be calibrated	
Hydraulic Pressure (gauge)				
Calibration of pressure indicating instruments and gauges	0.6 MPa to 6 MPa 6 MPa to 120 MPa	97 ppm 74 ppm		
END				



0654

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

#### Calibration, Maintenance and Repair Ltd

Issue No: 039 Issue date: 02 September 2021

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

### Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· $p$  + (0.12·10<sup>-6</sup>· $p$ ·10<sup>-6</sup>) + 4.0 Pa, where  $p$  is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 ·  $i$ , where  $i$  is the instrument indication.