


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0716 Accredited to ISO/IEC 17025:2017	Tru-Cal Metrology Ltd Issue No: 016 Issue date: 21 September 2021	
	Unit 1 Gorton Industrial Estate Froxmer Street Gorton Manchester M18 8EF	Contact: Mr R Desmond Tel: +44 (0)161 223 4028 Fax: +44 (0)161 223 6028 E-Mail: info@tru-cal.co.uk Website: www.tru-cal.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 1 Gorton Industrial Estate Froxmer Street Gorton Manchester M18 8EF	Local contact Mr R Desmond	Dimensional A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises The location must be suitable for the calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Local contact Mr R Desmond	Dimensional B



0716
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tru-Cal Metrology Ltd

Issue No: 016 Issue date: 21 September 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH			2. Calibrations may be given in inch units where applicable.	
Gauge blocks Inch (Steel and Tungsten Carbide)	As BS 4311:2007 0 in to 0.4 in 0.4 in to 1in 2 in 3 in 4 in	(See notes) Class C 3.0 4.0 5.0 μ in. 6.0 7.0	Comparison Class C uncertainties apply to the measurement of length of gauges by comparison with grade K standards of length of a similar material.	A
Millimetre (Steel and Tungsten Carbide)	BS EN ISO 3650:1999 0 to 10 10 in to 25 30, 40, 50 60, 70, 75 80, 90, 100	Class C 0.080 0.10 0.12 0.15 0.18	Class C uncertainties apply to new and used grade 0, 1 and 2 gauges to BS 4311:2007 and BS EN ISO 3650:1999.	
Plain plug gauge parallel, cylindrical setting standards and rollers	1 to 50 diameter 50 to 100 100 up to 150	0.8 on diameter 1.0 1.5	By comparison to length standards using a length measuring machine	A
Plain plug gauges (taper) parallel to 1 in 8 on diameter	0 to 100 diameter	3.0 on diameter	Using a length measuring machine	A
Plain ring gauges (parallel)	2 to 50 50 to 100 100 to 150 150 to 350	0.8 1.0 1.5 4.2	By comparison to master ring gauge using a length measuring machine	A
Plain ring gauges (taper) parallel to 1 in 8 on diameter	3 to 25 diameter 25 to 100	4.0 on diameter 12	Using a length measuring machine	A
Length gauges, flat and spherical ended	0 to 575	1.0 + (8.0 x length in m)	By comparison to length standards using a length measuring machine	A
Plain gap gauges (parallel)	2 to 125 125 to 150	3.0 4.0	By comparison to length standards	A
Feeler gauges	0.025 to 1	2.0	Calibration as BS 957:2008	A



0716
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tru-Cal Metrology Ltd
Issue No: 016 Issue date: 21 September 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH (cont'd)				
Screw plug gauges (parallel) including check and setting plugs	0 to 150 diameter	4.0 on pitch diameter See Note 3	3. Single start symmetrical thread forms only. By comparison length measuring machine and thread measuring cylinders	A
Screw plug gauges (taper)	0 to 150 diameter	5.0 on pitch diameter See Note 3	By comparison to length measuring machine and thread measuring cylinders	A
Screw ring gauges (parallel)	6 to 150 diameter	5.0	By comparison length standards using a length measuring machine	A
Screw ring gauges (parallel)	1 to 16 diameter	See Note 4	4. Functional test of size using check plugs	A
Screw ring gauges (taper)	6 to 150 diameter	6.0	By comparison length standards using a length measuring machine	A
Screw pitch	0.2 to 8	1.5	Using length measuring machine and a pitch attachment	A
Screw flank angle	0° to 52°	5.0 minutes of arc	By optical methods	A
Screw thread adjustable caliper gauges (parallel)	1 to 100 diameter	See Note 5	5. Functional test of size using setting plugs calibrated with a CMC of 4.0 µm	A
Parallels	5 to 50 x 100 x 400	2.5 to 5.0	Calibration as BS 906:Part 1:1972	A
Vee blocks	20 to 150	4.0 to 6.0	Calibration as BS 3731:1987	A
Receiver, position and profile gauges, jigs and fixtures See note 6	Maximum dimensions 0 to 400 x 700 x 600	Minimum per co- ordinate 3.0 + (20 x length in m)	6. Features and associated parts of these gauges can be measured to the uncertainties given for equivalent items listed in this schedule.	A



0716
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tru-Cal Metrology Ltd

Issue No: 016 Issue date: 21 September 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
ANGLE			1. The uncertainty quoted is for the departure from flatness, straightness, parallelism or squareness, ie the distance separating the two parallel planes which just enclose the surface under consideration.	
Squares				
Blade type	0 to 600	5.0 On Squareness See Note 1	Calibration as BS 939:2007	A
Block type	0 to 600	7.0	Calibration as BS 939:2007	A
Spirit levels	5 seconds of arc to 60 minutes of arc nominal sensitivity	Mean sensitivity: 10% of nominal Minimum 0.50 seconds of arc	Calibration as BS 958:1968 or BS 3509:1962	A
Electronic indicating levels	0 minutes of arc to 20 minutes of arc	1.0 % range Minimum 0.50 seconds of arc	Using a small angle generator	A
Right angle and box angle plates	50 to 600	Squareness 6.0 Parallelism 3.0 See note 1	Calibration as BS 5535:1978	A
FORM			1. The uncertainty quoted is for the departure from flatness, straightness, parallelism or squareness, ie the distance separating the two parallel planes which just enclose the surface under consideration.	
Surface plates Granite and Cast iron	160 x 100 to 1600 x 1000	1.5 + (0.80 x diagonal in m) See Note 1	Calibration as BS 817:2008	A, B



0716
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tru-Cal Metrology Ltd
Issue No: 016 Issue date: 21 September 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
MEASURING INSTRUMENTS AND MACHINES				
Micrometers				
External	0 to 600	Heads: 2.0	Calibration as BS 870:2008 Calibration as BS 959:2008 Calibration as BS 6468:2008	A
Internal	0 to 900	Setting and extension rods		
Depth	0 to 300	1.0 + (8.0 x length in m)		
3 and 2 point bore micrometers	2 to 6	Heads 3.0 Setting 2.5	Calibration by comparison to setting rings	A
3 point bore micrometers	6 to 25	Heads 3.0 Setting 2.5	Calibration by comparison to setting rings	A
	25 to 50	Heads 3.0 Setting 3.0		
	50 to 100	Heads 4.0 Setting 5.0		
	100 to 180	Heads 5.0 Setting 5.0	Calibration by comparison to setting ring and calibration fixture	A
Height setting micrometer	0 to 300	Heads 1.50 between any two points Stepped column 2.5 Overall Performance 3.0	By comparison to length standards	A
Riser blocks for above	150 300	3.0 5.0	By comparison to length standards using a length measuring machine	A
Vernier type gauge including dial and digital				A
Caliper,	0 to 600	Overall performance 10 + (30 x length in m)	Calibration as BS 887:2008 Calibration as BS 1643:2008 Calibration as BS 6365:2008	A
Height	0 to 600			
Depth	0 to 600			
Bevel protractor	0° to 360°	6 0 minutes of arc	Calibration as BS 1685:2008	A
Dial gauges and dial test indicators	0 to 50	1.5	Calibration as BS 907:2008 or BS 2795:1981	A



0716
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tru-Cal Metrology Ltd
Issue No: 016 Issue date: 21 September 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
MEASURING INSTRUMENTS AND MACHINES con't				
Comparators (external)	250 to 10 000 magnifications	1% of range Minimum 0.20	Calibration as BS 1054:1975	A
Profile projectors	10 to 100 magnification Linear 0 to 300 Angular 0° to 360°	125 at the screen 5.0 3.0 minutes of arc	By comparison to reference scales	A, B
Electronic microprocessor controlled height gauges	0 to 1000	7.0	By comparison to length standards	A, B
END				



0716
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tru-Cal Metrology Ltd
Issue No: 016 Issue date: 21 September 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$