

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0768</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Megger Instruments Limited</h3> <p>Issue No: 027 Issue date: 19 April 2022</p>	
	<p>Standards & Calibration Laboratory Archcliffe Road Dover Kent CT17 9EN</p>	<p>Contact: Mr Gary Potter Tel: +44 (0)1304 502121 Fax: +44 (0)1304 502268 E-Mail: gary.potter@megger.com Website: www.megger.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Expanded Uncertainty ($k = 2$)	Remarks
<p>Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct or ratio-metric comparison unless otherwise described in the remarks column.</p>			
DC RESISTANCE			
Generation / Sourcing			
Specific Values	0.0001 Ω 0.001 Ω 0.01 Ω 0.1 Ω 0.2 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω 10 G Ω 100 G Ω 1 T Ω	90 $\mu\Omega/\Omega$ 92 $\mu\Omega/\Omega$ 60 $\mu\Omega/\Omega$ 21 $\mu\Omega/\Omega$ 66 $\mu\Omega/\Omega$ 18 $\mu\Omega/\Omega$ 5.4 $\mu\Omega/\Omega$ 3.8 $\mu\Omega/\Omega$ 3.7 $\mu\Omega/\Omega$ 6.4 $\mu\Omega/\Omega$ 16 $\mu\Omega/\Omega$ $\mu\Omega/\Omega$ + 300 m Ω 25 $\mu\Omega/\Omega$ + 4 Ω 82 $\mu\Omega/\Omega$ + 200 Ω 410 $\mu\Omega/\Omega$ + 1.6 k Ω 0.13 % + 3 M Ω 0.13 % + 30 M Ω 0.67 % + 15 M Ω 1.6 % + 6.1 G Ω	Source values for the calibration of ohmmeters.
Other Values	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω 1 G Ω to 10 G Ω 10 G Ω to 500 G Ω	500 $\mu\Omega/\Omega$ + 0.30 m Ω 510 $\mu\Omega/\Omega$ + 1.0 m Ω 500 $\mu\Omega/\Omega$ + 2.7 m Ω 500 $\mu\Omega/\Omega$ + 29 m Ω 500 $\mu\Omega/\Omega$ + 780 m Ω 500 $\mu\Omega/\Omega$ + 12 Ω 0.20 % + 690 Ω 0.25 % + 32 k Ω 1.0 % + 59 k Ω 1.0 % + 10 M Ω 2.3 % + 53 M Ω	



0768
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Megger Instruments Limited
Issue No: 027 Issue date: 19 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Uncertainty (k = 2)	Remarks
Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct or ratio-metric comparison unless otherwise described in the remarks column.			
DC RESISTANCE (continued)			
Measurement	100 μΩ to 1 mΩ 1 mΩ to 10 mΩ 10 mΩ to 100 mΩ 100 mΩ to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω	91 μΩ/Ω 92 μΩ/Ω 61 μΩ/Ω 58 μΩ/Ω 58 μΩ/Ω + 14 μΩ 59 μΩ/Ω + 280 μΩ	Resistance sources can be calibrated to these uncertainties at currents of up to 100 A
	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ 2 GΩ to 20 GΩ	17 μΩ/Ω + 0.12 mΩ 16 μΩ/Ω + 0.12 mΩ 11 μΩ/Ω + 1.2 mΩ 8.9 μΩ/Ω + 1.2 mΩ 9.4 μΩ/Ω + 12 mΩ 8.8 μΩ/Ω + 48 mΩ 12 μΩ/Ω + 1.0 Ω 31 μΩ/Ω + 93 Ω 150 μΩ/Ω + 9.3 kΩ 0.15 % + 0.93 MΩ 0.18 % + 9.3 MΩ	Resistance sources can be calibrated to these uncertainties using a Fluke 8508A
	1 MΩ to 5 TΩ	0.12 % to 0.90 %	The applied voltage depends on the nominal resistance and may be up to 20 kV.
Simulated values to support Megger range of insulation testers	2 TΩ to 12 TΩ 12 TΩ to 35 TΩ	5.0 % 11 %	3 terminal devices
DC VOLTAGE			
Generation	0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1 kV	70 μV/V + 2.0 μV 58 μV/V + 6.0 μV 58 μV/V + 61 μV 64 μV/V + 660 μV 64 μV/V + 4.0 mV	Source values for the calibration of voltmeters
	1 kV to 20 kV 20 kV to 40 kV	0.10 % + 20 V 0.10 % + 120 V	
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV 1 kV to 15 kV 15 kV to 90 kV	6.5 μV/V + 1.2 μV 3.6 μV/V + 1.2 μV 3.6 μV/V + 4 μV 5.5 μV/V + 39 μV 5.6 μV/V + 40 μV 0.20 % + 0.14 V 1.20 %	Voltage sources can be calibrated to these uncertainties
Thermocouple Simulation Type K	-25 °C to +120 °C	0.16 % + 0.19 °C	Simulated using Fluke 5500A



0768
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Megger Instruments Limited
Issue No: 027 Issue date: 19 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Uncertainty ($k = 2$)	Remarks
<p>Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct or ratio-metric comparison unless otherwise described in the remarks column.</p>			
DC CURRENT			
Generation	0 μ A to 330 μ A 330 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 2.2 A 2.2 A to 11 A 11 A to 550 A	0.14 % + 290 nA 0.12 % + 350 nA 157 μ A/A + 60 nA 210 μ A/A + 12 μ A 210 μ A/A + 13 μ A 800 μ A/A + 870 μ A 0.76 % + 50 mA	Ammeters can be calibrated to these uncertainties Simulated current using 50 turn coil, for the calibration of clamp-on ammeters.
Measurement	0 pA to 200 pA 200 pA to 2 nA 2 nA to 20 nA 20 nA to 200 nA 200 nA to 2 μ A 2 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A 20 A to 200 A	1.2 % + 220 fA 0.23 % + 60 fA 0.23 % + 660 fA 0.23 % + 610 fA 0.12 % + 9.0 pA 14 μ A/A + 0.31 nA 14 μ A/A + 3.1 nA 15 μ A/A + 31 nA 47 μ A/A + 620 nA 170 μ A/A + 12 μ A 390 μ A/A + 120 μ A 60 μ A/A + 52 μ A	Current sources can be calibrated to these uncertainties.
AC VOLTAGE			
Generation	40 Hz to 10 kHz 1 mV to 33 mV 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1 kV 50 Hz 2 kV to 40 kV	0.17 % + 23 μ V 0.062 % + 26 μ V 0.037 % + 136 μ V 0.048 % + 700 mV 0.060 % + 23 mV 0.060 % + 95 mV 1.7 % + 120 V	Source values for the calibration of voltmeters
AC POWER			
Generation	45 Hz to 65 Hz 1 VA to 1100VA	0.40 %	
Measurement	1 mV to 200 mV 40 Hz to 10 kHz 10 kHz to 100 kHz 200 mV to 2 V 40 Hz to 10 kHz 10 kHz to 100 kHz 2 V to 20 V 40 Hz to 10 kHz 10 kHz to 100 kHz 20 V to 200 V 40 Hz to 10 kHz 10 kHz to 100 kHz	0.011 % + 3.9 μ V 0.067 % + 19 μ V 0.011 % + 19 μ V 0.050 % + 190 μ V 0.011 % + 190 μ V 0.050 % + 1.9 mV 0.011 % + 1.9 mV 0.051 % + 19 mV	Voltage sources can be measured to these uncertainties



0768
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Megger Instruments Limited
Issue No: 027 Issue date: 19 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Uncertainty ($k = 2$)	Remarks
Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct or ratio-metric comparison unless otherwise described in the remarks column.			
AC VOLTAGE (Continued)	200 V to 1 kV 40 Hz to 10 kHz	0.012 % + 19 mV	
	At 50 Hz 1 kV to 50 kV	1.7 % + 120 V	
AC VOLTAGE RATIO	40 Hz to 120 Hz 0.8 to 2000 2000 to 20000	0.011 % 0.038 % + 0.0050	At voltages of up to 200 V
AC CURRENT Generation	10 Hz to 1 kHz 0.03 μ A to 330 μ A 330 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA	0.29 % + 290 nA 0.23 % + 370 nA 0.23 % + 4.0 μ A 0.23 % + 50 μ A	Ammeters can be calibrated to these uncertainties
	10 Hz to 1 kHz 330 mA to 2.2 A 2.2 A to 11 A	0.24 % + 410 μ A 0.38 % + 2.7 mA	
	40 Hz to 400 Hz 11 A to 550 A	0.31 % + 600 mA	Simulated current using 50 turn coil, for the calibration of clamp-on ammeters.
Measurement	10 Hz to 5 kHz 12 nA to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA	490 μ A/A + 19 nA 300 μ A/A + 190 nA 290 μ A/A + 1.9 μ A 280 μ A/A + 19 μ A	Outputs of current sources can be measured with these uncertainties.
	10 Hz to 1 kHz 200 mA to 2 A 2 A to 20 A	560 μ A/A + 190 μ A 720 μ A/A + 1.9 mA	
	5 kHz 200 mA to 2 A	0.28 % + 650 μ A	
	50 Hz 11 A to 2 kA	0.85 % + 130 mA	
FREQUENCY Measurement	1 Hz to 10 MHz	1.0 part in 10^7	These outputs can be measured
Generation			
Specific Values	0.1 Hz, 1 Hz, 10 Hz, 100 Hz, 1 kHz, 10 kHz, 100 kHz, 1 MHz, 5 MHz and 10 MHz	1.4 parts in 10^7	Source values for measuring instruments
PHASE ANGLE	50 Hz and 60 Hz 0° to 360°	70 m°	



0768
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Megger Instruments Limited
Issue No: 027 Issue date: 19 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Uncertainty ($k = 2$)	Remarks
Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct or ratio-metric comparison unless otherwise described in the remarks column.			
CAPACITANCE Generation Specific Values Measurement	1 kHz 100 pF 1000 pF, 10 nF, 100 nF and 1 μF 1 μF 1 kHz 100 pF to 10 μF	0.020 % 0.010 % 0.020 % 0.20 % to 0.050 %	Standard capacitors available for the calibration of capacitance bridges, meters etc. Capacitors can be calibrated to these uncertainties.
INDUCTANCE Generation Specific Values Measurement	1 kHz 100 μH, 1 mH, 10 mH, 100 mH and 1 H 1 kHz 100 μH to 1 H	0.010 % 0.20 % to 0.050 %	Standard inductors available for the calibration of inductance bridges, meters etc. Inductors can be calibrated to these uncertainties.
AC RESISTANCE	40 Hz to 1 kHz 1 Ω to 10 kΩ 50 Hz 0.02 Ω to 5 Ω 5 Ω to 1 kΩ 40 Hz to 1592 Hz 1 Ω to 10 kΩ	0.10 % 0.58 % + 36 mΩ 0.58 % + 50 mΩ 0.050 % to 0.20 %	Sourcing suitable for measuring devices Nominal values suitable for the calibration of earth loop testers Measurement of AC resistors
RCD TRIP TIMES ELAPSED TIME To support Megger products: Oil test set & check meters; The OTS range, Oil tan delta test set (otd) and otd cc.	30 ms to 400 ms Nominal 60 s	0.010 % + 1.0 ms 5.0 ms	Suitable for RCD testers Electronically triggered
DC VOLTAGE AC VOLTAGE	0 V to 1 kV 200 mV to 1 kV; 55 Hz 10 kV to 100 kV; 50 Hz to 75 Hz	0.030 % + 0.10 V 0.070 % + 0.20 V 0.85 % + 80 V	



0768
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Megger Instruments Limited
Issue No: 027 Issue date: 19 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Uncertainty ($k = 2$)	Remarks
Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. The method used is by direct or ratio-metric comparison unless otherwise described in the remarks column.			
DISSIPATION FACTOR 750 V at 55 Hz 2 kV at 55 Hz	0 nominal 0.003 nominal 0.03 nominal 0.3 nominal 3 nominal 0.003 nominal to 0.3 nominal 0.003 nominal 0.03 nominal	0.030 % + 0.00001 0.080 % + 0.00003 0.030 % + 0.00017 0.080 % + 0.0015 0.030 % + 0.015 1.5 % + 0.00030 0.080 % + 0.00003 0.030 % + 0.00017	At temperatures of up to 100 °C
PERMITTIVITY 750 V and 2 kV at 55 Hz 200 pF applied RESISTANCE At 500 V RESISTIVITY At 500 V DC CAPACITANCE 750 V and 2 kV at 55 Hz	3 nominal 5 MΩ to 5 GΩ 40 MΩ·m to 40 GΩ·m 200 pF	0.060 % 0.15% 0.050 % 0.060 %	
END			



0768
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Megger Instruments Limited
Issue No: 027 **Issue date: 19 April 2022**

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$