


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION 0775</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Fluke (UK) Limited</h3> <p>Issue No: 022 Issue date: 04 January 2021</p>	
	<p>Fluke Industrial Temperature Laboratory UK 52 Hurricane Way Norwich Norfolk NR6 6JB</p>	<p>Contact: Mr Andrew Johns Tel: +44 (0)1603 256745 Fax : +44 (0)1603 256777 E-mail: andrew.johns@fluke.com Website: www.fluke.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
TEMPERATURE			Calibration by comparison with a reference standard
4-wire platinum resistance thermometers	-196 °C -80 °C to +100 °C 100 °C to 250 °C	50 mK 50 mK 100 mK	In Liquid Nitrogen In a liquid bath
Thermistors	-70 °C to +150 °C	0.050 °C	
Temperature indicators and recorders with temperature sensor(s)		(as above with an allowance for display resolution and short-term stability)	
Type T thermocouples	-196 °C -80 °C to +250 °C	0.26 °C 0.28 °C	In Liquid Nitrogen In a liquid bath
Other base metal thermocouples	-196 °C -80 °C to +250 °C	0.26 °C 0.28 °C	In Liquid Nitrogen In a liquid bath
Electronic thermometers with indicators	Ranges as for sensors	as for sensors	
Calibration in air chamber	-40 °C to +70 °C	0.25 °C	in air chamber
HUMIDITY			Calibration by comparison with dew-point hygrometer and Platinum Resistance Thermometers
Relative humidity	25 %rh to 90 %rh at ambient temperature	2.8 % of reading	The accreditation covers other units derived from those listed
Dew point	0 °C dp to 20 °C dp	0.25 °C dp	



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Fluke (UK) Limited
Issue No: 022 Issue date: 04 January 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL			Calibration by comparison with a reference standard
DC Voltage	-100 mV to -50 mV -50 mV to +50 mV 50 mV to 100 mV	10 ppm 500 nV 10 ppm	Generation and measurement for the calibration of voltage based transducers
DC Resistance	0 Ω to 200 Ω 80 Ω to 10 M Ω	17 m Ω 0.10%	For the calibration of thermistor indicators
DC Current	4 mA to 20 mA	3.0 μ A	Measurement of transducer output of 4 to 20 mA devices
Temperature indicators, calibration by electrical simulation			
Base metal thermocouples	-200 $^{\circ}$ C to 0 $^{\circ}$ C 0 $^{\circ}$ C to 1200 $^{\circ}$ C 1200 $^{\circ}$ C to 1370 $^{\circ}$ C	0.46 $^{\circ}$ C 0.26 $^{\circ}$ C 0.22 $^{\circ}$ C	Including cold junction compensation
Resistance thermometers (Pt100)	-200 $^{\circ}$ C to +200 $^{\circ}$ C 200 $^{\circ}$ C to 800 $^{\circ}$ C	0.050 $^{\circ}$ C 0.070 $^{\circ}$ C	
END			



0775

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Fluke (UK) Limited

Issue No: 022 Issue date: 04 January 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· p + (0.12·10⁻⁶· p ·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i , where i is the instrument indication.