

Schedule of Accreditation

issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Accredited to
ISO/IEC 17025:2017

Weightron Bilanciai Limited

Issue No: 018 Issue date: 02 October 2022

Unit 4B, Block 4
Westpark
Chelston
Wellington
TA21 9AD

Contact: Lee Pavey
Tel: +44 (0)1823 662355
E-Mail: lee.pavey@weightroncb.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Address Customers' sites or premises The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Mass - weighing machines (non-automatic)	S



0780

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Weightron Bilanciai Limited**Issue No:** 018 **Issue date:** 02 October 2022

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
NON AUTOMATIC WEIGHING MACHINES (self indicating)	200 mg 500 mg 1g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 100 kg 200 kg 500 kg 1000 kg 1500 kg 3000 kg 6000 kg	0.030 mg 0.040 mg 0.050 mg 0.060 mg 0.081 mg 0.10 mg 0.12 mg 0.16 mg 0.27 mg 0.55 mg 1.3 mg 2.7 mg 5.5 mg 13 mg 250 mg 500 mg 1.2 g 2.6 g 6.1 g 14 g 29 g 40 g 150 g 310 g	Weights are available in OIML Class F1 from 200 mg g to 2 kg Max grouped load 5 kg F2 from 20 g g to 2 kg Max grouped load 5 kg M1 from 5 kg to 20 kg Max. grouped load 2000 kg And the addition of, M ₁₋₂ 6000kg Max grouped load 8000 kg Methods consistent with EURAMET CG18	S

END



0780

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Weightron Bilanciai Limited**Issue No:** 018 **Issue date:** 02 October 2022**Calibration performed by the Organisation at the locations specified**

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$