


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0783</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>CET Structures Ltd. trading as CET Calibration</p> <p>Issue No: 014 Issue date: 02 February 2018</p>	
	<p>M1 Commerce Park Markham Lane Duckmanton Chesterfield S44 5HS</p>	<p>Contact: Mr Chris Locke Tel: +44 (0)1246 828318 Fax: +44 (0)1246 828319 E-Mail: chris.locke@cetcalibration.com Website: www.cetcalibration.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL MEASUREMENTS			
DC VOLTAGE			
Measurement	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	9.2 ppm + 0.49 μ V 9.0 ppm 10 ppm 13 ppm 13 ppm	
Generation	0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V	70 ppm + 4.0 μ V 58 ppm + 6.0 μ V 58 ppm + 60 μ V 67 ppm 65 ppm	
DC CURRENT			
Measurement	0 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 20 A	160 ppm + 24 nA 190 ppm 190 ppm 190 ppm 300 ppm 65 ppm	
Generation	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 2.2 A 2.2 A to 11 A	150 ppm + 60 nA 120 ppm + 0.40 μ A 120 ppm + 5.0 μ A 350 ppm + 60 μ A 700 ppm + 0.46 mA	



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL MEASUREMENTS (continued)			
DC RESISTANCE			
Measurement	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	18 ppm + 180 $\mu\Omega$ 16 ppm 14 ppm 13 ppm 14 ppm 20 ppm 43 ppm 450 ppm 2.1 %	
Generation			
Specific Values			
Decade Values	1 m Ω 10 m Ω 100 m Ω 1 Ω 10 M Ω 100 M Ω 1 G Ω 10 G Ω 100 G Ω 1 T Ω	75 ppm 59 ppm 63 ppm 63 ppm 1.2 % 1.3 % 2.6 % 1.0 % 1.2 % 1.7 %	At 100 V, 500 V and 1000 V At 100 V, 500 V and 1000 V At 100 V, 500 V and 1000 V At 100 V and 500 V At 100 V and 500 V At 100 V and 500 V
Decade and non-decade values	10 M Ω 20 M Ω 30 M Ω 40 M Ω 50 M Ω 60 M Ω 70 M Ω 80 M Ω 90 M Ω 100 M Ω 110 M Ω	23 ppm 46 ppm 69 ppm 92 ppm 120 ppm 140 ppm 160 ppm 190 ppm 210 ppm 230 ppm 250 ppm	
Other values	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω	580 ppm 120 ppm 120 ppm 120 ppm 120 ppm 150 ppm	



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL MEASUREMENTS (continued)			
AC VOLTAGE			
Measurement	2 mV to 100 mV 20 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	61 ppm + 16 μ V 67 ppm + 27 μ V 67 ppm + 61 μ V	
	100 mV to 1 V 20 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	170 ppm 280 ppm 710 ppm 0.47 % 2.3 %	
	1 V to 10 V 20 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	170 ppm 280 ppm 710 ppm 0.47 % 2.3 %	
	10 V to 100 V 20 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	170 ppm 280 ppm 710 ppm	
	100 V to 500 V 50 Hz to 10 kHz 10 kHz to 30 kHz	160 ppm 280 ppm	
	500 V to 1000 V 50 Hz to 10 kHz 10 kHz to 30 kHz	210 ppm 310 ppm	
Generation	1 mV to 330 mV 45 Hz to 10 kHz 10 kHz to 50 kHz	0.10 % 0.21 %	
	330 mV to 3.3 V 45 Hz to 10 kHz 10 kHz to 50 kHz	0.036 % 0.039 %	
	3.3 V to 33 V 45 Hz to 10 kHz 10 kHz to 50 kHz	0.046 % 0.22 %	
	33 V to 330 V 45 Hz to 20 kHz	0.10 %	
	330 V to 1100 V 45 Hz to 1 kHz 1 kHz to 20 kHz	0.058 % 0.23 %	



0783
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration
Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL MEASUREMENTS (continued)			
AC CURRENT			
Measurement	2 μ A to 100 μ A 50 Hz to 5 kHz	630 ppm + 0.13 μ A	
	100 μ A to 1 mA 50 Hz to 5 kHz	640 ppm	
	1 mA to 10 mA 50 Hz to 5 kHz	610 ppm	
	10 mA to 100 mA 50 Hz to 5 kHz	610 ppm	
	100 mA to 1 A 50 Hz to 1 kHz 1 kHz to 5 kHz	0.12 % 0.35 %	
	1 A to 20 A 50 Hz to 5 kHz	0.10 %	
Generation	30 μ A to 3.3 mA 10 Hz to 5 kHz	0.23 %	
	3.3 mA to 33 mA 10 Hz to 5 kHz	0.23 %	
	33 mA to 330 mA 10 Hz to 5 kHz	0.23 %	
	330 mA to 2.2 A 10 Hz to 1 kHz 1 kHz to 5 kHz	0.23 % 0.87 %	
	2.2 A to 11 A 45 Hz to 500 Hz 500 Hz to 1 kHz	0.12 % 0.38%	



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL TEMPERATURE SIMULATION			
Temperature Indicators and Simulators (t/c Types) calibration by Electrical Simulation			
Base Metal Thermocouples	Type K -200°C to -100°C	0.40°C	Including Cold Junction Compensation
	Type K -100°C to -25°C	0.24°C	
	Type K -25°C to 120°C	0.22°C	
	Type K 120°C to 1000°C	0.32°C	
	Type K 1000°C to 1372°C	0.48°C	Including Cold Junction Compensation
	Type J -210°C to -100°C	0.33°C	
	Type J -100°C to -30°C	0.22°C	
	Type J -30°C to 150°C	0.20°C	
	Type J 150°C to 760°C	0.23°C	
	Type J 760°C to 1200°C	0.29°C	
	Type N -200°C to -100°C	0.14°C	Including Cold Junction Compensation
	Type N -100°C to -25°C	0.12°C	
	Type N -25°C to 120°C	0.11°C	
	Type N 120°C to 410°C	0.11°C	
Type N 410°C to 1300°C	0.11°C	Including Cold Junction Compensation	
Type T -250°C to -150°C	0.74°C		
Type T -150°C to 0°C	0.30°C		
Type T 0°C to 120°C	0.22°C		
Type T 120°C to 400°C	0.20°C		
Nobel Metal Thermocouples	Type B 600°C to 800°C		0.59°C
	Type B 800°C to 1000°C	0.23°C	
	Type B 1000°C to 1550°C	0.21°C	
	Type B 1550°C to 1820°C	0.22°C	Including Cold Junction Compensation
	Type R 0°C to 250°C	0.68°C	
	Type R 250°C to 400°C	0.42°C	
	Type R 400°C to 1000°C	0.40°C	
	Type R 1000°C to 1767°C	0.48°C	
	Type S 0°C to 250°C	0.56°C	
	Type S 250°C to 1000°C	0.44°C	
	Type S 1000°C to 1400°C	0.44°C	
	Type S 1400°C to 1767°C	0.55°C	
	Type E -250°C to -100°C	0.59°C	Including Cold Junction Compensation
	Type E -100°C to -25°C	0.22°C	
Type E -25°C to 350°C	0.20°C		
Type E 350°C to 650°C	0.22°C		
Type E 650°C to 1000°C	0.70°C		



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL TEMPERATURE SIMULATION (continued)			
Cold Junction Compensation	At ambient temperature of 23°C ± 2°C	0.20°C	
Temperature Indicators and Simulators (PT100 Types) calibration by Electrical Simulation			
PT 100 Indicators	-200°C to 0°C 0°C to 100°C 100°C to 300°C 300°C to 400°C 400°C to 630°C 630°C to 800°C	0.077°C 0.095°C 0.120°C 0.130°C 0.150°C 0.270°C	
PT 100 Simulators	-200 °C to 0 °C 0 °C to 560 °C 560 °C to 850 °C	0.053 °C 0.050 °C 0.072 °C	



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	
DIMENSIONAL MEASUREMENTS				
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH				
Plain plug gauges (parallel)	1 to 50 diameter 50 to 100 diameter 100 to 200 diameter	1.0 on diameter 1.0 on diameter 1.5 on diameter	NOTES In addition to these items, other similar items, including parts of measuring instruments and machines, may be calibrated in accordance with the stated CMCs. Where the item or part calibrated is of lower quality due to wear, errors in geometry or form, or poor surface texture, or where any other factor adversely affects the measurement capability, greater uncertainties may be quoted. All linear calibrations may be given in inch units.	
Plain ring gauges (parallel)	2 to 10 diameter 10 to 50 diameter 50 to 100 diameter 100 to 200 diameter	1.5 on diameter 1.5 on diameter 2.0 on diameter 2.4 on diameter		
Length gauge, flat and spherical ended (excluding length bars)	0 to 1200	1.0 + (8.0 x length in m)		
Plain gap gauges (parallel)	BS 969:2008 2 to 100 100 to 200 200 to 300	3.0 5.0 8.0		
Feeler gauges	BS 957:2008 0.020 to 1	2.0		
Paint thickness setting foils	0.020 to 1	2.0		
Parallels	BS 906:Part 1:1972 up to 5 to 50 x 100 x 400	Dependent on size and grade From 2.0 to 5.0		
Rule – steel	BS 4372:1968 0 to 1000 1000 to 2000	5.0 + (50 x length in m) 10 + (50 x length in m)		
ANGLE				
Squares				
Blade type	BS 939:2007 50 to 300 300 to 600	3.0 5.0 On squareness	The CMCs are for the departure from flatness, straightness, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration.	



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DIMENSIONAL MEASUREMENTS (continued)			
Right angle and box angle plates	BS 5535:1978 50 to 600	Squareness: 3.0 + (1.0 per 100 mm) Parallelism 1.0 + (1.0 per 100 mm)	The CMCs are for the departure from flatness, straightness, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration.
Bevel protractors	BS 1685:2008 0° to 360°	7.5 min of arc	
MEASURING INSTRUMENTS AND MACHINES			
Micrometers			
External	As BS 870:2008 (and above) 0 to 1000	Heads: 2.0	
Internal	As BS 959:2008 0 to 900	Setting and extension rods	
Depth	As BS 6468:2008 0 to 300	1.0 + (8.0 x length in m)	
Vernier caliper, height and depth gauges	BS 887:2008 (and above) 0 to 1200 BS 1643:2008 0 to 1000 BS 6365:2008 0 to 600	Overall performance 10 + (30 x length in m)	
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 50	1.0	
Electronic indicators	0 to 5 0 to 50	Lever type 1.0 Linear type 0.7	



0783

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

CET Structures Ltd. trading as CET Calibration

Issue No: 014 Issue date: 02 February 2018

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
 - As an explicit function of the measurand or of a parameter (see below).
 - As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
 - As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V:

Over the range 100 mV to 1 V, the CMC is 0.0025 % \cdot V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 % \cdot p + (0.12 \cdot 10⁻⁶ \cdot p \cdot 10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 \cdot 0.01 \cdot i, where i is the instrument indication.

END