


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines Upon Thames. TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0789</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Campbell Associates Ltd</p> <p>Issue No: 019 Issue date: 23 August 2021</p>	
	<p>Sonitus House 5B Chelmsford Road Industrial Estate Great Dunmow Essex CM6 1HD</p>	<p>Contact: Mr David Egan Tel: +44 (0)1371 871030 Fax: +44 (0)1371 879106 E-Mail: ian@campbell-associates.co.uk Website: www.campbell-associates.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ACOUSTICS			
<u>Pistonphones & sound calibrators</u>			
Sound pressure level	250 Hz 1000 Hz	0.10 dB	Using Norsonic 1504 with NOR-1018 Software
Sound pressure level of multi-frequency calibrator	31.5 Hz to 63 Hz	0.13 dB	
	>63 to 5 kHz	0.10 dB	
	>5 kHz to 8 kHz	0.15 dB	
	>8 kHz to 12.5 kHz	0.21 dB	
Amplitude stability	>12.5 kHz to 16 kHz	0.30 dB	With WS2P microphone
	Dependent on instrument	0.02 dB	
Frequency	63 Hz to 16 kHz	0.10 % of reading	
Distortion	Dependent on instrument	14 % of reading	
Periodic testing of sound calibrators in accordance with IEC 60942:2003	90 to 140 dB	Uncertainties as listed above See also remarks	Periodic testing of sound calibrators Class LS, 1 or 2 using Insert voltage technique using WS2P or LS2Pmicrophone as
<u>Sound level meters</u>			
Verification of Sound Level Meters	BS 7580:Part 1:1997	See remarks	Verification of Type 0, 1 & 2 SLMs originally manufactured in accordance with BS EN 60651:1994 BS EN 60804:1994 and for which appropriate correction factors are known and agreed



0789
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines Upon Thames. TW18 3HR, UK

Campbell Associates Ltd
Issue No: 019 **Issue date:** 23 August 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ACOUSTICS (cont'd)			
<u>Sound level meters</u>			
Verification of Sound Level Meters	BS EN 61672-3:2006 (Withdrawn) as modified by UKAS TPS 49 Edition 2. June 2009.	See remarks	Verification of Class 1 & 2 SLMs originally manufactured in accordance with IEC 61672- 3:2006 and for which required correction factors are known and agreed, including measurement of self generated noise with microphone fitted at customers request.
Verification of Sound Level Meters	BS EN 61672-3:2013	See remarks	Verification of Class 1 & 2 SLMs originally manufactured in accordance with IEC 61672- 3:2013 and for which required correction factors are known and agreed, including measurement of self generated noise with microphone fitted at customers request.
Filters - sound level meter based octave band filters one-third octave band filters	16 Hz to 16 kHz 16 Hz to 20 kHz	0.13 dB 0.13 dB	Filters originally manufactured in accordance with IEC 61260:1995 (BS EN 61260:1996) or IEC 60225 in combination with a sound level meter
Reverberation time One- third octave bands	50 Hz to 10 kHz For R _t times of 0.1, 0.2, 0.5, 1 and 2 seconds 5 and 10 seconds	0.01 s 0.06 s	Verification of specific RT modules on sound level meters using transfer reference audio files i.e. computer generated multi-sine files to give the required decay curves
Microphones Pressure sensitivity of 1", ½" & ¼" microphones @ reference frequency	250 Hz	0.1 dB	WSM type microphones
Electrostatic actuator response of 1" microphones	100 Hz – 4 kHz >4 kHz – 8 kHz >8 kHz to 12.5 kHz	0.21 dB 0.24 dB 0.48 dB	By electrostatic actuator methods



0789
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines Upon Thames. TW18 3HR, UK

Campbell Associates Ltd
Issue No: 019 Issue date: 23 August 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ACOUSTICS (cont'd)			
Microphones			
Electrostatic actuator response of 1/2" microphones	100 Hz – 4 kHz	0.21 dB	By electrostatic actuator methods The upper frequency limit for high sensitivity 1/2" microphones is 20 kHz
	>4 kHz – 8 kHz	0.24 dB	
	>8 kHz to 16 kHz	0.48 dB	
	>16 kHz to 20 kHz	0.7 dB	
	>20 kHz to 50 kHz	0.9 dB	
Electrostatic actuator response of 1/4" microphones	100 Hz – 4 kHz	0.21 dB	By electrostatic actuator methods
	>4 kHz – 8 kHz	0.24 dB	
	>8 kHz to 16 kHz	0.48 dB	
	>16 kHz to 20 kHz	0.7 dB	
	>20 kHz to 50 kHz	0.9 dB	
Polarised self-capacitance of 1", 1/2" & 1/4" microphones @ 250 Hz	1 pF to 100 pF	0.3%	
Low frequency response of 1/2" microphones (with pressure equalisation vent exposed to sound field)	2 Hz to 4 Hz	0.45 dB	Using microphone test chamber
	>4 Hz to 25 Hz	0.27 dB	
	>25 Hz to 100 Hz	0.26 dB	



0789
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines Upon Thames. TW18 3HR, UK

Campbell Associates Ltd
Issue No: 019 Issue date: 23 August 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ACOUSTICS (cont'd)			
Tapping Machines -verification	In support of BS EN ISO 16283-2:2015 & BS EN ISO 140-7:1998 (Withdrawn)		
	Velocity 0.70 m/s to 1.00 m/s	0.01 m/s	
	Mass 480 g to 520 g	0.17 g	
	Time 50 ms to 150 ms	0.25 ms	
	Diameter 25 mm to 35 mm	0.03 mm	
	Radius of curvature 300 mm to 700 mm	11 mm	
	Angle of fall 0° to 0.6°	0.07°	
ACCELEROMETRY			Portable vibration field calibrators to documented in- house procedure TP-15 "Calibration of Vibration Calibrators"
<u>Portable vibration field calibrators</u>			
Acceleration:			
10 Hz to 20 Hz	1 ms ⁻² to 100 ms ⁻²	1.11 %	Certificate of Conformance to BS EN ISO 8041:2005 Annex A for devices with matching specification by periodic verification
20 Hz to 80 Hz	1 ms ⁻² to 100 ms ⁻²	0.72 %	
80 Hz	1 ms ⁻² to 100 ms ⁻²	0.56 %	
80 Hz to 1 kHz	1 ms ⁻² to 100 ms ⁻²	0.75 %	
1 kHz to 2 kHz	1 ms ⁻² to 100 ms ⁻²	1.50 %	
Frequency:			
8 Hz to 1280 Hz	1 ms ⁻² to 100 ms ⁻²	0.17 %	
Distortion (percentage of reading)	1 ms ⁻² to 100 ms ⁻²	1.05 %	

END



0789
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines Upon Thames. TW18 3HR, UK

Campbell Associates Ltd
Issue No: 019 **Issue date:** 23 August 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$