

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0805</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Testo Industrial Services Ltd</b>	
	Issue No: 022    Issue date: 05 October 2021	
	Stanley House Old Brick Kiln Monk Sherborne Road Ramsdell Hampshire RG26 5PR	Contact: Mr Paul Miller Tel: +44 (0)1252 648989 E-Mail: <a href="mailto:info@testotis.co.uk">info@testotis.co.uk</a> Website: <a href="http://www.testotis.co.uk">www.testotis.co.uk</a>
Calibration performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Stanley House Old Brick Kiln Monk Sherborne Road Ramsdell Hampshire RG26 5PR	<a href="#">Humidity</a> <a href="#">Temperature</a> <a href="#">Air velocity</a> <a href="#">Pressure</a>	P

#### Site activities performed away from the location listed above:

Location details	Activity	Location code
Any customer premises  The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	<a href="#">Temperature</a> <a href="#">Humidity</a> <a href="#">Pressure</a>	S



0805  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Testo Industrial Services Ltd**  
**Issue No: 022 Issue date: 05 October 2021**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	
HUMIDITY Relative humidity	for the temperature range 0 °C to 20 °C 10 %rh to 35 %rh 35 %rh to 65 %rh 65 %rh to 90 %rh	0.70 %rh 1.0 %rh 1.3 %rh	Calibration by comparison with a reference chilled mirror hygrometer and thermometers  Calibration of devices with an electrical output may be undertaken	P
	for the temperature range 20 °C to 30 °C 10 %rh to 40 %rh 40 %rh to 70 %rh 70 %rh to 80 %rh 80 %rh to 90 %rh	0.55 %rh 0.90 %rh 0.97 %rh 1.1 %rh		
	for the temperature range 30 °C to 50.5 °C 10 %rh to 35 %rh 35 %rh to 65 %rh 65 %rh to 90 %rh	0.45 %rh 0.80 %rh 1.0 %rh		
Relative humidity	for the temperature range 5 °C to 25 °C 10 %rh to 35 %rh 23 %rh to 65 %rh 65 %rh to 90 %rh	1.6 %rh 2.1 %rh 2.6 %rh	Calibration by comparison with a reference hygrometer and thermometers	S
	for the temperature range 25 °C to 50 °C 10 %rh to 35 %rh 23 %rh to 65 %rh 65 %rh to 90 %rh	1.6 %rh 2.3 %rh 3.0 %rh		
Dew point/Frost point	- 30 °C to + 48 °C	0.16 °C	Calibration by comparison with a reference chilled mirror hygrometer	P
TEMPERATURE			Calibration by comparison with a reference instruments	
Electronic thermometers with indicators and data- loggers	- 40 °C to 110 °C	0.039 °C	In liquid bath	P
	110 °C to 200 °C	0.055 °C		
	-70 °C to 100 °C	0.070 °C		
Resistance thermometers	100 °C to 180 °C	0.11 °C	In air chamber	P
	- 40°C to 200 °C	0.060 °C		
Electronic thermometers with indicators and data- loggers	- 70 °C to 180 °C	0.12 °C	In air chamber	P
	-80 °C to 180 °C	0.26 °C		



0805  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Testo Industrial Services Ltd**  
**Issue No: 022 Issue date: 05 October 2021**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>AIR VELOCITY</b>			Calibration by comparison with a reference in a characterised wind tunnel.	
Calibration of vane anemometers by comparison	0.2 m/s to 1 m/s 1 m/s to 15 m/s 15 m/s to 20 m/s 20 m/s to 30 m/s	0.10 m/s 0.20 m/s 0.32 m/s 0.60 m/s	Anemometer up to 100 mm diameter can be calibrated.	P
Calibration of thermal anemometers by comparison	0.1 m/s to 1 m/s 1 m/s to 10 m/s 10 m/s to 15 m/s 15 m/s to 20 m/s	0.13 m/s 0.37 m/s 0.75 m/s 0.91 m/s	Uncertainty and range achieved will depend on the size of the anemometer calibrated	
<b>PRESSURE</b>			Methods consistent with EURAMET CG17.	
Gas pressure gauge				
Calibration of pressure indicating instruments and gauges	-100 kPa to -2 kPa -2 kPa to -200 Pa -200 Pa to 0 Pa 0 Pa to 200 Pa 200 Pa to 20 kPa 20 kPa to 100 kPa 100 kPa to 7 MPa	2.7 kPa 17 Pa 4.9 Pa 1.4 Pa 16 Pa 2.7 kPa 0.019 % + 8.0 Pa	Absolute pressure calibrations can be undertaken using associated barometric pressure measurement correction. The uncertainties quoted will be increased by 72 Pa	P
Gas pressure absolute			Calibrations of pressure devices with an electrical output may be undertaken	
Calibration of pressure indicating instruments and gauges	5 kPa to 160 kPa	72 Pa		
Gas pressure gauge				S
Calibration of pressure indicating instruments and gauges	-85 kPa to -7 kPa -7 kPa to -2 kPa -2 kPa to -200 Pa -200 Pa to 200 Pa 200 Pa to 20 kPa 20 kPa to 35 kPa 35 kPa to 7 MPa	3.1 kPa 26 Pa 20 Pa 6.0 Pa 20 Pa 26 Pa 2 kPa		
END				



0805  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Testo Industrial Services Ltd**  
**Issue No: 022 Issue date: 05 October 2021**

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$