

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0820</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>SGS United Kingdom Limited</p> <p>Issue No: 017 Issue date: 09 February 2022</p>	
	<p>Address Metrology Laboratory Unit 1, Rossmore Business Park Ellesmere Port South Wirral Cheshire CH65 3EN</p>	<p>Contact: Darryl Devaney Tel: +44 (0)7889 939567 E-Mail: darryl.devaney@sgs.com Website: www.sgs.com</p>
<p>Calibration performed by the Organisation at the locations specified</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Metrology Laboratory Unit 1, Rossmore Business Park Ellesmere Port South Wirral Cheshire CH65 3EN</p> <p>Local contact Paul Hannam</p>	<p>Gas flow Hydrocarbon flow Volume tanks and provers</p>	A
<p>Address 2 Martin Close Blenheim Industrial Estate Bulwell Nottingham NG6 8UW</p> <p>Local contact Sam Cooke</p>	<p>Dimensional Electrical Pressure Temperature Humidity</p>	B
<p>Address Aurora Court Barton Road Riverside Park Middlesbrough TS2 1RY</p> <p>Local contact Jake Bird</p>	<p>Electrical Pressure Temperature</p>	C
<p>Address Aberdeen Laboratory SGS House Well Heads Drive Dyce Aberdeen AB21 7GQ</p> <p>Local contact Colin Brown</p>	<p>Volume tanks and provers Volume - pipe provers Flow meters</p>	D



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Paul Hannam Hydrocarbon flow Volume tanks and provers	A1
	Sam Cooke Dimensional Electrical Pressure Temperature	B1
	Colin Brown Volume tanks and provers Volume - pipe provers Flow meters	D1



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
FLOW				
Gas meters Volume passed	0.15 m ³ to 350 m ³ (At flow rates of 1.2 m ³ /hour to 4579 m ³ /hour)	0.60 %	Calibration against reference meters using air	A
	30 litres to 300 litres (At flow rates of 40 litres/hour to 65,000 litres/hour)	0.60 %	Calibration against bell provers using air	A
	30 litres to 300 litres (At flow rates of 40 litres/hour to 6,000 litres/hour)	0.60 %	Calibration against sonic nozzles using air.	A
Hydrocarbon oil flow meters Volume passed	500, 2500 & 5000 litres at flow rates from 50 l min ⁻¹ to 5000 l min ⁻¹	0.052 % of measured quantity	Calibrations can be carried out using Kerosene at ambient temperature and pressures of up to 3 bar.	A
	3000 litres at flow rates from 50 l min ⁻¹ to 3000 l min ⁻¹	0.052 % of measured quantity	Calibration carried out using diesel, gas oil, kerosene or motor oil and water at ambient temperature and pressures of up to 3 bar	A1
VOLUME				
Calibration of Prover Tanks	100 litres to 5000 litres	0.020 % of measured quantity	Calibration carried out using water	A + A1
Calibration of Pipe Provers and Compact Provers using Waterdraw Method	60 litres to 40000 litres	0.025 % of measured quantity	Calibration carried out using water	A + A1
Calibration of Pipe Provers and Compact Provers using Master Meter Method	300 litres to 40000 litres	0.030 % of measured quantity	Calibration carried out using water	A + A1
Calibration of Pipe Provers and Compact Provers using Master Meter Method	300 litres to 40000 litres	0.045 % of measured quantity	Calibration carried out using hydrocarbons	A1



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH				B
Plain plug gauges (parallel), cylindrical setting standards and rollers	1 to 50	0.50		B
	50 to 100	0.80		
	100 to 150	1.0		
	150 to 200	1.2		
	200 to 300	1.6		
Thread measuring cylinders	BS 5590 0.1 to 5	0.50	Notes Note 1: plain cylindrical limit gauges are calibrated using a single axis measuring machine and reference standards	B
Plain ring gauges (parallel) and setting standards	2 to 10 diameter	1.0	Note 2: using gauge block	B
	10 to 25	0.80		
	25 to 50	1.0		
	50 to 100	1.5		
	100 to 150 150 to 250	2.0 3.0		
Length gauges, flat and spherical ended	0 to 600	1.0 + (8.0 x length in m)	Note 3: using a single axis measuring machine	B
Plain gap gauges (parallel)	0.5 to 100	3.0	Note 4: thread gauges are calibrated using NPL methods described in Notes on Applied Science No. 1	B
	100 to 200	5.0		
	200 to 300	8.0		
Feeler Gauges and paint thickness foils	BS 957:2008 0.025 to 1	1.0		B
Screw plug gauges (parallel) including check and setting plugs See Note 3	1 to 100 diameter	2.5 on pitch diameter		B
	100 to 150	5.0 on pitch diameter		
	150 to 200	8.0 on pitch diameter		
Screw plug gauges (taper) including check plugs See Note 2	5 to 100 diameter	5.0 on pitch diameter		B
Screw ring gauges (parallel) See Notes 3 and 4	3 to 75 diameter	5.0 on pitch diameter		B
	75 to 150	7.0 on pitch diameter		
	150 to 250	10 on pitch diameter		
Screw ring gauges (taper) See Note 2	6 up to 150 diameter	7.0 on pitch diameter		B
Screw pitch Screw flank angle	0.2 to 8 0° to 52°	1.5 5.0 minutes of arc		B
Parallels	BS906:1972 5 to 50 x 100 x 400	1.5 up to 5.0		B
Vee blocks	BS 3731:1987 20 to 150	2.5 to 5.0		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Receiver, position and profile gauges, jigs, fixtures See note 5	Maximum dimensions 0 to 1000 x 750 x 500 (Limited to gauges where a specific procedure and uncertainty budget are available).	Minimum per co-ordinate: 3.0 + (10 x length in m)	Documented in-house methods using a cmm and first principles	B
ANGLE				B
Squares Blade type	BS 939:2007 50 to 300 300 to 600	3.0 On squareness 5.0 See Note 1		B
Angle plates and box angle plates	BS 5535:1978 50 to 600	Squareness: 3.0 + (1.0 per 100 mm) Parallelism: 1.0 + (1.0 per 100 mm) See Note 1		B
Sine bars and tables	BS 3064:1978 0 to 500 length	Linear dimensions: 1.0 + (10 x length in m) Overall performance: 3.0 seconds of arc		B
FORM				B
Surface plates Granite and Cast iron	BS 817:2008 and above 160 x 100 to 1600 x 1000	1.5 + (0.80 diagonal in m) See Note 1		B
Steel balls	1 to 25 diameter	0.50 on diameter	Single axis measuring machine and reference standards	B
MEASURING INSTRUMENTS AND MACHINES				B
Micrometers External	BS 870:2008 0 to 600	Heads: 2.0		B
Internal	BS 959:2008 0 to 900	Setting and extension rods 1.0 + (8.0 x length in m)		B
Depth	BS 6468:2008 0 to 300			B
Micrometer heads	BS 1734:1951 0 to 50	1.0		B
Bench micrometer	NPL MOY/SCMI 22 0 to 100	Overall performance 1.0		B
Vernier caliper, height and depth gauges	BS 887:2008 0 to 1000 BS 1643:2008 0 to 1000 BS 6365:2008 0 to 600	Overall performance 10 + (30 x length in m)		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Height gauges - (Simple) including vernier, dial and digital types.	As BS EN ISO 13225:2012 0 to 1000	Length measurement error (E): $2.0 + (5.0 \times \text{length in metres})$		B
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 50	1.0		B
Comparators (external)	BS 1054:1975 250 to 10 000 magnifications	1.0 % of range Minimum 0.20		B
Graduated rules	BS 4372:1968 0 to 1000	$5.0 + (10 \times \text{length in m})$		B
Bevel protractors	BS 1685:2008 0° to 360°	6.0 minutes of arc		B
Electronic height gauges with microprocessor control	0 to 1000	$2.0 + (5.0 \times \text{length in m})$	Documented in-house methods	B
Notes for length, angle and form				
<ol style="list-style-type: none"> The uncertainty quoted is for the departure from flatness, straightness, parallelism, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration. Single start, symmetrical thread forms only. Single and multi-start symmetrical and asymmetrical thread forms. Includes use of check plugs for screw rings from 1 mm to 14 mm diameter 				
Features and associated parts of these gauges can be measured to the uncertainties given for equivalent items listed in this schedule.				
ELECTRICAL				
DC Voltage	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V 1kV to 50 kV	10 $\mu\text{V/V} + 0.80 \mu\text{V}$ 7.0 $\mu\text{V/V} + 1.0 \mu\text{V}$ 8.0 $\mu\text{V/V} + 3.5 \mu\text{V}$ 8.0 $\mu\text{V/V} + 60 \mu\text{V}$ 8.0 $\mu\text{V/V} + 0.60 \text{mV}$ 0.20 % + 7.0 V	Electrical calibrations are performed as a comparison against a reference standard. These values can be generated for the calibration of measuring instruments, Outputs of instruments can be measured directly.	B B
DC Resistance	0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 1 G Ω	25 $\mu\Omega/\Omega + 24 \mu\Omega$ 11 $\mu\Omega/\Omega + 0.10 \text{m}\Omega$ 9.0 $\mu\Omega/\Omega + 0.80 \text{m}\Omega$ 9.0 $\mu\Omega/\Omega + 8.0 \text{m}\Omega$ 13 $\mu\Omega/\Omega + 80 \text{m}\Omega$ 22 $\mu\Omega/\Omega + 1.7 \Omega$ 38 $\mu\Omega/\Omega + 100 \Omega$ 200 $\mu\Omega/\Omega + 11 \text{k}\Omega$ 0.12 % + 0.11 M Ω		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (Cont'd)				B
DC Current	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	35 μ A/A + 0.80 nA 30 μ A/A + 6.0 nA 35 μ A/A + 52 nA 46 μ A/A + 1.2 μ A 70 μ A/A + 25 μ A		B
	2 A to 30 A	400 μ A/A + 0.30 mA		
AC Voltage	20 Hz to 60 Hz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	0.025 % + 5.0 μ V 0.024 % + 24 μ V 0.021 % + 0.24 mV 0.021 % + 2.3 mV 0.024 % + 12 mV		B
	60 Hz to 3 kHz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	0.022 % + 5.0 μ V 0.014 % + 24 μ V 0.014 % + 0.24 mV 0.015 % + 2.4 mV 0.020 % + 12 mV		B
	3 kHz to 30 kHz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V	0.025 % + 10 μ V 0.015 % + 50 μ V 0.015 % + 0.50 mV 0.016 % + 5.0 mV		B
	3 kHz to 10 kHz 200 V to 1000 V	0.040 % + 120 mV		
	30 kHz to 100 kHz 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V	0.060 % + 24 μ V 0.075 % + 0.24 mV 0.030 % + 2.4 mV 0.050 % + 23 mV		B
	100 kHz to 500 kHz 200 mV to 2 V 2 V to 20 V	0.20 % + 2.4 mV 0.30 % + 24 mV		B
	500 kHz to 1 MHz 200 mV to 2 V 2 V to 20 V	0.25 % + 24 mV 0.26 % + 230 mV		B
	50 Hz 1 kV to 50 kV	0.60 % + 60 V		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (Cont'd)				B
AC Current	40 Hz to 1 kHz 1 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA	0.040 % + 24 nA 0.040 % + 0.24 μ A 0.040 % + 2.4 μ A 0.045 % + 24 μ A		B
	40 Hz to 100 Hz 200 mA to 2 A	0.030 % + 0.50 mA		
	100 Hz to 1 kHz 200 mA to 2 A	0.065 % + 1.0 mA		
	1 kHz to 5 kHz 1 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	0.10 % + 24 nA 0.075 % + 0.24 μ A 0.075 % + 2.4 μ A 0.090 % + 24 μ A 0.14 % + 1.0 mA		B
	50 Hz to 60 Hz 2 A to 30 A	0.040 % + 2.5 mA		B
Electrical calibration of temperature indicators for the following sensors:				
Noble metal thermocouples	0 °C to 1600 °C	0.50 °C	Including cold junction compensation	B
	0 °C to 1600 °C	0.15 °C	Cold junction disabled	B
	0 °C to 1600 °C	0.90 °C	Including cold junction compensation	B1
Base metal thermocouples	-200 °C to +1300 °C	0.13 °C	Including cold junction compensation	B
	-200 °C to +1300 °C	0.060 °C	Cold junction disabled	B
	-200 °C to +1300 °C	0.70 °C	Including cold junction compensation	B1
Resistance sensors	-200 °C to +250 °C +250 °C to +650 °C -200 °C to +850 °C	5.0 m°C 30 m°C 0.30 °C		B B1
Cold junction compensation of thermocouple indicators and sources	At ambient temperature	0.10 °C		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (Cont'd)				B
Capacitance	1 nF	4.2 pF		B
Sourcing	10 nF	30 pF		
	20 nF	48 pF		
	50 nF	72 pF		
	100 nF	93 pF		
	1 μ F	2.0 nF		
	10 μ F	50 nF		
Frequency Measurement	0 Hz to 3 GHz 3 GHz to 26 GHz	4.0 in 10^{12} + 3.0 mHz 3.0 in 10^9		B
Generation	0 Hz to 6 GHz	4.0 in 10^{12}		
Period	10 ns to 1700 s	1.5 in 10^9		B
Time interval	1 s to 8×10^5 s	8.0 in 10^9		B
17th and 18th edition electrical test instruments			Using the calibrator to generate known quantities to simulate measurements made by the Electrical Test Instruments.	B
Insulation Testers:				B
Resistance	10 k Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	0.30 % 0.85 % 1.5 %		B
Output Voltage DC	50 V to 1000 V	0.40 %		B
Output Current DC	0 mA to 1 mA	0.060 % + 6.0 μ A		B
Continuity Resistance	50 m Ω to 1 Ω 1 Ω to 50 k Ω	0.30 % + 12 m Ω 0.40 %		B
Continuity Current DC	0 mA to 320 mA	0.70 mA		B
Voltage AC	100 V to 400 V at 50 Hz	0.40 %		B
Loop Testers:				B
Loop Resistance	0 Ω to 5 Ω 9 Ω to 1000 Ω	0.60 % + 24 m Ω 0.60 % + 45 m Ω		B
RCD Testers:				B
Current Measurement AC	0.1mA to 3000mA at 50 Hz	1.4 % + 0.070 mA		B
Trip time	0 ms to 390 ms 390 ms to 5 s	0.70 ms 8.2 ms		B
Appliance Testers:				B
Insulation Resistance	0.1 M Ω to 10 M Ω	0.30 %		B
Earth Bond Resistance	50 m Ω to 10 Ω 10 Ω to 1000 Ω	0.60 % + 8.0 m Ω 0.60 % + 35 m Ω		B
Earth Bond Current AC	0.1 mA to 500 mA at 50 Hz 500 mA to 30 A at 50 Hz	1.75 % + 7.0 mA 1.75 % + 75 mA		B
Leakage Current AC	1 mA to 8 mA at 50 Hz	1.75 % + 10 μ A		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (Cont'd)				B & B1
Flash Test (Accessory)				B
AC Voltage	0.1 kV to 3 kV at 50 Hz	2.0 % + 12 V		B
AC Leakage Current	0.1 mA to 3 mA at 50 Hz	6.0 % + 17 μ A		B
Electrical verification of ultrasonic flaw detection equipment	AS BS EN 12668-1:2010			B
	Stability after warm up (height)	0.30 % of screen height		B
	Stability after warm up (width)	0.020 % of screen width		B
	Jitter – screen height	0.30 % of screen height		B
	Jitter – screen width	0.30 % of screen width		B
	Stability against supply variations (height)	0.30 % of screen height		B
	Stability against supply variations (width)	0.020 % of screen width		B
	Pulser Voltage	3.6 V		B
	Pulser Rise Time	1.7 ns		B
	Pulser Reverberation	1.0 % of pulsar voltage		B
	Pulse Duration	1.7 ns		B
	Frequency Response	30 kHz at -3 dB point		B
	0.1 MHz to 25 MHz			
	Equivalent input noise	$2.4 \times 10^{-9} \text{ V}/\sqrt{\text{Hz}}$		B
	Calibrated Attenuator	0.60 dB		B
	0 MHz to 100 MHz			
	Vertical Linearity	0.70 % of screen height		B
	Linearity of Timebase	0.25 % relative to 20 % and 80 % of the horizontal scale		B
PRESSURE				B & B1
Gas Pressure Gauge			Methods consistent with EURAMET CG17	B & B1
Calibration of pressure indicating instruments and gauges	-95 kPa to -20 kPa -20 kPa to 20 kPa 20 kPa to 200 kPa 200 kPa to 2 MPa	0.012 % + 66 Pa 0.015 % + 10 Pa 0.012 % + 66 Pa 0.012 % + 590 Pa		
Gas Pressure Absolute				B & B1
Calibration of pressure indicating instruments and gauges	3.5 kPa to 120 kPa 120 kPa to 200 kPa	100 Pa 140 Pa		
Hydraulic Pressure Gauge				B & B1
Calibration of pressure indicating instruments and gauges	0 Pa to 13.5 MPa 13.5 MPa to 70 MPa	0.010 % + 6.0 kPa 0.015 % + 20 kPa		



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE			Calibration by Comparison with reference instruments.	B & B1
Resistance thermometers	-20 °C to +100 °C 100 °C to 250 °C 250 °C to 600 °C 50 °C to 300 °C 300 °C to 600 °C	0.055 °C 0.075 °C 0.15 °C 0.50 °C 1.3 °C	Liquid bath Block calibrator Block calibrator	B B1
Thermocouples	-20 °C to 250 °C 250 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1300 °C 0 °C to 130 °C 130 °C to 350 °C 350 °C to 600 °C	0.50 °C 0.90 °C 1.9 °C 2.8 °C 0.60 °C 0.85 °C 1.5 °C	Liquid bath Block calibrator Spherical furnace Block calibrator	B B1
Temperature indicators with probes	-20 °C to +100 °C 100 °C to 250 °C 250 °C to 600 °C	0.060 °C 0.090 °C 0.15 °C	Liquid bath Block calibrator	B
	Ambient (typically 20 °C) 0 °C to 130 °C 130 °C to 350 °C 350 °C to 600 °C	0.070 °C 0.50 °C 0.80 °C 1.4 °C	Ambient air Block calibrator	B B1
Dry Block Calibrators	-20 °C to +600 °C	0.13 °C		B
Radiation thermometers (Pyrometers)	0 °C to +100 °C 100 °C to 200 °C 200 °C to 350 °C 350 °C to 500 °C	1.0 °C 1.5 °C 2.5 °C 3.5 °C	Radiation thermometers working in the wavelength range 8 µm to 14 µm	B
Calibration in air chamber	5 °C to 23 °C 23 °C to 60 °C	0.10 °C 0.15 °C	Comparison in air chamber	B
HUMIDITY			Calibration by comparison with a reference hygrometer and reference thermometers.	B
Relative humidity meters	At 5 °C to 12 °C 15 %rh to 90 %rh	0.45 %rh to 1.9 %rh	Comparison in air chamber	B
	At 12 °C to 23 °C 10 %rh to 90 %rh	0.50 %rh to 1.9 %rh		B
	At 23 °C to 60 °C 10 %rh to 90 %rh	0.50 %rh to 1.6 %rh		B



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
DIMENSIONAL				B1
Surface plates Granite and Cast iron	BS 817:2008 and above 160 x 100 to 1600 x 1000	1.5 + (0.80 diagonal in m) See Note 1		B1
Electronic height gauges with microprocessor control	0 to 1000	2.0 + (5.0 x length in m)		B1
PRESSURE				C
Gas pressure gauge			Methods consistent with EURAMET CG17	C
Calibration of pressure indicating instruments and gauges	- 100 kPa to - 3.5 kPa - 3.5 kPa to 3.5 kPa 3.5 kPa to 7 MPa	0.010 % 0.69 Pa 0.0080 %	Calibration of pressure devices with an electrical output can be undertaken	C
Gas pressure Absolute			Absolute pressure calibrations can be undertaken using the gauge and absolute ranges with an additional uncertainty of 10 Pa.	C
Calibration of pressure indicating instruments and gauges	80 kPa to 120 kPa	10 Pa		C
Hydraulic pressure gauge				C
Calibration of pressure indicating instruments and gauges	0.6 MPa to 120 MPa	0.013 %		C
TEMPERATURE				C
Liquid-in-glass thermometers	- 50 °C to 250 °C	0.050 °C	Plus 1/10 scale division	C
Temperature indicators with sensors	- 50 °C to 250 °C 250 °C to 350 °C 350 °C to 650 °C	0.050 °C 0.20 °C 0.30 °C	Calibrations performed in a range of dry and liquid baths	C



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column.				
ELECTRICAL				
DC Resistance	0 Ω to 10 Ω	17 $\mu\Omega/\Omega$ + 84 $\mu\Omega$		C
	10 Ω to 100 Ω	14 $\mu\Omega/\Omega$ + 650 $\mu\Omega$		C
	100 Ω to 1 k Ω	12 $\mu\Omega/\Omega$ + 630 $\mu\Omega$		
	1 k Ω to 10 k Ω	12 $\mu\Omega/\Omega$ + 120 m Ω		
	10 k Ω to 100 k Ω	12 $\mu\Omega/\Omega$ + 1.2 Ω		
	100 k Ω to 1 M Ω	18 $\mu\Omega/\Omega$ + 18 Ω		
	1 M Ω to 10 M Ω	59 $\mu\Omega/\Omega$ + 720 Ω		
	10 M Ω to 100 M Ω	580 $\mu\Omega/\Omega$ + 58 k Ω		
	100 M Ω to 1 G Ω	0.64 % + 6.0 M Ω		
	DC Voltage	0 mV to 100 mV	11 $\mu\text{V}/\text{V}$ + 1.2 μV	
100 mV to 1 V		9.0 $\mu\text{V}/\text{V}$ + 9.3 μV		
1 V to 10 V		9.0 $\mu\text{V}/\text{V}$ + 92 μV		
10 V to 100 V		12 $\mu\text{V}/\text{V}$ + 1.2 mV		
100 V to 1 kV		12 $\mu\text{V}/\text{V}$ + 18 mV		
AC Voltage	100 μV to 100 mV			C
	1 kHz	0.010 % + 23 μV		
	100 mV to 1 V			
	40 Hz to 1 kHz	0.010 % + 88 μV		
	1 V to 10 V			
	40 Hz to 1 kHz	0.010 % + 840 μV		
	10 V to 100 V			
	40 Hz to 1 kHz	0.020 % + 23 mV		
	100 V to 700 V			
	50 Hz to 1 kHz	0.050 % + 32 mV		
Generation only (Increased frequency range)	10 mV to 300 mV			C
	10 Hz to 3 kHz	0.050 % + 140 μV		
	300 mV to 3 V			
	10 Hz to 3 kHz	0.050 % + 1.4 mV		
	3 V to 30 V			
	10 Hz to 3 kHz	0.050 % + 14 mV		
	30 V to 100 V			
10 Hz to 3 kHz	0.050 % + 47 mV			
100 V to 300 V				
40 Hz to 3 kHz	0.10 % + 280 mV			
300 V to 750 V				
40 Hz to 3 kHz	0.10 % + 700 mV			



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (Cont'd)				C
DC Current	0 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	56 μ A/A + 110 pA 25 μ A/A + 310 pA 24 μ A/A + 2.6 nA 24 μ A/A + 25 nA 24 μ A/A + 250 nA 41 μ A/A + 820 nA 130 μ A/A + 130 μ A		C
Generation only (Extended range)	1 A to 3 A 3 A to 10 A	700 μ A/A + 2.1 mA 640 μ A/A + 6.4 mA		C
	3 A to 30 A 30 A to 100 A 100 A to 150 A 150 A to 500 A	0.10 % 0.091 % 0.10 % 0.091 %	Using multi turned coil, suitable only for calibration of clamp- on ammeters.	C
AC Current	1 μ A to 100 μ A 1 kHz	0.080 % + 36 nA		C
	100 μ A to 1 mA 1 kHz	0.080 % + 250 nA		
	1 mA to 10 mA 45 Hz to 1 kHz	0.11 % + 2.5 μ A		
	10 mA to 100 mA 45 Hz to 1 kHz	0.11 % + 24 μ A		
	100 mA to 1 A 45 Hz to 1 kHz	0.15 % + 250 μ A		
Generation only (Extended range and frequency)	30 μ A to 300 μ A 10 Hz to 3 kHz	0.090 % + 350 nA		C
	300 μ A to 3 mA 10 Hz to 3 kHz	0.090 % + 350 nA		
	3 mA to 30 mA 10 Hz to 3 kHz	0.090 % + 25 μ A		
	30 mA to 300 mA 10 Hz to 3 kHz	0.10 % + 280 μ A		
	300 mA to 3 A 10 Hz to 3 kHz	0.10 % + 560 μ A		
	3 A to 10 A 10 Hz to 3 kHz	0.23 % + 23 mA		
	3 A to 150 A 10 Hz to 100 Hz	0.17 %	Simulation	
	100 A and 500 A 10 Hz to 100 Hz	0.33 %		



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (Cont'd)				C
AC Current (Cont'd) Generation only (Extended range and frequency) (cont'd)	100 A to 150 A 10 Hz to 100 Hz	0.17 %		C
	150 A to 500 A 10 Hz to 100 Hz	0.33 %		C
Temperature indicators and recorders, calibration by electrical simulation			Including reference junction compensation	C
Type J	- 210 °C to -100 °C - 100 °C to + 1200 °C	0.32 °C 0.15 °C		C
Type K	- 270 °C to - 250 °C - 250 °C to - 100 °C - 100 °C to + 1370 °C	3.8 °C 1.0 °C 0.13 °C		C
Type N	- 270 °C to - 200 °C - 200 °C to - 100 °C - 100 °C to 1370 °C	3.9 °C 0.39 °C 0.18 °C		C
Type R	600 °C to 1760 °C	1.5 °C		C
Type S	700 °C to 1700 °C	1.6 °C		C
Type T	- 270 °C to - 200 °C - 200 °C to - 100 °C - 100 °C to 0 °C 0 °C to 400 °C	4.1 °C 0.52 °C 0.30 °C 0.21 °C		C
Resistance Thermometers Pt 100	- 0200 °C to 0 °C 0 °C to 850 °C	4.0 m°C 17 m °C		C
FREQUENCY				C
Frequency reference	10 MHz	8.2 in 10 ¹²		C
Generation	1 Hz to 120 MHz	16 in 10 ¹¹		C
Measurement	0.1 Hz to 10 Hz 10 Hz to 225 MHz	0.16 µHz 18 in 10 ¹²		C
Pulse counting	1 Hz to 100 kHz	1 count		C
Periodic time	1 µs to 10 s	16 ps	Repetitive signals	C



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
VOLUME				D1
Calibration of small volume Provers by water draw method	40 l 60 l	0.031 % 0.035 %	Results are based on dynamic uncertainty, influenced by site conditions, pressure, temperature and calibration medium used.	D1
Calibration of Pipe Provers by master meter and compact prover method				D1
Water And Hydrocarbons	40 l to 20000 l At flow rates of 260 l/min to 2600 l/min	0.031 %		D1
Flow – volume				D1
Water or Hydrocarbons	5 l/min to 6600 l/min	0.031 %	Suitable for the calibration of flow meters with a pulsed output by compact prover method.	D1
END				



0820

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SGS United Kingdom Limited

Issue No: 017 Issue date: 09 February 2022

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$