


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0822</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Chamois Metrology Limited</h3> <p>Issue No: 080 Issue date: 21 August 2020</p>	
	<p>Unit 8 The Centre Holywell Business Park Northfield Road Southam Warwickshire CV47 0FP</p>	<p>Contact: Mr A Garthwaite Tel: +44 (0)1926 812066 Fax: +44 (0)1926 813569 E-Mail: lab@chamois.net Website: www.chamois.net</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Unit 8 The Centre Holywell Business Park Northfield Road Southam Warwickshire CV47 0FP</p> <p>Local contact Mr A Garthwaite</p>	<p><u>Mass calibration</u> <u>Electrical calibration</u> <u>Pressure calibration</u> <u>Temperature calibration</u> <u>Dimensional calibration</u></p>	UK
<p>Address Metrology division Unit K2 M7 Business Park Newhall Naas County Kildare Ireland</p> <p>Local contact Mr P Kinsella Tel. +353 (0) 45 896660 Fax. +353 (0) 45 896713 Email: info@classictechnology.ie</p>	<p><u>Pressure calibration</u> <u>Electrical calibration</u> <u>Temperature calibration</u> <u>Mass calibration</u> <u>Humidity calibration</u></p>	IRE

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer</p> <p>Local contact Mr A Garthwaite</p>	<p><u>Pressure calibration</u></p>	Site



0822

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
<p>PRESSURE</p> <p>Gas pressure (absolute)</p> <p>Calibration of pressure measuring instruments and gauges</p> <p>Gas pressure (gauge)</p> <p>Calibration of pressure measuring instruments and gauges and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set) and Effective area calibration of Dead Weight Testers</p> <p>Gas pressure (differential)</p> <p>Calibrations of differential pressure devices with low and high pressure ports at a common mode pressure of 3.5kPa</p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>2 Pa to 160 Pa 160 Pa to 1.4 kPa 1.4 kPa to 15.7 kPa 15.7 kPa to 710 kPa 710 kPa to 27.6 MPa 27.6 MPa to 41.4 MPa</p> <p>- 100 kPa to - 3.5 kPa - 3.5 kPa to 0 Pa 0 Pa to 1.4 kPa 1.4 kPa to 15.7 kPa 15.7 kPa to 710 kPa 710 kPa to 27.6 MPa 27.6 MPa to 41.4 MPa</p> <p>6 Pa to 10 kPa (Line pressure 3.5 kPa)</p> <p>0 Pa to (7 - line pressure) MPa (Line pressure 200 kPa to 7 MPa)</p> <p>7 MPa to (27.6 - line pressure) MPa (Line pressure 7 MPa to 27.6 MPa)</p> <p>0 Pa to (41.4 - line pressure) MPa (Line pressure 27.6 MPa to 41.4 MPa)</p>	<p>10 % 0.0040 % + 32 Pa 0.0030 % + 1.0 Pa 0.0025 % + 1.0 Pa 0.0025 % + 10 Pa 0.0045 % + 10 Pa</p> <p>0.0035 % 0.0095 % + 0.60 Pa 0.0040 % + 0.50 Pa 0.0022 % + 0.030 Pa 0.0017 % 0.0025 % 0.0045 %</p> <p>0.010 % + 0.060 Pa</p> <p>0.000060 % of line pressure, plus 0.0035 % of differential pressure, plus 5.0 Pa</p> <p>0.000060 % of line pressure, plus 0.0035 % of differential pressure, plus 10 Pa</p> <p>0.000065 % of line pressure, plus 0.0060 % of differential pressure, plus 16 Pa</p>	<p>Methods consistent with EURAMET CG3 and CG17</p> <p>Calibration of pressure measuring devices with an electrical output may be undertaken.</p> <p>Calibrations may also be performed over an environmental temperature range of -10 °C to +150 °C, with an uncertainty of ± 1 °C on the reported temperature. There will be an additional pressure uncertainty of $\pm (30 \text{ ppm} + 0.030 \text{ Pa})$.</p> <p>Differential pressure cells may be calibrated using digital communications protocols</p> <p>Calibrations may also be performed over an environmental temperature range of +2 °C to +8 °C with an uncertainty of ± 1 °C on the reported temperature.</p>	<p>UK & Site</p>



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
<p>PRESSURE (cont'd)</p> <p>Hydraulic pressure (gauge)</p> <p>Calibration of pressure measuring instruments and gauges. "Pressure equivalent" calibration of Dead Weight Testers (Pressure balance with associated mass set). Effective area calibration of Dead Weight Testers.</p>	<p>137 kPa to 200 kPa 200 kPa to 7 MPa 7 MPa to 172 MPa</p> <p>172 MPa to 345 MPa 345 MPa to 500 MPa</p>	<p>0.0040 % + 14 Pa 0.0035 % + 14 Pa 0.0035 % + 0.24 ppm/MPa 0.0080 % 0.015 %</p>		UK & Site
<p>Hydraulic pressure (absolute)</p> <p>Calibration of pressure measuring instruments and gauges.</p>	<p>200 kPa to 7 MPa 7 MPa to 172 MPa</p> <p>172 MPa to 345 MPa 345 MPa to 500 MPa</p>	<p>0.0040 % + 28 Pa 0.0035 % + 0.24 ppm/MPa + 15 Pa 0.0080 % 0.015 %</p>		
<p>Hydraulic pressure (differential)</p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>0 Pa to (172 - line pressure) MPa (Line pressure 1.7 MPa to 172 MPa)</p>	<p>0.000060 % of line pressure plus 0.0055 % of differential pressure plus 20 Pa</p>		
<p>MASS</p>	<p>Nominal value (g)</p> <p>26 000 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.002 0.001</p>	<p>(mg)</p> <p>26 20 10 5.0 2.0 1.0 0.50 0.20 0.10 0.060 0.050 0.040 0.032 0.024 0.020 0.016 0.012 0.010 0.0080 0.0060 0.0050 0.0040 0.0040 0.0040</p>	<p>Intermediate values can be calibrated with an uncertainty not less than that interpolated from the next higher and lower nominal value in the table.</p> <p>Calibrations can be given in other units as required.</p> <p>Calibration by substitution</p>	UK



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL				UK
All electrical measurements a carried out using the method of direct comparison or transfer to laboratory reference standards unless otherwise determined in the remarks column.				
DC Voltage	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1050 V	10 ppm + 0.5 μ V 10 ppm + 1.0 μ V 10 ppm + 10 μ V 10 ppm + 200 μ V 10 ppm + 2.0 mV	All electrical values can be sourced or measured by comparison unless otherwise stated	UK
DC Current	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 100 mA	20 ppm + 1.0 nA 20 ppm + 10 nA 20 ppm + 75 nA 40 ppm + 150 nA	Using nominal 10 Ω shunt	UK
DC Current	100 mA to 200 mA 200 mA to 2 A 2 A to 20 A	35 ppm + 0.70 μ A 250 ppm + 30 μ A 500 ppm + 1.0 mA	These values can be sourced	
DC Current	100 mA to 202 mA 202 mA to 2.02 A 2.02 A to 20 A	62 ppm + 5.5 μ A 90 ppm + 72 μ A 330 ppm + 8.0 mA	Simulation using multi turn coil	
DC Resistance	20 A to 1000 A	0.22 % + 100 mA		
DC Resistance	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 1 G Ω	15 ppm + 20 μ Ω 15 ppm + 20 μ Ω 15 ppm + 150 μ Ω 15 ppm + 1.0 m Ω 15 ppm + 15 m Ω 15 ppm + 100 m Ω 15 ppm + 1.5 Ω 20 ppm + 20 Ω 400 ppm + 500 Ω 0.35 %+ 12 k Ω		
AC VOLTAGE	1 mV to 200 mV 20 Hz to 55 Hz 55 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	150 ppm + 15 μ V 120 ppm + 15 μ V 350 ppm + 16 μ V 600 ppm + 20 μ V		UK
	200 mV to 2 V 20 Hz to 55 Hz 55 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 500 kHz	140 ppm + 40 μ V 120 ppm + 40 μ V 260 ppm + 40 μ V 350 ppm + 100 μ V 0.65 % + 15 mV		



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC VOLTAGE (cont'd)	2 V to 20 V 20 Hz to 55 Hz 55 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 500 kHz	150 ppm + 260 μ V 140 ppm + 260 μ V 260 ppm + 330 μ V 550 ppm + 1.2 mV 0.65 % + 120 mV		
	20 V to 200 V 20 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	140 ppm + 7.0 mV 260 ppm + 7.0 mV 600 ppm + 15 mV		
	200 V to 1 kV 45 Hz to 10 kHz 10 kHz to 30 kHz	200 ppm + 25 mV 380 ppm + 30 mV		
AC CURRENT	20 μ A to 200 μ A 55 Hz to 5 kHz 5 kHz to 10 kHz	0.050 % + 50 nA 0.060 % + 50 nA		
	200 μ A to 2 mA 55 Hz to 10 kHz	0.050 % + 500 nA		
	2 mA to 20 mA 55 Hz to 10 kHz	0.050 % + 5.0 μ A		
	20 mA to 200 mA 55 Hz to 10 kHz	0.050 % + 50 μ A		
	200 mA to 2 A 55 Hz to 1 kHz	0.060 % + 500 μ A		
	2 A to 20 A 55 Hz to 1 kHz	0.060 % + 4.0 mA		
AC CURRENT	25 μ A to 200 μ A 40 Hz to 45 Hz 45 Hz to 1 kHz	0.17 % + 410 nA 0.080 % + 390 nA	These values can be sourced	UK
	200 μ A to 2 mA 40 Hz to 45 Hz 45 Hz to 1 kHz	0.18 % + 1.0 μ A 0.075 % + 0.70 μ A		
	2 mA to 20 mA 40 Hz to 45 Hz 45 Hz to 1 kHz	0.18 % + 1.1 μ A 0.073 % + 7.4 μ A		
	20 mA to 200 mA 40 Hz to 45 Hz 45 Hz to 1 kHz	0.18 % + 120 μ A 0.077 % + 86 μ A		
	200 mA to 2 A 40 Hz to 45 Hz 45 Hz to 1 kHz	0.18 % + 1.1 mA 0.085 % + 770 μ A		



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC CURRENT (Cont'd)	2 A to 20 A 40 Hz to 45 Hz 45 Hz to 100 Hz	0.16 % + 11 mA 0.037 % + 6.6 mA		UK
	20 A to 100 A at 50 Hz	0.22 % + 100 mA	Simulation using a multi turn coil	
AC RESISTANCE	100 A to 1000 A at 50 Hz	0.22 % + 400 mA		
55 Hz to 1 kHz	30 $\mu\Omega$ to 10 m Ω 10 m Ω to 100 m Ω 100 m Ω to 1 Ω 1 Ω to 10 10 Ω to 100 Ω 100 Ω to 1 k Ω	26 $\mu\Omega$ 0.26 % 0.26 % 0.16 % 0.16 % 0.16 %		
CAPACITANCE	1 nF 10 nF 20 nF 50 nF 100 nF 1 μ F 10 μ F	29 pF 61 pF 99 pF 220 pF 370 pF 5.1 nF 78 nF	For the calibration of measuring devices	
FREQUENCY	10 MHz Clock frequency 10 mHz to 80 MHz	1.0 parts in 10^8 5.0 parts in 10^8	Frequency may also be expressed in terms of time; $1/f$, for repetitive signals or in other units such as revolutions per minute.	UK
	1 mHz to 80 MHz	5.0 parts in 10^8 + 5.0 μ Hz	Calibration of measuring devices Calibration of sources	
TIME INTERVAL	0 s to 1 day	100 ms	Manually triggered single events.	UK
RPM	60 RPM to 60000 RPM	50 ppm + 0.01 RPM	Generate	UK



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
OSCILLOSCOPES				UK
Vertical deflection coefficients				
DC	30 mV to 300 mV 300 mV to 120 V	1.1 % 0.30 %	Square-wave & DC signals appropriate for the calibration of oscilloscope vertical deflection coefficients	
Peak to Peak Voltage 1 kHz	30 mV to 300 mV 300 mV to 6 V	1.3 % 0.70 %		
Horizontal deflection coefficients				UK
Time	10 ns to 1 s	0.10 %	Pulse markers The uncertainties quoted above are based on the readout resolution of typical oscilloscopes.	
ELECTRICAL SIMULATION OF TEMPERATURE				
Thermocouple capabilities listed below are given for type T Base and Type S Noble, using EMF sensitivity values as listed in BS EN 60584-1:2013. Other Thermocouple types can be calibrated, the uncertainties will correspond to the appropriate sensitivities listed. Calibrations which include the internal reference junction (CJC) are available for types: J, K, N, T, E, R, S, B & C				
Base Metal Thermocouples Noble Metal Thermocouples	-200 °C to +1400 °C 0 °C to 500 °C 500 °C to 1800 °C	0.050 °C 0.080 °C 0.050 °C	Excluding automatic CJC	
Base Metal Thermocouples Noble Metal Thermocouples	-200 °C to +1400 °C 0 °C to 500 °C 500 °C to 1800 °C	0.17 °C 0.26 °C 0.25 °C	Including automatic CJC	
Cold Junction Compensation	0 °C to 30 °C	0.15 °C		
Resistance thermometers by simulation				
Values below are based upon a PT100. Other resistance thermometer types, including thermocouples with a defined resistance scale can also be calibrated based on the resistance capabilities listed above.				
Resistance thermometer simulation	-200 °C to 830 °C	0.0050 °C		



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Temperature indicators and recorders, with temperature sensor(s)	0.01 °C (Water Triple Point) -95 °C to 140 °C 140 °C to 150 °C 150 °C to 660 °C	0.0050 °C 0.055 °C 0.080 0.10 °C	Fixed point Calibrations within both metal and liquid media	UK
Resistance thermometers	0.01 °C (Water Triple Point) -95 °C to 140 °C 140 °C to 150 °C 150 °C to 660 °C	0.0050 °C 0.055 °C 0.080 °C 0.10 °C	Calibrations within both metal and liquid media	
Thermocouples Base Metal	-95 °C to 0 °C 0 °C to 30 °C 30 °C to 660 °C	0.40 °C 0.10 °C 0.40 °C	Calibrations within both metal and liquid media	
Noble Metal Type R and S Type B	0 °C to 660 °C 0 °C to 660 °C	0.40 °C 0.70 °C	Calibrations within both metal and liquid media	
Metal Block Calibrators and portable liquid baths	-100 °C to 250 °C 250 °C to 660 °C	0.050 °C 0.13 °C	Includes axial, radial and stability information	
LENGTH			NOTES	UK
Orifice plates	BS EN ISO 5167-2:2003 Bore diameter 10 to 700	8	RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED	
Micrometers, External	BS 870:2008 0 to 125	Heads:2.0 between any two points	The uncertainty quoted is for the departure from either flatness, straightness, parallelism, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration.	
Calliper gauges (inc. Vernier, dial and digital)	BS 887:2008 0 to 150	Overall performance 20		
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 100	1.0		
Length gauges, flat and spherical ended	0 to 100 BS 870:2008	1.0 + (8.0 x length in m)		
ANGLE				
Squares, blade type	BS 939:2007 50 to 300	3.0 (see note)		
Spirit levels	BS 3509:1962 and BS 958:1968 5 seconds of arc to 60 minutes of arc nominal sensitivity	Mean sensitivity 10 % of nominal Minimum 0.50 seconds of arc		



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
MASS	Nominal value (g)	(mg)		IRE
	20 000	10	Intermediate values can be calibrated with an uncertainty not less than that interpolated from the next higher and lower nominal value in the table. Calibrations can be given in other units as required. Calibration by substitution	
	10 000	5.3		
	5 000	2.7		
	2 000	1.0		
	1 000	0.53		
	500	0.27		
	200	0.10		
	100	0.053		
	50	0.033		
	20	0.027		
	10	0.020		
	5	0.017		
	2	0.013		
	1	0.010		
	0.5	0.0083		
	0.2	0.0067		
	0.1	0.0053		
	0.05	0.0040		
	0.02	0.0033		
0.01	0.0027			
0.005	0.0020			
0.002	0.0020			
0.001	0.0020			
ELECTRICAL DC VOLTAGE	0 mV to 200 mV	10 ppm + 1.5 μ V		IRE
	200 mV to 2 V	10 ppm + 1.5 μ V		
	2 V to 20 V	10 ppm + 10 μ V		
	20 V to 200 V	10 ppm + 200 μ V		
	200 V to 1 kV	10 ppm + 2.0 mV		
DC CURRENT	0 μ A to 200 μ A	20 ppm + 1 nA	Using nominal 10 Ω shunt	IRE
	200 μ A to 2 mA	20 ppm + 10 nA		
	2 mA to 20 mA	20 ppm + 75 nA		
	20 mA to 40 mA	50 ppm + 150 nA		
	40 mA to 200 mA	35 ppm + 700 nA		
	200 mA to 2 A	250 ppm + 30 μ A		
	2 A to 20 A	500 ppm + 1.0 mA		
100 mA to 202 mA	130 ppm + 6.6 μ A	Calibration of measuring devices by comparison		
202 mA to 2.02 A	150 ppm + 180 μ A			
2.02 A to 20 A	420 ppm + 1.5 mA			
	20 A to 1500 A	0.22 % + 100 mA	Simulation using a multi turn coil	



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
DC RESISTANCE	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 1 G Ω	15 ppm + 20 $\mu\Omega$ 15 ppm + 20 $\mu\Omega$ 15 ppm + 150 $\mu\Omega$ 15 ppm + 1.0 m Ω 15 ppm + 15 m Ω 15 ppm + 100 m Ω 15 ppm + 1.5 Ω 20 ppm + 20 Ω 400 ppm + 500 Ω 0.35 % + 12 k Ω		IRE
AC VOLTAGE	1 mV to 200 mV 20 Hz to 55 Hz 55 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 200 mV to 2 V 20 Hz to 55 Hz 55 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 500 kHz	150 ppm + 15 μV 120 ppm + 15 μV 350 ppm + 16 μV 600 ppm + 20 μV 140 ppm + 40 μV 120 ppm + 40 μV 260 ppm + 40 μV 350 ppm + 100 μV 0.65 % + 15 mV	AC Values can be sourced or measured by comparison up to 1 kHz, above that frequency is for measurement only.	IRE
AC VOLTAGE	2 V to 20 V 20 Hz to 55 Hz 55 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 500 kHz 20 V to 200 V 20 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 200 V to 1 kV 50 Hz to 10 kHz 10 kHz to 30 kHz	150 ppm + 260 μV 140 ppm + 260 μV 260 ppm + 330 μV 550 ppm + 1.2 mV 0.75 % + 120 mV 140 ppm + 7 mV 260 ppm + 7 mV 600 ppm + 15 mV 200 ppm + 25 mV 380 ppm + 30 mV		IRE
AC CURRENT	10 μA to 200 μA 55 Hz to 5 kHz 5 kHz to 10 kHz 200 μA to 2 mA 55 Hz to 10 kHz 2 mA to 20 mA 55 Hz to 10 kHz 20 mA to 200 mA 55 Hz to 10 kHz	0.050 % + 50 nA 0.060 % + 50 nA 0.050 % + 500 nA 0.050 % + 5.0 μA 0.050 % + 50 μA		IRE



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
AC CURRENT (cont'd)	200 mA to 2 A 55 Hz to 10 kHz	0.085 % + 500 μ A	Calibration of measuring devices by comparison	IRE
	2 A to 20 A 55 Hz to 5 kHz	0.20 % + 5.0 mA		
	25 μ A to 202 μ A 40 Hz to 45 Hz 45 Hz to 999 Hz	0.28 % + 420 nA 0.099 % + 390 nA		
	202 μ A to 2.02 mA 40 Hz to 45 Hz 45 Hz to 999 Hz	0.22 % + 1.2 μ A 0.094 % + 0.80 μ A		
	2.02 mA to 20.2 mA 40 Hz to 45 Hz 45 Hz to 999 Hz	0.23 % + 12 μ A 0.094 % + 7.9 μ A		
	20.2 mA to 202 mA 40 Hz to 45 Hz 45 Hz to 999 Hz	0.22 % + 120 μ A 0.94 % + 90 μ A		
	202 mA to 2.02 A 40 Hz to 45 Hz 45 Hz to 999 Hz	0.25 % + 1.2 mA 0.11 % + 0.11 mA		
	2.02 A to 20 A 40 Hz to 45 Hz 45 Hz to 999 Hz	0.34 % + 13 mA 0.073 % + 4.4 mA		
CAPACITANCE	20 A to 100 A, 40 Hz to 60 Hz	0.25 % + 100 mA	Simulation using a multi turn coil	IRE
	100 A to 1500 A, 40 Hz to 60 Hz	0.25 % + 400 mA		
FREQUENCY	1 nF	29 pF	Calibration of measuring devices by comparison	IRE
	10 nF	58 pF		
	20 nF	92 pF		
	50 nF	190 pF		
	100 nF	360 pF		
	1 μ F	5.1 nF		
	10 μ F	74 nF		
	100 μ F	840 nF		
	1 mF	13 μ F		
	10 mF	130 μ F		
	0.01 Hz to 50 MHz	5.0 parts in 10^8	Frequency may also be expressed in terms of time; 1/f, for repetitive signals or in other units such as revolutions per minute.	IRE



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
RPM (Revolutions per minute)	2 rpm to 10 rpm 10 rpm to 100 rpm 100 rpm to 1000 rpm 1 000 rpm to 10 000 rpm 10 000 rpm to 100 000 rpm	10 ppm + 0.00050 rpm 10 ppm + 0.0020 rpm 10 ppm + 0.020 rpm 10 ppm + 0.20 rpm 10 ppm + 2.0 ppm		IRE
TIME INTERVAL	Longer than 100 ms Longer than 80 ns	2.0 ppm + 20 ms 2.0 ppm + 80 ns	Manually triggered single events. Electronically triggered single events	IRE IRE
RCD				
Trip current	2 mA to 3 A 20 ms to 190 ms	5.8 % + 240 μ A		IRE
	2 mA to 3 A 190 ms to 5 s	1.4 % + 80 μ A		
Trip time	20 ms to 400 ms 400 ms to 5 s	1.0 ms 10 ms		
AC resistance for Loop 50 Hz			Laboratory loop 0.20 Ω	IRE
Nominal Ranges	0.2 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω	0.6% + 4.8 m Ω 0.6% + 19 m Ω 0.6% + 36 m Ω		
Earth Bond Resistance	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω	0.60 % + 4.8 m Ω 0.60 % + 19 m Ω 0.60 % + 36m Ω		IRE
Earth bond current 50 Hz	10 mA to 500 mA 100 mA to 10 A 10 A to 30 A	1.8 % + 7.0 mA 1.8 % + 70 mA 1.8 % + 70 mA		
Load	0.13 kVA	6.0 %		
Leakage Current At nominal 240 V 50 Hz	2 mA to 8 mA	1.8 % + 11 μ A		IRE
Insulation Test Voltage	50 V to 1000 V	1.2 % + 950 mV		
Insulation Resistance	10 k Ω to 100 k Ω 101 k Ω to 1 M Ω 1.01 M Ω to 10 M Ω 10.1 M Ω to 100 M Ω 101 M Ω to 1 G Ω 1.01 G Ω to 10 G Ω	0.12 % + 200 m Ω 0.12 % 1.2 % 1.2 % 1.4 % 7.0 %		
AC Voltage				
Nominal 50 Hz	100 V to 400 V	0.25 % + 160 mV		IRE



0822
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
Continuity Resistance	20 mΩ to 1000 Ω	0.30 % + 30 mΩ		IRE
Continuity Current At a nominal 1 Ω	10 mA to 300 mA	1.6 % + 0.80 mA		
<p>ELECTRICAL SIMULATION OF TEMPERATURE</p> <p>Thermocouple capabilities listed below are given for type T Base and Type C Noble, using EMF sensitivity values as listed in BS EN 60584-1:2013. Other Thermocouple types can be calibrated, the uncertainties will correspond to the appropriate sensitivities listed. Calibrations which include the internal reference junction (CJC) are available for types: J, K, N, T, E, R, S, B & C</p>				
Base Metal Thermocouples	-200 °C to +1400 °C	0.10 °C	Excluding automatic CJC	IRE
Noble Metal Thermocouples	0 °C to 2315 °C	0.10 °C		
Base Metal Thermocouples	-200 °C to +1400 °C	0.25 °C	Including automatic CJC	
Noble Metal Thermocouples	500 °C to 1800 °C	0.26 °C		
Cold Junction Compensation	0 °C to 30 °C	0.15 °C		
<p>Resistance thermometers by simulation</p> <p>Values below are based upon a PT100. Other resistance thermometer types, including thermocouples with a defined resistance scale can also be calibrated based on the resistance capabilities listed above.</p>				
Resistance thermometer simulation	-200 °C to 840 °C	0.0050 °C		



0822

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited
Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
<p>TEMPERATURE</p> <p>Temperature indicators and recorders, with temperature sensor(s)</p> <p>Platinum Resistance Thermometers (4 wire)</p> <p>Metal Block Calibrators and portable liquid baths</p>	<p>0.01 °C (Water Triple Point) -196 °C (LN2) -95 °C to -80 °C -80 °C to +300 °C 300 °C to 450 °C 450 °C to 650 °C</p> <p>0.01 °C (Water Triple Point) -196 °C (LN2) -95 °C to -80 °C -80 °C to +300 °C 300 °C to 450 °C 450 °C to 650 °C</p> <p>-100 °C to +100 °C 100 °C to 300 °C 300 °C to 420 °C 420 °C to 650 °C 0 °C</p>	<p>0.0030 °C 0.015 °C 0.025 °C 0.015 °C 0.027 °C 0.094 °C</p> <p>0.0030 °C 0.015 °C 0.025 °C 0.015 °C 0.027 °C 0.094 °C</p> <p>0.030 °C 0.038 °C 0.15 °C 0.16 °C 0.020 °C</p>	<p>In a range of liquid and metal media baths</p> <p>In a range of liquid and metal media baths</p> <p>Suitable zero reference baths</p>	IRE
<p>HUMIDITY</p> <p>Dew-point</p> <p>Temperature sensors in air</p> <p>Relative humidity</p>	<p>-25 °C to +60 °C</p> <p>0 °C to 60 °C</p> <p>Example conditions</p> <p>At 0 °C: 10 %rh to 90 %rh</p> <p>At 23 °C: 5 %rh to 95 %rh</p> <p>At 60 °C: 5 %rh to 90 %rh</p>	<p>0.17 °C</p> <p>0.10 °C</p> <p>Corresponding to above dew-point and temperature uncertainties</p> <p>0.20 %rh to 1.1 %rh</p> <p>0.20 %rh to 1.1 %rh</p> <p>0.20 %rh to 0.80 %rh</p>	<p>Calibrations undertaken in an air chamber</p>	IRE
END				



0822

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Chamois Metrology Limited

Issue No: 080 Issue date: 21 August 2020

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %-p + (0.12 · 10⁻⁶ · p · 10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.