


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> CALIBRATION  0832  Accredited to ISO/IEC 17025:2017	<b>MDS Aero Support Ltd</b>  Issue No: 016    Issue date: 13 April 2021	
	37 Longbridge Lane Off Ascot Drive Derby DE24 8UJ	Contact: Mr Tim Rennie Tel: +44 (0)1332 754180 Fax: +44 (0)1332 753531 E-Mail: <a href="mailto:timothy.rennie@mdsaero.co.uk">timothy.rennie@mdsaero.co.uk</a> Website: <a href="http://www.mdsaero.com">www.mdsaero.com</a>

**Calibration performed by the Organisations at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 37 Longbridge Lane Off Ascot Drive Derby DE24 8UJ  <b>Local contact</b> Mr Tim Rennie Tel: +44 (0)1332 754180 Fax: +44 (0)1332 753531 Email: <a href="mailto:timothy.rennie@mdsaero.co.uk">timothy.rennie@mdsaero.co.uk</a> Website: <a href="http://www.mdsaero.com">www.mdsaero.com</a>	Electrical Pressure	Lab

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Electrical	Site



0832  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**MDS Aero Support Ltd**  
Issue No: 016 Issue date: 13 April 2021

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>ELECTRICAL</b>			Electrical calibrations are performed by comparison with a reference standard measuring the output of instruments or generating the output, as stated.
DC Voltage			
Generation	0 V to 100 mV 100 mV to 10 V	0.022 mV 2.0 mV	Lab & Site
Measurement	0 V to 100 mV	10 ppm + 2.0 $\mu$ V	Lab
	0 V to 200 mV 200 mV to 2 V 2 V to 20 V	110 ppm + 10 $\mu$ V 100 ppm + 20 $\mu$ V 100 ppm + 170 $\mu$ V	Site
DC Resistance			
Measurement	0 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 1 k $\Omega$ to 100 k $\Omega$ 100 k $\Omega$ to 1 M $\Omega$	20 ppm + 0.60 m $\Omega$ 20 ppm + 0.60 m $\Omega$ 20 ppm + 2.0 $\Omega$ 20 ppm + 10 $\Omega$	Lab
Generation	100 $\Omega$ 150 $\Omega$ 2 k $\Omega$ 12 k $\Omega$ 30 k $\Omega$ 200 k $\Omega$	14 m $\Omega$ 24 m $\Omega$ 0.18 $\Omega$ 0.60 $\Omega$ 3.8 $\Omega$ 5.5 $\Omega$	Lab
<b>PRESSURE</b>			
<u>Gas Pressure (Gauge)</u>			Methods consistent with EURAMET CG17.
Calibration of pressure indicating instruments	-70 kPa to -35 kPa -35 kPa to 35 kPa 35 kPa to 100 kPa 100 kPa to 350 kPa 350 kPa to 850 kPa 850 kPa to 1.3 MPa 1.3 MPa to 2.2 MPa 2.2 MPa to 4.5 MPa 4.5 MPa to 7 MPa	19 Pa 15 Pa 26 Pa 86 Pa 0.13 kPa 0.17 kPa 0.42 kPa 0.56 kPa 0.80 kPa	Lab
END			



0832  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**MDS Aero Support Ltd**  
Issue No: 016 Issue date: 13 April 2021

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
  - As an explicit function of the measurand or of a parameter (see below).
  - As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
  - As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

**Expression of CMCs - symbols and units**

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10<sup>-6</sup>·p·10<sup>-6</sup>) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.