

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0834</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Bellingham & Stanley Limited</p> <p>Issue No: 013 Issue date: 26 August 2021</p>	
	<p>Longfield Road Tunbridge Wells Kent TN2 3EY</p>	<p>Contact: Mr Andrew Darkins Tel: +44 (0)1892 500400 Fax: +44 (0)1892 543115 E-Mail: sales.bs.uk@xylem.com Website: www.bellinghamandstanley.com</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
POLARIMETERS			
Optical rotation	- 34°Z to 105°Z	0.017°Z	Using a digital polarimeter Values at 20.0°C and 589.44nm
	546 nm to 900 nm between 17 °C and 30 °C	0.017°Z	Values for Quartz Control Plates calculated using ICUMSA equations 3 and 6 from SPS-1 (2007) and ICUMSA 22nd Session, Berlin, 1998
REFRACTOMETERS			
Refractive index Calibration samples (RI)	Oils and Solid Plates		Using high accuracy refractometer Values at 20.0 °C and 589.3 nm
	1.30 to 1.54	0.000030	
	1.54 to 1.62	0.00010	
Calibration samples (°Brix)	a) Sucrose		Using gravimetric method
	0 to 30	0.010	
	30 to 60	0.020	
	b) AG Fluids		By comparison
	0 to 40	0.020	

END



0834
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation (DRAFT)
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Bellingham + Stanley Ltd
Issue No: 013 **Issue date:** 26 August 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$