


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|---|--|---|
|  <p>Accredited to ISO/IEC 17025:2005</p> | MCS Test Equipment Limited | |
| | Issue No: 001 | Issue date: 21 September 2018 |
| | Unit 8 New Vision Business Park Glascoed Road St Asaph LL17 0LP | Contact: Mr Alan Horner Tel: +44 (0)1492 550 398 E-Mail: alan.horner@mcs-testequipment.com Website: https://mcs-testequipment.com/ |
| Calibration performed at the above address only | | |

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|--|--|--|---------|
| DC Voltage | | | |
| Generation | 0 V to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V | 30 ppm + 2.1 μ V 20 ppm + 7.0 μ V 20 ppm + 120 μ V 30 ppm + 1.0 mV 30 ppm + 4.0 mV | |
| Measurement | 0 V to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V | 10 ppm + 2.1 μ V 18 ppm + 7.0 μ V 25 ppm + 120 μ V 55 ppm + 1.0 mV 70 ppm + 4.0 mV | |
| Resistance | | | |
| Generation | 0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 k Ω 1.1 k Ω to 3.3 k Ω 3.3 k Ω to 11 k Ω 11 k Ω to 33 k Ω 33 k Ω to 110 k Ω 110 k Ω to 330 k Ω 330 k Ω to 1.1 M Ω 1.1 M Ω to 3.3 M Ω 3.3 M Ω to 11 M Ω 11 M Ω to 33 M Ω 33 M Ω to 110 M Ω | 100 ppm + 6.0 m Ω 100 ppm + 40 m Ω 100 ppm + 40 m Ω 100 ppm + 200 m Ω 100 ppm + 500 m Ω 100 ppm + 700 m Ω 100 ppm + 500 m Ω 100 ppm + 2.0 Ω 100 ppm + 20 Ω 100 ppm + 5.0 k Ω 100 ppm + 2.1 k Ω 100 ppm + 6.0 k Ω 200 ppm + 3.0 k Ω 400 ppm + 7.5 k Ω 700 ppm + 20 k Ω | |
| Measurement | 0 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω | 150 ppm + 2.1 m Ω 150 ppm + 690 m Ω 150 ppm + 7.0 Ω 150 ppm + 120 Ω 150 ppm + 1.2 k Ω 550 ppm + 12 k Ω 0.27 % + 350 k Ω | |



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

MCS Test Equipment Limited

Issue No: 001 Issue date: 21 September 2018

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2) | Remarks |
|--|---|---|---------|
| DC Current | | | |
| Generation | 0 A to 330 μ A 330 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 3.3 A 3.3 A to 11 A | 200 ppm + 30 nA 200 ppm + 70 nA 200 ppm + 300 nA 200 ppm + 10 μ A 500 ppm + 140 μ A 700 ppm + 2.0 mA | |
| Measurement | 0 A to 20 mA 20 mA to 100 mA 100 mA to 1 A 1 A to 3 A | 770 ppm + 2.5 μ A 470 ppm + 95 μ A 0.12 % + 500 μ A 0.14 % + 2.5 mA | |
| AC Voltage | | | |
| Generation | <i>10 Hz to 45 Hz</i> 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V | 0.040 % + 27 μ V 0.050 % + 70 μ V 0.050 % + 3.5 mV | |
| | <i>45 Hz to 1 kHz</i> 30 μ V to 33 mV 33 V to 330 V 330 V to 1 kV | 0.20 % + 8.0 μ V 0.030 % + 17 mV 0.040 % + 40 mV | |
| | <i>45 Hz to 10 kHz</i> 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V | 0.020 % + 12 μ V 0.020 % + 130 μ V 0.020 % + 5.0 mV | |
| | <i>1 kHz to 10 kHz</i> 33 V to 330 V | 0.030 % + 13 mV | |
| | <i>10 kHz to 20 kHz</i> 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V | 0.020 % + 12 μ V 0.030 % + 130 μ V 0.040 % + 2.0 mV 0.030 % + 27 mV | |
| | <i>20 kHz to 50 kHz</i> 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V | 0.050 % + 16 μ V 0.040 % + 130 μ V 0.050 % + 2.0 mV 0.040 % + 27 mV | |
| | <i>50 kHz to 100 kHz</i> 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V | 0.10 % + 38 μ V 0.10 % + 3.5 mV 0.20 % + 4.0 mV 0.30 % + 95 mV | |



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

MCS Test Equipment Limited

Issue No: 001 Issue date: 21 September 2018

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|--|---|---|---------|
| AC Voltage (continued) Generation (continued) Measurement | <i>100 kHz to 500 kHz</i> 33 mV to 330 mV 330 mV to 3.3 V | 0.30 % + 5.0 mV 0.30 % + 35 mV | |
| | <i>10 Hz to 20 kHz</i> 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V | 0.10 % + 37 μ V 0.10 % + 110 μ V 0.10 % + 12 mV 0.10 % + 37 mV | |
| | <i>10 Hz to 1 kHz</i> 100 V to 750 V | 0.10 % + 88 mV | |
| | <i>20 kHz to 50 kHz</i> 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V | 0.20 % + 72 μ V 0.20 % + 600 μ V 0.20 % + 6.5 mV 0.20 % + 65 mV | |
| | <i>50 kHz to 100 kHz</i> 100 mV to 1 V 1 V to 10 V | 1.0 % + 990 μ V 0.95 % + 9.5 mV | |
| | <i>100 kHz to 300 kHz</i> 1 V to 10 V | 6.2 % + 65 mV | |
| AC Current Generation | <i>45 Hz to 1 kHz</i> 30 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3.3 A | 0.20 % + 800 nA 0.070 % + 4.5 μ A 0.070 % + 5.5 mA 0.080 % + 16 mA 0.80 % + 80 mA | |
| | <i>1 kHz to 5 kHz</i> 30 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3.3 A | 0.35 % + 3.5 μ A 0.15 % + 8.0 μ A 0.20 % + 150 μ A 1.0 % + 5.5 mA 1.0 % + 45 mA | |
| | <i>5 kHz to 10 kHz</i> 30 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3.3 A | 0.80 % + 3.5 μ A 0.35 % + 8.5 μ A 0.35 % + 2.1 mA 4.0 % + 15 mA 4.0 % + 70 mA | |



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

MCS Test Equipment Limited

Issue No: 001 Issue date: 21 September 2018

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|---|---|---|---------|
| AC Current (continued) Generation (continued) | <i>10 kHz to 30 kHz</i> 30 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA | 1.6 % + 4.5 μ A 0.65 % + 9.5 μ A 0.65 % + 350 μ A | |
| Measurement | 3.3 A to 11 A <i>45 Hz to 100 Hz</i> <i>100 Hz to 1 kHz</i> <i>1 kHz to 5 kHz</i> | 0.10 % + 3.0 mA 2.0 % + 3.0 mA 5.0 % + 3.0 mA | |
| | <i>10 Hz to 3 kHz</i> 100 mA to 1 A 1 A to 3 A | 0.040 % + 27 mA 0.050 % + 70 mA | |
| Capacitance | <i>3 kHz to 5 kHz</i> 1 A to 3 A | 0.30 % + 210 mA | |
| Sourcing by simulation | 1.1 nF to 3.3 nF 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF 330 nF to 1.1 μ F 1.1 μ F to 3.3 μ F 3.3 μ F to 11 μ F 11 μ F to 33 μ F 33 μ F to 110 μ F 110 μ F to 330 μ F 330 μ F to 1.1 mF 1.1 mF to 3.3 mF 3.3 mF to 11 mF 11 mF to 33 mF | 0.60 % + 25 pF 0.30 % + 40 pF 0.30 % + 160 pF 0.30 % + 370 pF 0.30 % + 1.0 nF 0.30 % + 750 pF 0.30 % + 11 nF 0.30 % + 40 pF 0.50 % + 150 pF 0.55 % + 600 nF 0.55 % + 2.5 μ F 0.55 % + 8.5 μ F 0.55 % + 4.5 μ F 0.55 % + 60 μ F 0.90 % + 500 μ F | |
| Frequency | 10 MHz | 6.0 in 10^{12} | |
| | 1 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 160 MHz | 22 ppm 14 ppm 150 ppm 3.8 in 10^8 3.5 in 10^9 1.0 in 10^{10} | |
| | 160 MHz to 1.3 GHz 1.3 GHz to 10 GHz 10 GHz to 18 GHz | 3.0 in 10^{10} 7.5 in 10^{10} 1.4 in 10^9 | |



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

MCS Test Equipment Limited

Issue No: 001 Issue date: 21 September 2018

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|--|---|---|---------|
| RF Measurements | | | |
| Power | | | |
| Ranges | <i>9 kHz to 6 GHz</i> 200 pW to 40 μ W 20 nW to 4 mW 2 μ W to 200 mW <i>10 MHz to 18 GHz</i> 2 nW to 40 μ W 200 nW to 40 mW 2 μ W to 2 W | 3.0 % 2.7 % 2.8 % 9.5 % 4.5 % 5.0 % | |
| Calibration Factor | 0 % to 100 % 9 kHz 18 GHz | 2.0 % | |
| END | | | |



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

MCS Test Equipment Limited

Issue No: 001 Issue date: 21 September 2018

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
 - As an explicit function of the measurand or of a parameter (see below).
 - As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
 - As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V: $0.0025 \% + 5.0 \mu\text{V}$:

Over the range 100 mV to 1 V, the CMC is $0.0025 \% \cdot V + 5.0 \mu\text{V}$, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: $0.0036 \% + 0.12 \text{ ppm/MPa} + 4.0 \text{ Pa}$

Over the range 0.5 MPa to 140 MPa, the CMC is $0.0036 \% \cdot p + (0.12 \cdot 10^{-6} \cdot p \cdot 10^{-6}) + 4.0 \text{ Pa}$, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means $1.5 \cdot 0.01 \cdot i$, where i is the instrument indication.