


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> CALIBRATION  <b>10723</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>JTF Wireless Limited</b>	
	Issue No: 002    Issue date: 23 August 2021	
	Unit 9 Galena Close Amington Tamworth B77 4AS	Contact: Mr Jake Farrow Tel: +44 (0) 121 794 8092 E-Mail: <a href="mailto:jake@jtfwireless.com">jake@jtfwireless.com</a> Website: <a href="http://www.jtfwireless.com">www.jtfwireless.com</a>
Calibration performed by the Organisation at the locations specified		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Unit 9 Galena Close Amington Tamworth B77 4AS  <b>Local contact</b> Jake Farrow  Tel: +44 (0) 121 794 8092 E-Mail: <a href="mailto:jake@jtfwireless.com">jake@jtfwireless.com</a>	Temperature	P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Any Customer Premises  The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer  <b>Local contact</b> Jake Farrow  Tel: +44 (0) 121 794 8092 E-Mail: <a href="mailto:jake@jtfwireless.com">jake@jtfwireless.com</a>	Temperature	S



10723

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**JTF Wireless Limited**  
**Issue No:** 002 **Issue date:** 23 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
TEMPERATURE Temperature indicators and recorders with temperature sensor(s)				
Dry Block Bath	-90 °C to +135 °C	0.10 °C	By comparison with reference standards	P & S
Calibration in air chamber	-35 °C to +40 °C	0.38 °C	By comparison with reference standards	P & S
END				



10723

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

#### JTF Wireless Limited

Issue No: 002 Issue date: 23 August 2021

Calibration performed by the Organisation at the locations specified

### Appendix - Calibration and Measurement Capabilities

#### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$