

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> CALIBRATION  10843  Accredited to ISO/IEC 17025:2017	<b>seca Limited</b>	
	Issue No: 002    Issue date: 23 August 2021	
	40 Barn Street Birmingham B5 5QB United Kingdom	Contact: Simon Ford Tel: +44 (0) 121 643 9349 E-Mail: sales@seca.co.uk Website: www.seca.co.uk
Calibration performed by the Organisations at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 40 Barn Street Birmingham B5 5QB England	<b>Local contact</b> Simon Ford	Calibration of Non Automatic Weighing Machine  P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
At Customers Premises  The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Calibration Engineer  Calibration of Non Automatic Weighing Machine	S



10843  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**seca Limited**

**Issue No: 002    Issue date: 23 August 2021**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
WEIGHING INSTRUMENTS  Non-automatic weighing machines (Electronic self indicating)	100 g 200 g 500 g  1 kg 2 kg 5 kg 10 kg 20 kg	0.30 mg 0.53 mg 1.4 mg  2.7 mg 5.4 mg 39 mg 78 mg 160 mg	1. Weights are available in OIML Class:  E2 from 1 mg to 1 kg, Max. grouped load 2.1 kg  F1 from 1 g to 10 kg, Max. grouped load 20 kg  2. Other loads within the overall listed range may also be used.  3. Method based on the requirements of Euramet guide cg-18	S
END				



10843  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**seca Limited**  
**Issue No:** 002    **Issue date:** 23 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$