

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS CALIBRATION 1893 Accredited to ISO/IEC 17025:2017	Technological Laboratory of Uruguay	
	Issue No: 003 Issue date: 25 August 2021	
	Av. Italia 6201 C.P. 11500 Montevideo Uruguay	Contact: Ms Monica Trias Tel: +598 2 601 3724 Int 1228 E-Mail: mtrias@latu.org.uy Website: www.latu.org.uy
Calibration performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Av. Italia 6201 C.P. 11500 Montevideo Uruguay	Local contact Ms Monica Trias Tel: +598 26013724 int 1228 Email: mtrias@latu.org.uy Website: www.latu.org.uy	<u>Site Calibration:</u> Temperature profiling Lab UPT

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer premises where ovens are situated, managed from Montevideo	Oven qualification and performance qualification (performance testing)	Site (UPT)



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Technological Laboratory of Uruguay
Issue No: 003 Issue date: 25 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				Lab + Site
Temperature surveys			Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	
Ovens, incubators, environmental chambers, fridges and freezers	-35 °C to 0 °C 0 °C to 80 °C	0.12 °C 0.14 °C	Calibration by comparison	
	80 °C to 200 °C 200 °C to 300 °C	0.13 °C 0.45 °C	with reference thermocouples	
	-30 °C to 0 °C 0 °C to 80 °C 80 °C to 135 °C	0.08 °C 0.10 °C 0.08 °C	Calibration by comparison With reference PRTs	
Static Sterilisers	100 °C to 135 °C	0.08 °C	Calibration by comparison With reference PRTs	
END				



1893

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Technological Laboratory of Uruguay
Issue No: 003 **Issue date:** 25 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$