


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 20765 Accredited to ISO/IEC 17025:2017	The Equipment Calibration Business Ltd	
	Issue No: 006 Issue date: 01 December 2023	
	1A Golf Link Villas Downley Common High Wycombe Buckinghamshire HP13 5YH United Kingdom	Contact: Mr Richard Harper Tel: +44 (0) 1491 822691 E-Mail: ecb@equipment-calibration.com Website: www.equipment-calibration.com
Calibration performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address The Equipment Calibration Business The Annex W-17 Howbery Business Park Wallingford OX10 8BA United Kingdom Local contact Mr Richard Harper Tel: +44 (0)1491 822691 E-Mail: Richard-harper@equipment-calibration.com Website: www.equipment-calibration.com	Electrical calibration: Calibration of EMC- related test and measurement equipment.	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Address The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. Local contact Mr Richard Harper Tel: +44 (0)1491 822691 Mobile: +44 (0)7380 950803 E-Mail: Richard-harper@equipment-calibration.com Website: www.equipment-calibration.com	Electrical calibration: Calibration of EMC-related test and measurement equipment.	S

NOTE: Where EN Standards have exact equivalents in IEC, or BS EN Standards, these are also included in the accreditation.



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL CALIBRATION				
All electrical measurements are carried out using the method of direct comparison or transfer to laboratory reference standards unless otherwise determined in the remarks column. The measurement and generation headings in the first column declare the laboratory's ability to either measure outputs of submitted test items or to generate values as a stimulus for test items which measure.				
DC RESISTANCE Measurement	0 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω	0.010 % + 4.0 m Ω 0.010 % + 10 m Ω 0.010 % + 0.10 Ω 0.010 % + 1.0 Ω 0.010 % + 10 Ω 0.040 % + 0.10 k Ω 0.80 % + 10 k Ω		P and S
DC VOLTAGE Measurement	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	50 μ V/V + 3.5 μ V 40 μ V/V + 7.0 μ V 35 μ V/V + 50 μ V 45 μ V/V + 0.60 mV 45 μ V/V + 10 mV		P and S
DC CURRENT Measurement	0 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 5 A 5 A to 10 A	50 μ V/V + 25 nA 50 μ V/V + 60 nA 50 μ V/V + 2.0 μ A 50 μ V/V + 5.0 μ A 0.10 % + 0.10 mA 0.20 % + 0.60 mA 0.12 % + 1.0 mA 0.32 % + 1.0 mA		P and S
AC VOLTAGE Measurement	1 mV to 100 mV 3 Hz to 5 Hz 5 Hz to 10 Hz 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	1.0 % + 30 μ V 0.35 % + 30 μ V 0.060 % + 30 μ V 0.12 % + 50 μ V 0.60 % + 80 μ V 4.0 % + 0.50 mV		P and S
	100 mV to 1 V 3 Hz to 5 Hz 5 Hz to 10 Hz 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	1.0 % + 0.30 mV 0.35 % + 0.30 mV 0.060 % + 0.30 mV 0.12 % + 0.50 mV 0.60 % + 0.80 mV 4.0 % + 5.0 mV		P and S
	1 V to 10 V 3 Hz to 5 Hz 5 Hz to 10 Hz 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	1.0 % + 3.0 mV 0.35 % + 3.0 mV 0.060 % + 3.0 mV 0.12 % + 5.0 mV 0.60 % + 8.0 mV 4.0 % + 50 mV		P and S



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
AC VOLTAGE Measurement (continued)	10 V to 100 V 3 Hz to 5 Hz 5 Hz to 10 Hz 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 100 V to 750 V 3 Hz to 5 Hz 5 Hz to 10 Hz 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	1.0 % + 30 mV 0.35 % + 30 mV 0.060 % + 30 mV 0.12 % + 50 mV 0.60 % + 80 mV 4.0 % + 0.50 V 1.0 % + 0.30 V 0.35 % + 0.30 V 0.060 % + 0.30 V 0.12 % + 0.50 V 0.60 % + 0.80 V 4.0 % + 3.8 V		P and S
AC CURRENT Measurement	10 µA to 100 µA 3 Hz to 5 kHz 100 µA to 1 mA 3 Hz to 5 kHz 1 mA to 10 mA 3 Hz to 5 kHz 10 mA to 100 mA 3 Hz to 5 kHz 100 mA to 1 A 3 Hz to 5 kHz 1 A to 3 A 3 Hz to 5 kHz 3 A to 10 A 3 Hz to 5 kHz	0.10 % + 40 nA 0.10 % + 0.40 µA 0.10 % + 4.0 µA 0.10 % + 40 µA 0.10 % + 0.40 mA 0.23 % + 1.2 mA 0.15 % + 4.0 mA		P and S
CAPACITANCE Measurement	0.0010 nF to 1.0000 nF 1.0000 nF to 10.000 nF 10.000 nF to 100.00 nF 100.00 nF to 1.0000 µF 1.0000 µF to 10.000 µF 10.000 µF to 100.00 µF	0.50 % + 5.0 pF 0.40 % + 10 pF 0.40 % + 100 pF 0.40 % + 1.0 nF 0.40 % + 10 nF 0.40 % + 100nF		P and S
FREQUENCY Measurement	3 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 300 kHz	0.10 % 0.030 % 0.010 %		P and S
FREQUENCY Measurement	10 mHz to 20 kHz 20 kHz to 200 kHz 200 kHz to 1 MHz	15 µHz/Hz 30 µHz/Hz 120 µHz/Hz	Using a digital sampling oscilloscope	P and S
FREQUENCY Measurement (continued)	1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 120 MHz	0.12 % 1.2 % 1.8 %		



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (<i>k</i> = 2)	Remarks	Location Code	
CALIBRATION OF TRANSIENT GENERATORS					
PULSE VOLTAGE	±10 V to ±15 kV	2.1 % of measured Voltage	Using a Digital Sampling Oscilloscope with Voltage or Current Probes as appropriate, with due consideration to EN 60469:2013.	P and S	
PULSE ABERRATIONS Peak voltage, overshoot and undershoot	±10 V to ±15 kV	2.1 % of measured Voltage			
PULSE CURRENT	±1 A to ±5000 A	2.1 % of measured Current			
PULSE ABERRATIONS Peak current, overshoot and undershoot	±1 A to ±5000 A	2.1 % of measured Current			
PULSE RISETIME			For the calibration of Transient Generators with due consideration to EN 60469:2013 using a Digital Sampling Oscilloscope with Voltage or Current Probes as appropriate and in accordance with their respective Operators Manuals.		
Voltage	1 ns to 1 µs 1 µs to 20 µs 20 µs to 100 s	0.12 ns 0.26 ns 13 µs/s			
Current	25 ns to 100 ns 100 ns to 500 ns 500 ns to 10 µs 10 µs to 20 µs 20 µs to 100 s	6.7 ns to 3.2 ns 3.2 ns to 0.66 ns 0.66 ns to 0.17 ns 0.17 ns to 0.26 ns 13 µs/s			
PULSE DURATION Voltage and current	1 ns to 1 µs 1 µs to 20 µs 20 µs to 100 s	0.12 ns 0.26 ns 13 µs/s			
PHASE ANGLE	0° to 360°	0.20°			
CALIBRATION OF ESD GENERATORS					
DC Voltage	1 kV to 32 kV	0.5 % of measured Voltage	For the calibration of ESD generators to EN 61000-4-2:2009, ISO 10605:2008 +A1:2014, & ISO 10605:2023, including instruments designed to comply with earlier versions of these standards.		P and S
Voltage Hold Time	1 s to 100 s	14 ms			
Transition Time	300 ps to 800 ps 800 ps to 1500 ps	28 ps (16 % to 6.0 %) 28 ps (6.0 % to 3.2 %)	The maximum uncertainty permitted by the standard is 15 % at the standardized value of 800 ps.		
Peak Current	0.1 A to 150 A	2.6%			
Current at 30 ns	0.1 A to 150 A	3.7 %			
Current at 60 ns	0.1 A to 150 A	3.7 %			
Current at 65 ns	0.1 A to 150 A	3.7 %			
Current at 130 ns	0.1 A to 150 A	3.7 %			
Current at 180 ns	0.1 A to 150 A	3.7 %			
Current at 400 ns	0.1 A to 150 A	4.4 %			
Current at 360 ns	0.1 A to 150 A	4.4 %			
Current at 800 ns	0.1 A to 150 A	6.0 %			



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
CALIBRATION OF EFTB GENERATORS				P and S
Peak Voltage into 50 Ω	± 1 V to ± 7.2 kV	2.2 % of measured Voltage	For the calibration of EFTB Generators to EN 61000-4-4:2012 ISO 7637-2:2011 including instruments designed to comply with earlier versions of these standards.	
Peak Voltage into 1 k Ω	± 1 V to ± 7.2 kV	3.3 % of measured Voltage		
Rise Time	1 ns to 10 ns	0.12 ns		
Pulse Width	10 ns to 50 μ s	0.59 ns		
Burst Duration	50 μ s to 50 ms	0.58 μ s		
Burst Period	50 ms to 500 ms	6.1 μ s		
Repetition Rate	1 Hz to 150 kHz	0.040 % of the measured frequency		
CALIBRATION OF SURGE GENERATORS				P and S
Open Circuit Voltage	± 10 V to ± 15 kV	2.1 % of measured Voltage	For the calibration of Surge Generators to EN 61000-4-5:2014/A1, ISO 7637-2:2011 ISO 16750-2:2012 and earlier versions of these standards.	
Voltage Undershoot	0 to 60 %	2.1 % of measured Voltage		
Voltage Front Time	100 ns to 20 μ s	0.30 ns		
Voltage Duration	1 μ s to 1 ms	16 ns		
S/C Current Peak	± 1 A to ± 5000 A	2.1 % of measured Current		
Current Undershoot	0 % to 60 %	2.1 % of measured Current		
Current Front Time	100 ns to 20 μ s	2.1 ns		
Current Duration	1 μ s to 500 μ s	6.4 ns		
Phase Angle	0° to 360°	0.20°		
Output Impedance	0.1 Ω to 100 Ω	2.9 %		
VOLTAGE DIPS AND INTERRUPTS (AC / DC)				P and S
AC / DC Voltage	100 mV to 750 V	0.26 % of measured Voltage	For the calibration of Voltage Dips and Interrupt Generators to EN 61000-4-11:2020, EN 61000-4-34:2007 +A1:2009 EN 61000-4-29:2001 and earlier versions of these standards.	
AC Frequency	10 Hz to 100 Hz	0.040 %		
Dip Levels	0 % to 100 %	0.27 % of measured Voltage		
Rise and Fall Times	500 ns to 10 μ s	0.20 ns		
Dip Durations	1 ms to 10 s	12 μ s/S		
Over / Undershoot	0 % to 100 %	2.1 % of measured Voltage		
Phase Angle	0° to 360°	0.20°		
Inrush Current	1 A to 1000 A	3.4 %		



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (<i>k</i> = 2)	Remarks	Location Code
POWER FREQUENCY MAGNETIC FIELDS				
AC Voltage	100 mV to 750 V	0.27 %	For the Calibration of Power Frequency Magnetic fields test Systems to EN 61000-4-8:2010 and earlier versions of these standards.	P and S
AC Frequency	50 Hz	0.12 Hz		
Loop Current	1 A to 3.45 A 3.45 A to 45.5 A 45.5 A to 400 A 400 A to 1000 A 1000 A to 2000 A	0.15 A 0.15 A to 0.55 A 0.55 A to 4.8 A 8.2 A to 18.2 A 18.2 A to 35 A		
Total Harmonic Distortion (THD)	0 % to 100 %	0.26 % of measured THD		
Magnetic Field	50 A/m to 150 A/m	2.4 %		
Coil Factor	0 to 1	2.7 % of the measured coil factor		
IMPULSE MAGNETIC FIELDS				
S/C Current Peak	±1.0 A to ±2500 A	2.1 %	For the calibration of Impulse Magnetic Fields Test system to EN 61000-4- 9:2016 and earlier versions of these standards	P and S
Current Undershoot	0 % to 60 %	2.1 % of the measured current		
Current Front Time	100 ns to 20 µs	0.40 ns		
Current Duration	1 µs to 500 µs	5.8 ns		
Phase Angle	0° to 360°	0.20°		
RING WAVE				
Open Circuit Voltage	±10 V to ±7.2 kV	2.1 %	For the Calibration of Ring Wave Immunity Test Generators to EN 61000-4-12:2017 and earlier versions of these standards.	P and S
Voltage Rise Time	100 ns to 20 µs	0.30 ns		
Oscillation Frequency	50 kHz to 150 kHz	21 µHz/Hz		
Voltage Decay Ratio	0.1 to 1.5	0.50 %		
S/C Current Peak	±1.0 A to ±1000 A	2.1 %		
Current Rise Time	100 ns to 20 µs	2.0 ns		
Phase Angle	0° to 360°	0.20°		
Impedance	1 Ω to 100 Ω	3.0 %		



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
CALIBRATION OF LINE IMPEDANCE STABILIZING NETWORKS & COUPLING / DECOUPLING NETWORKS				
IMPEDANCE MAGNITUDE				P and S
1 kHz to 400 MHz	0 Ω to 10 Ω	9.8 %	For the calibration of Line Impedance Stabilizing Networks (LISN's) and Coupling / Decoupling Networks (CDN's) Where the CMC is expressed as a "%" this indicates that the uncertainty is expressed as a percentage of the measured value.	
	10 Ω to 20 Ω	9.8 % to 5.3 %		
	20 Ω to 30 Ω	5.3 % to 3.4 %		
	30 Ω to 50 Ω	3.4 % to 2.7 %		
	50 Ω to 70 Ω	2.7 %		
	70 Ω to 100 Ω	2.7 % to 3.5 %		
	100 Ω to 150 Ω	3.5 % to 5.7 %		
	150 Ω to 200 Ω	5.7 % to 7.1 %		
INSERTION LOSS/VOLTAGE DIVISION FACTOR				
1 kHz to 400 MHz	0 dB to 40 dB	0.43 dB	This Insertion Loss capability applies to all insertion loss measurements within the stated frequency range.	P and S
	40 dB to 60 dB	0.47 dB		
	60 dB to 70 dB	0.55 dB		
IMPEDANCE PHASE=				
1 kHz to 400 MHz	+50 to - 50 degrees	1.94 % of measured phase		P and S
ISOLATION/DECOUPLING LOSS				
1 kHz to 400 MHz	0 dB to 70 dB	2.9 dB		P and S
END				



20765
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd
Issue No: 006 Issue date: 01 December 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$