Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address The Equipment Calibration Business The Annex W-17 Howbery Business Park Wallingford OX10 8BA United Kingdom	Local contact Mr Richard Harper Tel: +44 (0)01491 822691 E-Mail: Richard-harper@equipment-calibration.com Website: www.equipment-calibration.com	Electrical calibration: Calibration of EMC- related test and measurement equipment.	Ρ

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Address The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Local contact Mr Richard Harper Tel: +44 (0)01491 822691 Mobile: +44 (07380 950803 E-Mail: Richard-harper@equipment-calibration.com Website: www.equipment-calibration.com	Electrical calibration: Calibration of EMC- related test and measurement equipment.	S

NOTE: Where EN Standards have exact equivalents in IEC, or BS EN Standards, these are also included in the accreditation.

	Un 2 Pine Tre	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 20765 Accredited to ISO/IEC 17025:2017		The Equipment Calibrat Issue No:008 Issue dat	tion Business Ltd te: 05 March 2025		
	Calibration performe	d by the Organisation at the location	ons specified		
	Calibration a	and Measurement Capability (CMC)		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code	
All electrical measurements a	n are carried out using the me	thod of direct comparison or transfer t	o laboratory reference standard	s unless otherwise	
measure	outputs of submitted test ite	ms or to generate values as a stimulu	s for test items which measure.	ability to either	
DC RESISTANCE Measurement	0 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ	0.014 Ω 0.11 Ω 1.1 Ω 11 Ω 110 Ω 2.2 kΩ 0.83 MΩ		P and S	
DC VOLTAGE Measurement	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	8.7 μV 48 μV 410 μV 5.2 mV 56 mV		P and S	
DC CURRENT Measurement	0 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	0.077 μΑ 0.57 μΑ 7.1 μΑ 56 μΑ 1.1 mA		P and S	

	100 mA to 1 A 1 A to 3 A 3 A to 10 A	1.1 mA 6.7 mA 13 mA	
AC VOLTAGE Measurement	1 mV to 100 mV 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	92 μV 170 mV 690 mV	P and S
	100 mV to 1 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	920 μV 1.7 mV 7.0 mV	P and S
	1 V to 10 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	9.2 μV 17 mV 70 mV	P and S
	10 V to 100 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	92 mV 170 mV 700 mV	P and S

	Uni 2 Pine Tree	Schedule of Ac issued ted Kingdom Acc es, Chertsey Lane, Staine	creditation by reditation Service s-upon-Thames, TW18 3	HR, UK
20765 Accredited to ISO/IEC 17025:2017	The Equipment Calibration Business Ltd Issue No:008 Issue date: 05 March 2025			
	Calibration performed	by the Organisation at the locat	tions specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC VOLTAGE Measurement (continued)	100 V to 750 V 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	680 mV 1.3 V 5.2 V	AC Voltage measurement in this range is limited to a maximum 8 x 10 ⁷ Volt- Hertz.	
AC CURRENT Measurement	0 μA to 100 μA 10Hz to 5 kHz	140 nA		P and S
	100 μA to 1 mA <i>10 Hz to 5 kHz</i>	1.4 µA		
	1 mA to 10 mA 10 Hz to 5 kHz	15 μΑ		
	10 mA to 100 mA 10 Hz to 5 kHz	140 µA		
	100 mA to 1 A 10 Hz to 5 kHz	1.4 mA		
	1 A to 3 A 10 Hz to 5 kHz	8.2 mA		
	3 A to 10 A 10 Hz to 5 kHz	19 mA		
CAPACITANCE Measurement	0 nF to 1 nF 1 nF to 10 nF 10 nF to 100 nF 100 nF to 1 μF 1 μF to 10 μF 10 μF to 100 μF	1.0 pF 5.1 pF 51 pF 510 pF 5.1 nF 51 nF		P and S
FREQUENCY Measurement	10 Hz to 100 Hz	0.031 % of reading		P and S
	100 Hz to 300 Hz	0.010 % of reading		
FREQUENCY Measurement	10 mHz to 20 kHz 20 kHz to 200 kHz 200 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 120 MHz	15 μHz/Hz 30 μHz/Hz 120 μHz/Hz 0.12 % 1.2 % 1.8 %	Using a digital sampling oscilloscope	P and S

UKAS CALIBRATION 20765 Accredited to	Unite 2 Pine Trees, T	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK The Equipment Calibration Business Ltd Issue No:008 Issue date: 05 March 2025		
Calibration performed by the Organisation at the locations specified				
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code

		(k = 2)		
TRANSIENT GENERATORS				
PULSE VOLTAGE	±1 V to ±15 kV	1.6 % of measured Voltage		P and S
PULSE ABERRATIONS Peak voltage, overshoot and undershoot	±10 V to ±15 kV	1.6 % of measured Voltage		
PULSE CURRENT	±1 A to ±5 kA	1.7 % of measured Current	Using a Digital Sampling Oscilloscope with Voltage	
PULSE ABERRATIONS Peak current, overshoot and undershoot	±1 A to ±5 kA	1.7 % of measured Current	or Current Probes as appropriate, with due consideration to EN 60469:2013.	
PULSE RISETIME			For the calibration of	
Voltage	1 ns to 1 μs 1 μs to 20 μs 20 μs to 100 s	0.12 ns 0.26 ns 13 μs/s	Transient Generators with due consideration to EN 60469:2013 using a Digital Sampling Oscilloscope with Voltage	
Current	25 ns to 100 ns 100 ns to 500 ns 500 ns to 10 μs 10 μs to 20 μs 20 μs to 100 s	6.7 ns to 3.2 ns 3.2 ns to 0.66 ns 0.66 ns to 0.17 ns 0.17 ns to 0.26 ns 13 μs/s	or Current Probes as appropriate and in accordance with their respective Operators Manuals.	
PULSE DURATION Voltage and current	1 ns to 1 μs 1 μs to 20 μs 20 μs to 100 s	0.12 ns 0.26 ns 13 µs/s	Polotivo to power line and	
PHASE ANGLE	0° to 360°	0.8°	Using a Digital Sampling Oscilloscope and a HV probe.	
ESD GENERATORS			For the calibration of ESD	P and S
DC Voltage	1 kV to 32 kV	0.5 % of measured Voltage	generators to EN 61000-4-2:2009,	
Voltage Hold Time	1 s to 100 s	14 ms	& ISO 10605:2003 +A12014, & ISO 10605:2023, including instruments designed to comply with earlier versions of these standards.	
Transition Time	300 ps to 800 ps	28 ps		
	800 ps to 1500 ps	28 ps	The maximum uncertainty	
Peak Current	0.1 A to 150 A	2.6%	permitted by the standard is 15 % at the standardized	
Current at 30 ns Current at 60 ns Current at 65 ns Current at 130 ns Current at 180 ns Current at 400 ns	0.1 A to 150 A 0.1 A to 150 A	3.7 % 3.7 % 3.7 % 3.7 % 3.7 % 4.4 %	value of 800 ps.	
Current at 360 ns	0.1 A to 150 A 0.1 A to 150 A	4.4 % 6.0 %		



Accredited to ISO/IEC 17025:2017 Schedule of Accreditation ^{issued by} Jnited Kingdom Accreditation S

United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Equipment Calibration Business Ltd

Issue No:008

Issue date: 05 March 2025

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
EFTB GENERATORS				P and S
Peak Voltage into 50 Ω	±1 V to ±7.2 kV	2.2 % of measured Voltage		
Peak Voltage into 1 k Ω	±1 V to ±7.2 kV	3.3 % of measured Voltage	For the calibration of	
Rise Time	1 ns to 10 ns	0.12 ns	EFTB Generators to EN 61000-4-4:2012	
Pulse Width	10 ns to 50 µs	0.59 ns	ISO 7637-2:2011 including instruments	
Burst Duration	50 µs to 50 ms	0.58 µs	earlier versions of these	
Burst Period	50 ms to 500 ms	6.1 µs	standards.	
Repetition Rate	1 Hz to 150 kHz	0.040 % of the measured frequency		
SURGE GENERATORS				P and S
Open Circuit Voltage	±1 V to ±25 kV	1.6 % of measured Voltage		
Voltage Undershoot	0 to 60 %	1.6 % of measured Voltage		
Voltage Front Time	100 ns to 20 µs	2.5 ns	For the calibration of Surge Generators to EN 61000-4-5:2014/A1, ISO 7637-2:2011 ISO 16750-2:2012	
Voltage Duration	1 µs to 1 ms	365 ns		
S/C Current Peak	±1 A to ±5 kA	1.7 % of measured Current		
Current Undershoot	0 % to 60 %	1.7 % of measured Current	these standards.	
Current Front Time	100 ns to 20 µs	35 ns		
Current Duration	1 µs to 500 µs	89 ns		
Phase Angle	0° to 360°	0.8°		
Output Impedance	0.1 Ω to 500 Ω	2.3 %		
VOLTAGE DIPS AND INTER	RUPTS (AC / DC)			P and S
AC Voltage <i>50 Hz /</i> DC Voltage	100 mV to 750 V	0.26 % of measured Voltage		
AC Frequency	10 Hz to 100 Hz	0.040 %	For the calibration of	
Dip Levels	0 % to 100 %	0.27 % of measured Voltage	Voltage Dips and Interrupt Generators to	
Rise and Fall Times	500 ns to 10 µs	0.20 ns	EN 61000-4-11:2020, EN 61000-4-34:2007 +A1:2009 EN 61000-4-29:2001 and earlier versions of these standards.	
Dip Durations	1 ms to 10 s	12 µs/S		
Over / Undershoot	0 % to 100 %	2.1 % of measured Voltage		
Phase Angle	0° to 360°	0.20°		
Inrush Current	1 A to 1000 A	3.4 % of measured inrush current		



Instrument or Gauge	Range	Uncertainty $(k=2)$	Remarks	Code
POWER FREQUENCY MAGNETIC FIELDS				D and O
AC Voltage 50 Hz	100 mV to 750 V	0.27 %		P and S
AC Frequency	50 Hz	0.12 Hz		
Loop Current	1 A to 3.45 A 3.45 A to 45.5 A 45.5 A to 400 A 400 A to 1 kA 1 kA to 2 kA	0.15 A 0.15 A to 0.55 A 0.55 A to 4.8 A 8.2 A to 18 A 18 A to 35 A	For the calibration of Power Frequency Magnetic fields test Systems to EN 61000- 4-8:2010 and earlier versions of these standards.	
Total Harmonic Distortion (THD)	0 % to 100 %	0.26 % of measured THD		
Magnetic Field	50 A/m to 150 A/m	2.4 %		
Coil Factor	0 to 1	2.7 % of the measured coil factor		
IMPULSE MAGNETIC FIELD	s			P and S
S/C Current Peak	±1.0 A to ±2500 A	2.1 % of the measured current		
Current Undershoot	0 % to 60 %	2.1 % of the measured current	For the calibration of Impulse Magnetic Fields	
Current Front Time	100 ns to 20 µs	0.40 ns	9:2016 and earlier versions	
Current Duration	1 μs to 500 μs	5.8 ns		
Phase Angle	0° to 360°	0.20°		
RING WAVE				P and S
Open Circuit Voltage	±10 V to ±7.2 kV	2.1 %		
Voltage Rise Time	100 ns to 20 µs	0.30 ns		
Oscillation Frequency	50 kHz to 150 kHz	21 µHz/Hz	For the calibration of Ring	
Voltage Decay Ratio	0.1 to 1.5	0.50 %	Wave Immunity Test Generators to	
S/C Current Peak	±1.0 A to ±1000 A	2.1 %	EN 61000-4-12:2017 and earlier versions of these	
Current Rise Time	100 ns to 20 µs	2.0 ns	standards.	
Phase Angle	0° to 360°	0.20°		
Impedance	1 Ω to 100 Ω	3.0 %		



LINE IMPEDANCE STABILISATION NETWORKS & COUPLING / DECOUPLING NETWORKS				
IMPEDANCE MAGNITUDE				P and S
1 kHz to 400 MHz	0 Ω to 10 Ω	9.8 %		
	10 Ω to 20 Ω	9.8 % to 5.3 %	For the calibration of Line Impedance Stabilizing	
	20 Ω to 30 Ω	5.3 % to 3.4 %	Networks (LISN's) and Coupling / Decoupling	
	30 Ω to 50 Ω	3.4 % to 2.7 %	Networks (CDN's)	
	50 Ω to 70 Ω	2.7 %	Where the CMC is expressed as a "%" this	
	70 Ω to 100 Ω	2.7 % to 3.5 %	indicates that the uncertainty in expressed as	
	100 Ω to 150 Ω	3.5 % to 5.7 %	measured value.	
	150 Ω to 200 Ω	5.7 % to 7.1 %		
INSERTION LOSS/VOLTAGE DIVISION FACTOR				
1 kHz to 400 MHz	0 dB to 40 dB	0.43 dB	This Insertion Loss capability applies to all insertion loss	P and S
	40 dB to 60 dB	0.47 dB	measurements within the stated frequency range.	
	60 dB to 70 dB	0.55 dB		
IMPEDANCE PHASE				
1 kHz to 400 MHz	+50 to – 50 degrees	1.9 degrees		P and S
ISOLATION/DECOUPLING LOSS				
1 kHz to 400 MHz	0 dB to 90 dB	2.9 dB		P and S
END				



Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: Q[a, b] = [a² + b²]^{1/2}