


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>4006</b> Accredited to <b>ISO/IEC 17025:2017</b>	<b>Guymark UK Limited</b> Issue No: 011 Issue date: 26 August 2021	
	Veronica House Old Bush Street Brierley Hill West Midlands DY5 1UB	Contact: Mr A Sewell Tel: +44 (0) 1384 890600 Fax: +44 (0)1384 890609 E-Mail: <a href="mailto:adrian@guymark.com">adrian@guymark.com</a> Website: <a href="http://www.guymark.com">www.guymark.com</a>

Calibration performed by the Organisations at the locations specified below

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Veronica House Old Bush Street Brierley Hill West Midlands DY5 1UB  <b>Local contact</b> Mr A Sewell	Audiometer calibrations	Lab

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Any suitable customer premises	Audiometer calibrations	Site



**4006**  
Accredited to  
**ISO/IEC 17025:2017**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Guymark UK Limited**  
**Issue No: 011 Issue date: 26 August 2021**

**Calibration performed by the Organisation at the locations specified**

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC)	Remarks	Location Code
AUDIOMETERS: Pure tone air conduction SPL	125 Hz to 8000 Hz	0.5 dB 0.7 dB	Types 1 to 4, as classified in BS EN 60645-1:1995 (Withdrawn), BS EN 60645-1:2001 (Withdrawn), BS EN 60645-1:2015 (Withdrawn) BS EN 60645-1:2017	Lab Site
Narrowband masking	125 Hz to 8000 Hz	0.7 dB 1.0 dB		Lab Site
Pure tone bone conduction FL	250 Hz to 4000 Hz	1.50 dB		Lab & Site
Pure tone frequency	125 Hz to 1000 Hz 1000 Hz to 8000 Hz	0.10 Hz 1.00 Hz		Lab & Site
Total harmonic distortion	125 Hz to 8000 Hz	0.3% THD		Lab & Site
Attenuator linearity 130 dB to 40 dB HLS <40 dB to 0 dB HLS <0 dB to -10 dB HLS	1000 Hz	0.20 dB 0.40 dB 0.50 dB	Measured between successive Hearing Level Settings	Lab
END				



4006

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Guymark UK Limited**

**Issue No: 011 Issue date: 26 August 2021**

**Calibration performed by the Organisation at the locations specified**

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$