

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



4280

Accredited to  
ISO/IEC 17025:2017

### The Radiological Protection Centre Incorporating the John Perry Laboratory St George's University Hospitals NHS Foundation Trust

Issue No: 016 Issue date: 02 September 2021

Unit 5  
The Observatory  
24 Deer Park Road  
London  
SW19 3UA

Contact: Mehboob Rathod  
Tel: +44 (0)20 8725 1498  
Fax: +44 (0)20 8417 1338  
E-Mail: mehboob.rathod@stgeorges.nhs.uk  
Website: www.sghrpc.co.uk

Calibration performed at the above address only

#### Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks   |
|---|---|---|---|
| <b>RADIOLOGICAL</b>   |   |   |   |
| <u>Air Kerma</u>  | BS EN 61267:2006 all RQR qualities<br><br>BS EN 61267:2006 all RQR-M qualities<br><br>BS EN 61267:2006 RQT8 and RQT9 qualities<br><br>BS EN 61267:2006 RQA5 quality<br><br>75kV with 2.5 mm Al added filtration and HVL of 2.5mm Al<br>70 - 75kV with added filtration of 2.5 mm Al + 2.0mm Cu and HVL of 9.2mm Al<br>125kV with added filtration of 2.5mm Al + 0.2mm Cu and HVL of 8.5mm Al<br>125kV with added filtration of 2.5mm Al + 1.0mm Cu and HVL of 11.3mm Al<br>70kV with added filtration of 2.5mm Al + 21.0mm Al and HVL of 6.5mm Al<br><br>25kV-35kV with total filtration 30µm Mo + 1mm Be | 1.6%  | Calibration of dosimeters intended for use in diagnostic radiology at the beam qualities stated<br><br>Calibration against a secondary standard instrument.   |
| <u>kVp Meters</u><br><u>Peak Practical Voltage</u><br><u>Peak Average Voltage</u> | Beam qualities:<br><br>50-150kV with total filtration of 2.5mm Al or 3.0mm Al on the diagnostic tube<br><br>24-35kV with total filtration of 30µm Mo + 1mm Be on the mammography tube   | 0.7%  | <u>Non-invasive kV calibrations</u><br><br>Calibration of non-invasive kVp meters intended for use in diagnostic radiology at the beam qualities stated<br><br>Calibration against an Invasive Dynalyzer High Voltage Unit. |



4280  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**The Radiological Protection Centre**  
**Incorporating the John Perry Laboratory**  
**St George's University Hospitals NHS Foundation Trust**

**Issue No: 016 Issue date: 02 September 2021**

Calibration performed at main address only

| Measured Quantity<br>Instrument or Gauge | Range                       | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  |
|--|-----------------------------|---|--|
| <u>Dynalyzer High Voltage Unit</u>       |                             |   | Calibration of the Dynalyzer High Voltage Unit (HVU) |
| Anode current                            | 10 mA to 250 mA             | 0.40 %  | Calibration against an electrical multimeter.        |
| Filament current                         | 1 A to 10 A                 | 0.6%  |  |
| Anode frequency response                 | 10 kHz to 170 kHz at 1000 V | 1.0 %   |  |
| Cathode frequency response               | 10 kHz to 170 kHz at 1000 V | 1.0 %   |  |
| Cathode kV                               | 5 kV to 75 kV               | 0.40 %  |  |
| Anode kV                                 | 5 kV to 75 kV               | 0.40 %  |  |
| END                                      |                             |   |  |



4280  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**The Radiological Protection Centre**  
**Incorporating the John Perry Laboratory**  
**St George's University Hospitals NHS Foundation Trust**

Issue No: 016 Issue date: 02 September 2021

Calibration performed at main address only

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

### Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10<sup>-6</sup>·p·10<sup>-6</sup>) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.