

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION</p> <p>4432</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Caltest Instruments Limited</h3> <p>Issue No: 009      Issue date: 15 November 2019</p>	
	<p>Unit 2 Viceroy Court Bedford Road Petersfield Hampshire GU32 3LJ</p>	<p>Contact: Ian Norman Tel: +44 (0)1483 302700 E-Mail: inorman@caltest.co.uk Website: www.caltest.co.uk</p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
DC RESISTANCE	0 $\Omega$ 1 m $\Omega$ to 10 m $\Omega$ 10 m $\Omega$ to 100 m $\Omega$ 100 m $\Omega$ to 1 $\Omega$ 1 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$ 1 G $\Omega$ to 10 G $\Omega$ 100 G $\Omega$	0.3 $\mu\Omega$ 70 ppm 60 ppm 25 ppm 5.0 ppm 5.0 ppm 6.0 ppm 4.0 ppm 4.0 ppm 7.0 ppm 8.0 ppm 15 ppm 35 ppm 190 ppm	Outputs of instruments within these values can be measured to the listed uncertainties
Set Values	1.0 m $\Omega$ 10 m $\Omega$ 0.1 $\Omega$ 1.0 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1.0 k $\Omega$ 10 k $\Omega$  1 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$	70 ppm 60 ppm 25 ppm 5.0 ppm 5.0 ppm 3.0 ppm 4.0 ppm 3.0 ppm  10 ppm 5.0 ppm 4.0 ppm 3.0 ppm 4.0 ppm 4.0 ppm 10 ppm 30 ppm	These values can be generated for the calibration of measuring instruments
DC VOLTAGE	0 V to 1.1 V 1.1 V to 11 V 11 V to 110 V 110 V to 1 kV	0.50 ppm + 0.82 $\mu$ V 0.50 ppm + 2.0 $\mu$ V 0.5 ppm + 2.1 $\mu$ V 1.0 ppm	These values can be measured and generated for the calibration of sourcing and measuring instruments



4432  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Caltest Instruments Limited**  
**Issue No: 009 Issue date: 15 November 2019**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
DC VOLTAGE (cont'd)	2 kV to 10 kV	0.30 %	Measurement only
	1 kV to 2 kV	0.15 %	
DC CURRENT	1 $\mu$ A to 100 mA	17 ppm	These values can be measured and generated for the calibration of sourcing and measuring instruments
	100 mA to 1 A	55 ppm	
	1 A to 10 A	75 ppm	
	10 A to 50 A	200 ppm	
	50 A to 100 A	300 ppm	
AC VOLTAGE	2 mV to 200 mV	0.0060 % + 4.0 $\mu$ V 0.0070 % + 4.0 $\mu$ V 0.010 % + 8.0 $\mu$ V 0.033 % + 20 $\mu$ V	These values can be measured and generated for the calibration of sourcing and measuring instruments
	40 Hz to 1 kHz		
	1 kHz to 10 kHz		
	10 kHz to 20 kHz		
	20 kHz to 100 kHz		
	200 mV to 2 V	0.0030 % + 20 $\mu$ V 0.0007 % + 20 $\mu$ V 0.0060 % + 40 $\mu$ V 0.012 % + 200 $\mu$ V	
	40 Hz to 1 kHz		
	1 kHz to 10 kHz		
	10 kHz to 20 kHz		
	20 kHz to 100 kHz		
	2 V to 20 V	0.0030 % + 200 $\mu$ V 0.0050 % + 200 $\mu$ V 0.0050 % + 400 $\mu$ V 0.011 % + 2.0 mV	
	40 Hz to 1 kHz		
	1 kHz to 10 kHz		
	10 kHz to 20 kHz		
	20 kHz to 100 kHz		
20 V to 200 V	0.0040 % + 2.0 mV 0.0040 % + 2.0 mV 0.0060 % + 4.0 mV 0.012 % + 20 mV		
40 Hz to 1 kHz			
1 kHz to 10 kHz			
10 kHz to 20 kHz			
20 kHz to 100 kHz			
200 V to 500 V	0.0080 % + 20 mV 0.011 % + 20 mV 0.032 % + 40 mV		
40 Hz to 1 kHz			
1 kHz to 10 kHz			
10 kHz to 20 kHz			
500 V to 700 V	0.010 % + 20 mV 0.013 % + 20 mV 0.038 % + 40 mV		
40 Hz to 1 kHz			
1 kHz to 10 kHz			
10 kHz to 20 kHz			
700 V to 1000 V	0.021 % + 20 mV 0.023 % + 20 mV 0.075 % + 40 mV		
40 Hz to 1 kHz			
1 kHz to 10 kHz			
10 kHz to 20 kHz			
50 Hz to 60 Hz	0.35 % 0.90 %	Measurement only	
1 kV to 2 kV			
2 kV to 10 kV			



4432  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Caltest Instruments Limited**  
**Issue No: 009 Issue date: 15 November 2019**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	
AC CURRENT	40 Hz to 1 kHz 5 mA to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 20 A	46 ppm 65 ppm 100 ppm 120 ppm	These values can be measured and generated for the calibration of sourcing and measuring instruments	
	45 Hz to 60 Hz 20 A to 100 A	180 ppm		
FREQUENCY	40 Hz to 225 MHz	2.0 parts in $10^8$	Frequency results may be expressed in terms of average periodic time.	
ACTIVE POWER AND ENERGY	45 to 65 Hz 15 W to 2.4 kW 0.8 to unity PF 1 Ø 0.8 to unity PF 3 Ø	75 ppm 60 ppm	Limiting voltage 60 V to 480 V * Limiting current 0.25 A to 5 A	
	0.5 Power Factor 1 Ø 0.5 Power Factor 3 Ø	85 ppm 85 ppm		
	0.3 Power Factor 1 Ø 0.3 Power Factor 3 Ø	110 ppm 110 ppm	* The CMC is increased to 200 ppm at voltages greater than 120 V for which power factors other than unity are used.	
	45 to 65 Hz 0.6 W to 120W, 300 W to 9.6 kW 0.8 to unity PF 1 Ø 0.8 to unity PF 3 Ø	75 ppm 80 ppm		
	0.5 Power Factor 1 Ø 0.5 Power Factor 3 Ø	100 ppm 100 ppm		
	0.3 Power Factor 1 Ø 0.3 Power Factor 3 Ø	130 ppm 130 ppm		
	45 Hz to 65 Hz 0.6W to 48 kW 0.3 to unity PF 1 Ø 0.3 to unity PF 3 Ø	200 ppm 200 ppm	Limiting voltage 60 V to 480 V * Limiting current 0.01 A to 100 A	
				* The CMC is increased to 200 ppm at voltages greater than 120 V for which power factors other than unity are used.



4432  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Caltest Instruments Limited**  
**Issue No: 009 Issue date: 15 November 2019**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
REACTIVE POWER AND ENERGY	48 Hz to 52 Hz		Limiting voltage 60 V to 480 V * Limiting current 0.05 A to 20 A
	3 VAr to 9.6 kVAr		
	Zero Power Factor 1 Ø	90 ppm	* The CMC is increased to 200 ppm at voltages greater than 120 V for which power factors other than unity are used.
	Zero Power Factor 3 Ø	100 ppm	
	0.866 Power Factor 1 Ø	100 ppm	
	0.866 Power Factor 3 Ø	110 ppm	
	48 Hz to 52 Hz		Limiting voltage 60 V to 480 V * Limiting current 20 A to 100 A
	1.2 kVAr to 48 kVAr		
Zero Power Factor 1 Ø	110 ppm	* The CMC is increased to 200 ppm at voltages greater than 120 V for which power factors other than unity are used.	
Zero Power Factor 3 Ø	120 ppm		
0.866 Power Factor 1 Ø	120 ppm		
0.866 Power Factor 3 Ø	130 ppm		
END			



4432

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Caltest Instruments Limited**  
Issue No: 009 Issue date: 15 November 2019

Calibration performed at main address only

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

### Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10<sup>-6</sup>·p·10<sup>-6</sup>) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.