


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS CALIBRATION 4635 Accredited to ISO/IEC 17025:2017	Avidity Science Ltd	
	Issue No: 011 Issue date: 02 September 2021	
	Unit D4 Drakes Drive Long Crendon Industrial Estate Long Crendon Buckinghamshire HP18 9BA	Contact: Victoria Plater Tel: +44 (0)1844 203651 Fax: +44 (0)1844 203650 E-Mail: Victoria.Plater@AvidityScience.com Website: www.AvidityScience.com
Calibration performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. Local contact Victoria Plater Contact details as above	Time Pressure Temperature	S



4635
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Avidity Science Ltd
Issue No: 011 **Issue date:** 02 September 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
<p>PRESSURE</p> <p>Gas pressure (gauge)</p>	-80 kPa to 300 kPa	3 kPa	Methods consistent with EURAMET CG3 and CG17.	S
<p>TEMPERATURE</p> <p>Temperature controlled, autoclaves, incubators, ovens, environmental chambers and liquid baths.(inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)</p>	30 °C to 135 °C	0.55 °C	Calibration by comparison with reference thermometers Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	S
Time Interval	10 s to 4 hours	2.0 s	Calibration by comparison with a reference timer	S
END				



4635
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Avidity Science Ltd
Issue No: 011 **Issue date:** 02 September 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$