


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>5204</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Airbus Operations Ltd</p> <p>Issue No: 010 Issue date: 23 June 2021</p>	
	<p>AWIC Airbus Operations Building 07Y Aerospace Avenue Filton Bristol BS34 7PA</p>	<p>Contact: Mr James Steeds Tel: +44(0)11793 62254 Fax: +44 (0)11793 62522 E-Mail: james.steeds@airbus.com Website: calibration@airbus.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
<p>Electrical Electrical measurement and sourcing capabilities listed below follow the method of direct comparison against laboratory references or established ratio technique unless otherwise stated in the remarks column. This includes Time, Frequency and Temperature Simulation</p>			
DC Voltage	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV	0.6 μ V 5.3 ppm 3.2 ppm 3.1 ppm 4.7 ppm	
DC Resistance	0 Ω to 1 Ω 1 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω 2 G Ω to 20 G Ω	13 μ Ω 13 ppm 8.3 ppm 7.7 ppm 7.7 ppm 7.7 ppm 7.9 ppm 9.6 ppm 19 ppm 110 ppm 0.12 % 0.42 %	
Current carrying resistors	1 m Ω at:- 3 A 10 A 20 A 10 m Ω at:- 3 A 10 A 20 A	0.33 μ Ω 0.15 μ Ω 0.76 μ Ω 2.6 μ Ω 1.4 μ Ω 0.70 μ Ω	
DC Current	0 μ A to 20 μ A 20 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA	1.7 nA 11 ppm 11 ppm 11 ppm 41 ppm	



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 Issue date: 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DC Current (continued)	200 mA to 2 A 2 A to 20 A	180 ppm 410 ppm	
AC Voltage	2 mV to 200 mV 10Hz to 40Hz 40Hz to 100Hz 100Hz to 2kHz 2kHz to 10kHz 10kHz to 30kHz 30kHz to 100kHz	150 ppm 130 ppm 120 ppm 130 ppm 350 ppm 810 ppm	
	200 mV to 2 V 10Hz to 40Hz 40Hz to 100Hz 100Hz to 2kHz 2kHz to 10kHz 10kHz to 30kHz 30kHz to 100kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	110 ppm 93 ppm 73 ppm 96 ppm 220 ppm 610 ppm 0.40 % 2.0 %	
	2 V to 20 V 10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	110 ppm 93 ppm 74 ppm 96 ppm 220 ppm 610 ppm 0.40 % 2.0 %	
	20 V to 200 V 10Hz to 40Hz 40Hz to 100Hz 100Hz to 2kHz 2kHz to 10kHz 10kHz to 30kHz 30kHz to 100kHz	120 ppm 97 ppm 76 ppm 98 ppm 220 ppm 610 ppm	
	200 V to 1 kV 10Hz to 40Hz 40Hz to 30 kHz	110 ppm 250 ppm	
AC Current	10 Hz to 10 kHz 200 μ A 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	110 nA 370 ppm 350 ppm 360 ppm 830 ppm 910 ppm	



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 Issue date: 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DC Voltage 5700A system	0 mV to 200 mV 200 mV to 2 V 2 V to 11 V 11 V to 20 V 20 V to 200 V 200 V to 1 kV	7.0 ppm + 0.70 μ V 4.2 ppm + 1.3 μ V 4.2 ppm + 4.1 μ V 3.8 ppm + 8.1 μ V 4.2 ppm + 92 μ V 4.8 ppm + 580 μ V	
DC Resistance 5700A system	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	50 $\mu\Omega$ 75 ppm 95 ppm 28 ppm 27 ppm 17 ppm 17 ppm 13 ppm 13 ppm 12 ppm 12 ppm 14 ppm 14 ppm 20 ppm 21 ppm 40 ppm 47 ppm 110 ppm	
DC Current 5700A system	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	7.5 ppm + 1.8 nA 26 ppm + 4.6 nA 26 ppm + 92 nA 33 ppm + 290 nA 48 ppm + 6.9 μ A	
5520A system	2 A to 3 A 3 A to 20 A	780 μ A 1.3 mA	
AC Voltage 5700A system	5 μ V to 2 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	550 ppm + 4.5 μ V 210 ppm + 4.5 μ V 110 ppm + 4.5 μ V 370 ppm + 4.5 μ V 850 ppm + 7.0 μ V 0.11 % + 13 μ V 0.17 % + 25 μ V 0.34 % + 25 μ V	
	2 mV to 20 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	550 ppm + 5.0 μ V 210 ppm + 5.0 μ V 110 ppm + 5.0 μ V 370 ppm + 5.0 μ V 850 ppm + 7.0 μ V 0.11 % + 13 μ V 0.17 % + 25 μ V 0.34 % + 25 μ V	



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 Issue date: 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	
AC Voltage (continued) 5700A system	20 mV to 200 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	550 ppm + 13 µV 210 ppm + 8.0 µV 110 ppm + 8.0 µV 370 ppm + 8.0 µV 850 ppm + 25 µV 500 ppm + 25 µV 0.13 % + 35 µV 0.27 % + 80 µV		
	200 mV to 2 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	550 ppm + 80 µV 210 ppm + 25 µV 110 ppm + 6.0 µV 370 ppm + 16 µV 850 ppm + 70 µV 500 ppm + 130 µV 0.13 % + 350 µV 0.27 % + 850 µV		
	2 V to 20 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	550 ppm + 800 µV 160 ppm + 250 µV 75 ppm + 60 µV 120 ppm + 160 µV 250 ppm + 350 µV 500 ppm + 15 mV 0.13 % + 4.3 mV 0.27 % + 8.5 mV		
	20 V to 200 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	550 ppm + 8.0 mV 160 ppm + 3.0 mV 80 ppm + 1.0 mV 220 ppm + 4.0 mV 500 ppm + 8.0 mV 0.15 % + 90 mV 0.47 % + 90 mV 1.2 % + 190 mV		
	200 V to 1.1 kV 15 Hz to 50 Hz 50 Hz to 1 kHz	400 ppm + 15 mV 80 ppm + 4 mV		
	AC Current 5700A system	20 µA to 200 µA 55 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	140 ppm+ 16 nA 600 ppm + 40 nA 0.16 % + 80 nA	
		200 µA to 2 mA 55 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	140 ppm+ 35 nA 600 ppm + 400 nA 0.16 % + 800 nA	



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 **Issue date:** 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	
AC Current (continued) 5700A system	2 mA to 20 mA 55 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	140 ppm+ 350 nA 600 ppm + 4.0 µA 0.16 % + 8.0 µA		
	20 mA to 200 mA 55 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	140 ppm+ 3.5 µA 600 ppm + 40 µA 0.16 % + 80 µA		
	200 mA to 2.0 A 55 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	650 ppm+ 35 µA 750 ppm + 80 µA 0.85 % + 160 µA		
	5520A system	2 A to 3 A 45 Hz to 1 kHz 1 kHz to 5 kHz		1.2 mA 1.1 mA
		3 A to 20 A 45 Hz to 1 kHz 1 kHz to 5 kHz		23 mA 26 mA
Capacitance	0.19 nF to 0.4 nF 10 Hz to 10 kHz	8.9 pF		
	0.4 nF to 1.1 nF 10 Hz to 10 kHz	12 pF		
	1.1 nF to 3.3 nF 10 Hz to 3 kHz	19 pF		
	3.3 nF to 11 nF 10 Hz to 1 kHz	27 pF		
	11 nF to 33 nF 10 Hz to 1 kHz	136 pF		
	33 nF to 110 nF 10 Hz to 1 kHz	270 pF		
	110 nF to 330 nF 10 Hz to 1 kHz	810 pF		
	330 nF to 1.1 µF 10 Hz to 600 Hz	2.7 nF		
	1.1 µF to 3.3 µF 10 Hz to 300 Hz	8.1 nF		
	3.3 µF to 11 µF 10 Hz to 150 Hz	27 nF		
	11 µF to 33 µF 10 Hz to 120 Hz	120 nF		
	33 µF to 110 µF 10 Hz to 80 Hz	430 nF		
	110 µF to 330 µF DC to 50 Hz	1.3 µF		
330 µF to 1.1 mF DC to 20 Hz	4.3 µF			

These values are sourced by simulation.



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 Issue date: 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Capacitance continued	1.1 mF to 3.3 mF <i>DC to 6 Hz</i> 3.3mF to 11 mF <i>DC to 2 Hz</i> 11 to 33 mF <i>DC to 0.6 Hz</i> 33 to 110 mF DC to 0.2 Hz	13 uF 43 uF 200 uF 930 uF	
AC Resistance	50 Hz to 1 kHz 0 Ω to 1 Ω	5 m Ω	
Spot frequency 1 kHz \pm 10 Hz	5 m Ω to 1 Ω	0.31 %	Calibration of Aircraft loop test box.
AC Conductivity At a nominal 60 kHz 0.8 % IACS to 15 % IACS 15 % IACS to 59.56% IACS 59.56 % IACS to 100 % IACS	0.14 MS/m to 0.26 MS/m 0.27 MS/m to 1.03 MS/m 1.03 MS/m to 58 MS/m	1.4 % of value 0.85 % of value 1.6 % of value	Note; 58.0 MS/m = 100 % on the International Annealed Copper Scale
Frequency	10 MHz reference 1 mHz to 1 GHz	5 in 10^{10} 1 in 10^9	
Time Repetitive events Single shot Mechanically triggered	10 ns to 103 s 10 ns to 103 s 100 ms to 103 s	1 ns 2 ns 50 ms	
Temperature Simulation Calibrators and displays Thermocouples Type:			Including reference junction compensation
K	-270 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to 120 $^{\circ}$ C 120 $^{\circ}$ C to 1372 $^{\circ}$ C	0.23 $^{\circ}$ C 0.14 $^{\circ}$ C 0.28 $^{\circ}$ C	
T	-250 $^{\circ}$ C to -150 $^{\circ}$ C -150 $^{\circ}$ C to 0 $^{\circ}$ C 0 $^{\circ}$ C to 400 $^{\circ}$ C	0.43 $^{\circ}$ C 0.19 $^{\circ}$ C 0.14 $^{\circ}$ C	
Calibrators and displays Thermocouples Type:			
J	-210 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to 150 $^{\circ}$ C 150 $^{\circ}$ C to 1200 $^{\circ}$ C	0.19 $^{\circ}$ C 0.12 $^{\circ}$ C 0.16 $^{\circ}$ C	



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 Issue date: 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Temperature Simulation continued			
Resistance thermometer (Pt 100)	-200 °C to 800 °C	0.070 °C	
Reference Junction Compensation	Nominal Ambient 20 °C to 21 °C	0.050 °C	
<u>Force Measuring Devices</u>			
Calibration of load cells (excluding proving devices) by force proving instruments in tension and compression	0.25 kN to 1200 kN	0.36 %	
<u>Pressure</u>			
All pressure calibrations performed by direct comparison to reference instruments Pressure devices with an electrical output can be calibrated.			Calibration of pressure indicating instruments and gauges
<u>Hydraulic pressure (gauge)</u>	600 kPa to 6 MPa 6 MPa to 120 MPa	$\sqrt{(0.020 \% p)^2 + (180 \text{ Pa})^2}$ $\sqrt{(0.014 \% p)^2 + (180 \text{ Pa})^2}$	p in Pa
<u>Gas pressure (gauge)</u>			Absolute pressure calibrations can be undertaken using associated barometric pressure measurement correction. The uncertainties quoted will be increased by 25 Pa
Calibration of pressure indicating instruments and gauges	-99 kPa to 0 0 Pa to 160 kPa 160 kPa to 7 MPa	25 Pa 25 Pa 500 Pa	
<u>Gas pressure (absolute)</u>			
Calibration of pressure indicating instruments and gauges	84 kPa to 115 kPa	25 Pa	
<u>Temperature</u>			
PRT (4 wire)	-90 °C to -40 °C -40 °C to 125 °C 125 °C to 250 °C 0.01 °C	0.15 °C 0.024 °C 0.022 °C 0.0055 °C	Using a metal block calibrator In silicon oil bath In silicon oil bath In Triple Point of Water Cell
Temperature sensors and indicators	-90 °C to -40 °C -40 °C to 125 °C 125 °C to 250 °C	0.15 °C 0.024 °C 0.022 °C	Using a metal block calibrator In silicon oil bath In silicon oil bath



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 **Issue date:** 23 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Temperature continued Thermocouples (base Metal)	-90 °C to -40 °C -40 °C to 125 °C 125 °C to 250 °C	0.50 °C 0.38 °C 0.42 °C	Using a metal block calibrator In silicon oil bath In silicon oil bath
Humidity Calibration of hygrometers and temperature in air	10 %rh to 95 %rh over a temperature range 5 °C to 50 °C	1.5% of value +0.2 %rh 0.20 °C	HygroGen Chamber
END			



5204
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airbus Operations Ltd
Issue No: 010 Issue date: 23 June 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10⁻⁶·p·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.