


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 7789 Accredited to ISO/IEC 17025:2005	VQS Limited Trading as Valcomm	
	Issue No: 003 Issue date: 14 December 2018	
	Unit 3 Holts Court Threshers Bush Harlow Essex CM17 0NS	Contact: Mr Paul Humphries Tel: +44 (0)1992 441029 / (0)1279 909130 E-Mail: paul.humphries@valcomm.co.uk Website: www.valcomm.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 3 Holts Court Threshers Bush Harlow Essex CM17 0NS Local contact Mr Paul Humphries Tel: +44 (0)1992 441029 +44 (0)1279 909130 Email: paul.humphries@valcomm.co.uk	Temperature and relative humidity.	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Temperature and relative humidity	S



7789

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

VQS Limited Trading as Valcom
Issue No: 003 Issue date: 14 December 2018

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Calibration of temperature loggers and probes in an air chamber	5 °C to 10 °C 10 °C to 50 °C	0.23 °C 0.16 °C	Including temperature probes built in to humidity instruments.	P & S
Temperature indicators and recorders, with temperature sensor(s)	-90 °C to -80 °C -80 °C to 0 °C 0 °C to +100 °C 100 °C to +155 °C +155 °C to +200 °C +200 °C to +300 °C	0.10 °C 0.015 °C 0.025 °C 0.040 °C 0.11 °C 0.35 °C		P
	0.01 °C	0.010 °C	In triple point of water cell	
	-90 °C to +140 °C +140 °C to +250 °C	0.10 °C 0.40 °C		S
Temperature controlled fridges, freezers, incubators, ovens and environmental chambers, including associated recorders, indicators and controllers	-80 °C to 140 °C 140 °C to 250 °C	0.18 °C 0.50 °C	Single or multit-point	S
HUMIDITY				
Relative humidity instruments	5 °C to 10 °C 10 %rh to 90 %rh	2.3 %rh		P & S
	10 °C to 40 °C 5 %rh to 95 %rh	1.60 %rh		
	40 °C to 50 °C 5 %rh to 90 %rh	1.90 %rh		
Humidity controlled chambers, including associated recorders, indicators and controllers	5 °C to 50 °C 10 %rh to 90 %rh	3.0 %rh	Single or multit-point	S
END				