


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>8211</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>DM Systems and Test Limited</p> <p>Issue No: 011 Issue date: 11 April 2022</p>	
	<p>Icon House 3 Iceni Court Icknield Way Letchworth Garden City Hertfordshire SG6 1TN United Kingdom</p>	<p>Contact: Mr Chris Augier Tel: +44 (0)1462 650620 Fax: +44 (0)1462 650622 E-Mail: Info@dplusm.co.uk Website: www.dplusm.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
<p>Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column</p>			
DC VOLTAGE	0 V to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	5.8 $\mu\text{V/V}$ + 1.5 μV 15 $\mu\text{V/V}$ 6 $\mu\text{V/V}$ 11 $\mu\text{V/V}$ 21 $\mu\text{V/V}$	
DC CURRENT	0 to 100 nA 100 nA to 1 μA 1 μA to 10 μA 10 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 20 A	35 $\mu\text{A/A}$ + 60 pA 520 $\mu\text{A/A}$ 175 $\mu\text{A/A}$ 120 $\mu\text{A/A}$ 85 $\mu\text{A/A}$ 85 $\mu\text{A/A}$ 100 $\mu\text{A/A}$ 250 $\mu\text{A/A}$ 650 $\mu\text{A/A}$	
Simulation	20 A to 1000 A	0.20 % + 100 mA	Simulation using a multiturn coil – for the calibration of clampmeters only
DC RESISTANCE	0 Ω to 100 m Ω 100 m Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	660 $\mu\Omega/\Omega$ + 0.15 $\mu\Omega$ 18 $\mu\Omega/\Omega$ + 60 $\mu\Omega$ 15 $\mu\Omega/\Omega$ + 0.6 m Ω 20 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 45 $\mu\Omega/\Omega$ 180 $\mu\Omega/\Omega$ 700 $\mu\Omega/\Omega$ 0.60 %	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC VOLTAGE	<i>1 Hz to 40 Hz</i> 3 μ V to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V <i>40 Hz to 1 kHz</i> 1 μ V to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V <i>1 kHz to 20 kHz</i> 1 μ V to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V <i>20 kHz to 100 kHz</i> 1 μ V to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V <i>100 kHz to 300 kHz</i> 1 μ V to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V <i>300 kHz to 1 MHz</i> 1 μ V to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V	350 μ V/V + 3.6 μ V 550 μ V/V 550 μ V/V 550 μ V/V 700 μ V/V 930 μ V/V 240 μ V/V + 1.5 μ V 320 μ V/V 320 μ V/V 320 μ V/V 470 μ V/V 700 μ V/V 350 μ V/V + 1.5 μ V 400 μ V/V 400 μ V/V 400 μ V/V 470 μ V/V 930 μ V/V 0.58 % + 1.5 μ V 0.12 % 0.12 % 0.12 % 0.17 % 0.37 % 4.6 % + 1.5 μ V 0.47 % 0.47 % 0.47 % 0.58 % 1.2 % + 2.5 μ V 1.30 % 1.30 % 1.90 %	
Generation only	<i>45 Hz to 10 kHz</i> 700 V to 1020 V	0.24 %	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC CURRENT	<p><i>10 Hz to 20 Hz</i> 0.01 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A</p> <p><i>20 Hz to 45 Hz</i> 0.01 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A</p> <p><i>45 Hz to 100 Hz</i> 0.01 μA to 100 μA 100 μA to 0.5 mA 0.5 mA to 1 mA 1 mA to 5 mA 5 mA to 10 mA 10 mA to 50 mA 50 mA to 100 mA 100 mA to 1 A 1 A to 20 A</p> <p><i>100 Hz to 1 kHz</i> 0.01 μA to 100 μA 100 μA to 0.5 mA 0.5 mA to 1 mA 1 mA to 5 mA 5 mA to 10 mA 10 mA to 50 mA 50 mA to 100 mA 100 mA to 1 A 1 A to 20 A</p> <p><i>1 kHz to 5 kHz</i> 0.01 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A</p> <p><i>5 kHz to 20 kHz</i> 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A</p> <p><i>20 kHz to 50 kHz</i> 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A</p>	<p>0.47 % + 50 nA 0.70 % 0.70 % 0.70 % 0.70 %</p> <p>0.18 % + 50 nA 0.41 % 0.41 % 0.41 % 0.42 %</p> <p>700 μA/A + 50 nA 0.31 % 0.13 % 0.31 % 0.12 % 0.18 % 0.12 % 0.065 % 0.075 %</p> <p>700 μA/A + 50 nA 0.27 % 0.088 % 0.27 % 0.082 % 0.27 % 0.082 % 0.065 % 0.075%</p> <p>700 μA/A + 50 nA 0.27 % 0.27 % 0.27 % 0.35 %</p> <p>770 μA/A + 0.40 μA 0.31 % 0.31 % 0.58 %</p> <p>0.47 % + 0.60 μA 0.93 % 0.93 % 1.6 %</p>	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC CURRENT (continued)	50 kHz to 100 kHz 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA	0.64 % + 1.8 µA 2.4 % 2.4 %	
Simulation	45 Hz to 60 Hz 20 A to 500 A	0.50 % + 100 mA	Simulation using a multiturn coil – for the calibration of clampmeters only
CAPACITANCE Generation	DC, 50 Hz to 1 kHz 500 pF to 1.1 nF 1.1 nF to 3.3 nF 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF 330 nF to 1.1 µF 1.1 µF to 3.3 µF DC, 50 Hz to 400 Hz 3.3 µF to 11 µF 11 µF to 33 µF DC, 50 Hz to 100 Hz 33 µF to 110 µF 110 µF to 330 µF	2.9 % 1.7 % 0.94 % 1.4 % 0.66 % 0.61 % 0.66 % 0.73 % 0.76 % 0.79 % 0.94 % 1.2 %	
FREQUENCY	0.1 Hz to 1 Hz 1 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 1 MHz 1 MHz to 10 GHz 10 GHz to 40 GHz 1 MHz and 10 MHz	1 x 10 ⁻³ 1 x 10 ⁻⁴ 1 x 10 ⁻⁵ 1 x 10 ⁻⁶ 2 x 10 ⁻⁷ 2 x 10 ⁻⁸ 2 x 10 ⁻⁹ 2 x 10 ⁻¹⁰ 5 x 10 ⁻¹²	Calibration of stable fixed value oscillators.
TIME INTERVAL	1 s to 12 days	40 nS	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
OSCILLOSCOPE CHARACTERISTICS			
Vertical deflection as a Voltage	0 to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 11 V 11 V to 130 V	580 μ V/V + 52 μ V 0.25 % 0.15 % 800 μ V/V 0.13 %	Into 1 M Ω
	0 to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 6.6 V	0.29 % + 47 μ V 0.48 % 0.35 % 0.30 %	Into 50 Ω
Square wave peak to peak Voltage	0 V to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 11 V 11 V to 130 V	580 μ V/V + 52 μ V 0.31 % 0.19 % 0.13 % 0.16 %	Into 1 M Ω
	0 to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 6.6 V	0.29 % + 53 μ V 0.48 % 0.35 % 0.30 %	Into 50 Ω
Level sine wave flatness for bandwidth with respect to set point value.	50 kHz to 100 MHz 50 kHz to 300 MHz 100 kHz to 2 GHz	2.5 % 3.0 % 3.0 %	3 dB point uncertainty will be reported as a frequency
Resistance	40 Ω to 60 Ω 500 k Ω to 1.5 M Ω	0.12 % 0.12 %	
Capacitance	5 pF to 50 pF	0.90 pF	
Time Markers	2 ns to 20 ns 50 ns to 100 ns 100 ns to 1 s 1 s to 5 s	4.0 μ s/s 150 μ s/s 0.12 % 0.59 %	
Risetime	250 ps Nominal 1 kHz to 10 MHz 250 mV to 500 mV pp	100 ps	
ELECTRICAL SIMULATION			
PRT Indicators			
Resistance thermometer (PT 100)	0 $^{\circ}$ C to 250 $^{\circ}$ C	0.42 $^{\circ}$ C	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
TEMPERATURE INDICATORS BY ELECTRICAL SIMULATION Including Reference junction compensation			
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.42 °C 0.27 °C 0.25 °C 0.34 °C 0.49 °C	
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.36 °C 0.25 °C 0.23 °C 0.26 °C 0.31 °C	
Type E	-250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C	0.61 °C 0.25 °C 0.23 °C 0.25 °C 0.29 °C	
Type B	600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1550 °C 1550 °C to 1820 °C	0.58 °C 0.46 °C 0.41 °C 0.43 °C	
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.51 °C 0.31 °C 0.28 °C 0.27 °C 0.35 °C	
Type R	0 °C to 250 °C 250 °C to 400 °C 400 °C to 1000 °C 1000 °C to 1767 °C	0.72 °C 0.46 °C 0.43 °C 0.50 °C	
Type S	0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1400 °C 1400 °C to 1767 °C	0.61 °C 0.47 °C 0.47 °C 0.57 °C	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
TEMPERATURE INDICATORS BY ELECTRICAL SIMULATION (continued)			
Including Reference junction compensation (continued)			
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.77 °C 0.33 °C 0.25 °C 0.23 °C	
Excluding reference junction compensation			
Type K	-200 °C to -150 °C -150 °C to -100 °C -100 °C to 1370 °C	0.16 °C 0.12 °C 0.10 °C	
Type J	-200 °C to -150 °C -150 °C to -100 °C -100 °C to 1200 °C	0.12 °C 0.10 °C 0.090 °C	
Type T	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 400 °C	0.36 °C 0.15 °C 0.11 °C	
Type N	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 1300 °C	0.77 °C 0.24 °C 0.13 °C	
Type E	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 1300 °C	0.24 °C 0.11 °C 0.080 °C	
Type B	300 °C to 500 °C 500 °C to 1000 °C 1000 °C to 1820 °C	0.76 °C 0.46 °C 0.26 °C	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
TEMPERATURE INDICATORS BY ELECTRICAL SIMULATION (continued) Excluding Reference junction compensation (continued) Type R Type S Measurement of ambient in support of temperature simulation	 -50 °C to 100 °C 100 °C to 300 °C 300 °C to 1768 °C -50 °C to 100 °C 100 °C to 300 °C 300 °C to 1768 °C 18 °C to 30 °C	 0.62 °C 0.32 °C 0.25 °C 0.59 °C 0.32 °C 0.26 °C 0.10 °C	
RF ELECTRICAL MEASUREMENTS RF Power 1 mW reference Flatness at nominal 1 mW (0.8 mW to 1.2 mW) RF Power	 1 mW 50 MHz 100 kHz to 50 MHz 50 MHz to 5 GHz 5 GHz to 15 GHz 15 GHz to 18 GHz 100 KHz to 10 MHz 0.01 mW to 0.12 mW 0.12 mW to 0.8 mW 1.2 mW to 10 mW 10 mW to 100 mW 10 MHz to 18 GHz 0.01 µW to 1.2 µW 1.2 µW to 10 µW 0.01 mW to 0.12 mW 0.12 mW to 0.8 mW 1.2 mW to 10 mW 10 mW to 100 mW	 0.30 % 0.95 % 1.0 % 1.1 % 1.3 % 1.6 % 1.2 % 1.6 % 4.3 % 2.9 % 4.9 % 1.8 % 1.5 % 1.8 % 4.4 %	 For outputs with a VRC of up to 0.05



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
MODULATION (continued) Phase modulation Rate 200 Hz to 20 kHz Deviation	0.4 rad to 4 rad 4 rad to 400 rad	3.7 % 3.5 %	
VOLTAGE REFLECTION COEFFICIENT	<p><i>300 kHz to 3 GHz</i></p> 0 to 0.1 0.1 to 0.2 0.2 to 0.3 0.3 to 0.4 0.4 to 0.5 0.5 to 0.6 0.6 to 0.7 0.7 to 0.8 0.8 to 0.9 0.9 to 1.0 <p><i>3 GHz to 8 GHz</i></p> 0 to 0.1 0.1 to 0.2 0.2 to 0.3 0.3 to 0.4 0.4 to 0.5 0.5 to 0.6 0.6 to 0.7 0.7 to 0.8 0.8 to 0.9 0.9 to 1.0 <p><i>8 GHz to 18 GHz</i></p> 0 dB to 0.1 0.1 dB to 0.2 0.2 dB to 0.3 0.3 dB to 0.4 0.4 dB to 0.5 0.5 dB to 0.6 0.6 dB to 0.7 0.7 dB to 0.8 0.8 dB to 0.9 0.9 dB to 1.0	0.010 0.011 0.012 0.014 0.016 0.019 0.023 0.026 0.031 0.036 0.015 0.016 0.017 0.019 0.021 0.024 0.027 0.031 0.035 0.040 0.015 0.016 0.018 0.021 0.025 0.029 0.034 0.040 0.047 0.054	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
ATTENUATION	300 kHz to 3 GHz		System A
	0 dB to 20 dB	0.080 dB	
	20 dB to 40 dB	0.080 dB	
	40 dB to 60 dB	0.13 dB	
	60 dB to 70 dB	0.33 dB	
	2.5 MHz to 1.3 GHz		System B
	0 dB to 10 dB	0.060 dB	
	10 dB to 40 dB	0.080 dB	
	40 dB to 80 dB	0.080 dB	
	80 dB to 110 dB	0.080 dB	
	1.3 GHz to 8 GHz		0.17 dB
	0 dB to 10 dB	0.17 dB	
10 dB to 40 dB	0.13 dB		
40 dB to 95 dB	0.13 dB		
95 dB to 100 dB	0.25 dB		
8 GHz to 18 GHz		0.24 dB	
0 dB to 10 dB	0.24 dB		
10 dB to 40 dB	0.17 dB		
40 dB to 90 dB	0.17 dB		
90 dB to 100 dB	0.27 dB		
Impedance and Voltage division factor (LISNs and CDNs)			
IMPEDANCE 3 Ω to 250 Ω	9 kHz to 100 kHz	2.5 %	
	100 kHz to 10 MHz	2.5 %	
	10 MHz to 100 MHz	3.5 %	
	100 MHz to 230 MHz	4.5 %	
	230 MHz to 400 MHz	8.5 %	
IMPEDANCE PHASE 3 Ω to 100 Ω	9 kHz to 100 kHz	0.46 °	
	100 kHz to 10 MHz	0.48 °	
	10 MHz to 100 MHz	1.6 °	
INSERTION LOSS / VDF 0 dB to 40 dB	9 kHz to 100 kHz	0.10 dB	
	100 kHz to 10 MHz	0.20 dB	
	10 MHz to 100 MHz	0.30 dB	
	100 MHz to 200 MHz	0.50 dB	
	200 MHz to 400 MHz	0.70 dB	
40 dB to 60 dB	9 kHz to 100 kHz	0.50 dB	
	100 kHz to 10 MHz	0.50 dB	
	10 MHz to 100 MHz	0.60 dB	
	100 MHz to 200 MHz	0.70 dB	
	200 MHz to 400 MHz	0.80 dB	



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
INSERTION LOSS / VDF (continued) 60 dB to 80 dB	9 kHz to 100 kHz 100 kHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 200 MHz 200 MHz to 400 MHz	0.90 dB 0.90 dB 1.0 dB 1.0 dB 1.1 dB	
END			



8211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DM Systems and Test Limited
Issue No: 011 Issue date: 11 April 2022

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$