

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



8239

Accredited to
ISO/IEC 17025:2017

PASS (Portable Appliance Safety Services) Ltd

Issue No: 011 Issue date: 08 June 2021

1 Wilson Street
Thornaby
Stockton-On-Tees
TS17 7AR
United Kingdom

Contact: Mr Ibrahim Ibrahim
Tel: +44 (0) 1642 626148
Fax: +44 (0) 870 143 1869
E-Mail: ibrahim@calibrate.co.uk
Website: www.calibrate.co.uk

Calibration performed at the above address only

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL			All electrical calibrations are performed as a comparison against a reference standard
DC RESISTANCE			
Fixed value sources for the calibration of measuring instruments	0.1 Ω 0.2 Ω 0.3 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω	5.8 m Ω 41 m Ω 41 m Ω 6.0 m Ω 8.2 m Ω 490 $\mu\Omega$ 4.8 m Ω 49 m Ω 980 m Ω 33 Ω 3.1 k Ω 210 k Ω 12 M Ω	
For generating a stimulus that can be applied to measuring instruments also for measuring a stimulus provided by the device being calibrated	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	17 ppm + 59 $\mu\Omega$ 14 ppm + 590 $\mu\Omega$ 12 ppm + 740 $\mu\Omega$ 12 ppm + 6.6 m Ω 12 ppm + 76 m Ω 17 ppm + 3.3 Ω 58 ppm + 130 Ω 580 ppm + 2.8 k Ω 0.58 % + 94 k Ω	
DC VOLTAGE			
Values can be generated for the calibration of measuring instruments	0 mV to 202 mV 202 mV to 1 V 1 V to 2.02 V 2.02 V to 10 V 10 V to 20.2 V 20.0 V to 100 V 100 V to 202 V 202 V to 1020 V	17 ppm + 2.5 μ V 10 ppm + 3.6 μ V 10 ppm + 7.6 μ V 9.8 ppm + 43 μ V 9.8 ppm + 72 μ V 14 ppm + 430 μ V 14 ppm + 720 μ V 14 ppm + 2.8 mV	
For measurement of instrument Outputs	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	5.8 ppm + 390 nV 4.6 ppm + 430 nV 4.6 ppm + 1.2 μ V 6.9 ppm + 54 μ V 15 ppm + 1.3 mV	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DC CURRENT			
Values can be generated for the calibration of measuring instruments	0 μ A to 202 μ A 202 μ A to 1 mA 1 mA to 2.02 mA 2.02m A to 10 mA 10 mA to 20.2 mA 20.2 mA to 100 mA 100 mA to 202 mA 202 mA to 1 A 1 A to 2.02 A 2.02 A to 10 A 10 A to 20.2 A 20.2 A to 30 A	120 ppm + 12 nA 58 ppm + 35 nA 58 ppm + 49 nA 58 ppm + 230 nA 58 ppm + 440 nA 58 ppm + 2.3 μ A 58 ppm + 9.0 μ A 150 ppm + 36 μ A 150 ppm + 100 μ A 350 ppm + 590 μ A 350 ppm + 760 μ A 580 ppm + 4.4 mA	
For measurement of instrument outputs	20 A to 1500 A	0.26 % + 13 mA	Simulation with coil
	0 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 5 A 5 A to 10 A	24 ppm 48 pA 23 ppm +130 pA 23 ppm +950 pA 23 ppm +6.0 nA 23 ppm + 60 nA 40 ppm + 630 nA 130 ppm + 13 μ A 0.23 % + 750 μ A 0.14 % + 2.6 mA 0.27 % + 4.4 mA	
AC VOLTAGE			
Values can be generated for the calibration of measuring instruments	20 mV to 202 mV 10 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 500 kHz	920 ppm + 62 μ V 190 ppm + 62 μ V 230 ppm + 56 μ V 0.12 % + 84 μ V 0.46 % + 2.5 mV	
	202 mV to 2.02 V 10 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	580 ppm + 320 μ V 180 ppm + 280 μ V 240 ppm + 450 μ V 750 ppm + 530 μ V	
	2.02 V to 20.2 V 10 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	580 ppm + 3.0 mV 180 ppm + 2.7 mV 240 ppm + 4.4 mV 690 ppm + 5.3 mV	
	20.2 V to 202 V 30 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 40 kHz	580 ppm + 33 mV 170 ppm + 28 mV 270 ppm + 30 mV 350 ppm + 53 mV	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Generation (cont'd)	202 V to 1020 V 30 Hz to 44 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	640 ppm + 250 mV 230 ppm + 110 mV 290 ppm + 200 mV	
For measurement of instrument outputs	10 μ V to 10 mV 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	230 ppm + 2.6 μ V 350 ppm + 2.6 μ V 0.12 % + 2.6 μ V	
	10 mV to 100 mV 40 Hz to 1 kHz 1 kHz to 20 kHz	82 ppm + 3.3 μ V 160 ppm + 3.3 μ V	
	100 mV to 1 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	82 ppm + 48 μ V 82 ppm + 27 μ V 160 ppm + 28 μ V 350 ppm + 30 μ V 920 ppm + 31 μ V	
	1 V to 10 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	82 ppm + 510 μ V 82 ppm + 270 μ V 160 ppm + 270 μ V 350 ppm + 310 μ V 920 ppm + 320 μ V 0.35 % ppm + 1.2 mV 1.2 % + 3.2 mV	
	10 V to 100 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	230 ppm + 4.9 mV 230 ppm + 2.8 mV 230 ppm + 2.9 mV 400 ppm + 3.2 mV 0.14 % + 3.8 mV	
	100 V to 700 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz	460 ppm + 52 mV 460 ppm + 26 mV 690 ppm + 28 mV	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
AC CURRENT Values can be generated for the calibration of measuring Instruments	20 μ A to 202 μ A 40 Hz to 1 kHz	580 ppm + 180 nA	
	202 μ A to 2.02 mA 40 Hz to 1 kHz	460 ppm + 460 nA	
	2.02 mA 20.2 mA 40 Hz to 1 kHz	400 ppm + 4.6 μ A	
	20.2 mA to 202 mA 40 Hz to 1 kHz	400 ppm + 46 μ A	
	202 mA to 2.02 A 40 Hz to 1 kHz	460 ppm + 550 μ A	
	2.02 A to 20 A 40 Hz to 100 Hz	650 ppm + 6.0 mA	
	20 A to 30 A 40 Hz to 100 Hz	650 ppm + 13 mA	
	20 A to 1500 A 40 Hz to 60 Hz	0.26 % + 13 mA	Simulation using coil
For measurement of instrument outputs	50 nA to 100 μ A 100 Hz to 5 kHz	700 ppm + 46 nA	
	100 μ A to 1 mA 100 Hz to 5 kHz	350 ppm + 230 nA	
	1 mA to 10 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 5 kHz	0.46 % + 2.3 μ A 0.17 % + 2.3 μ A 690 ppm + 2.3 μ A	
	10 mA to 100 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 5 kHz	0.17 % + 24 μ A 690 ppm + 24 μ A 350 ppm + 24 μ A	
	100 mA to 1 A 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 5 kHz	0.46 % + 240 μ A 0920 ppm + 240 μ A 0.12 % + 240 μ A	
	1 A to 3 A 10 Hz to 5 kHz	0.27 % + 580 μ A	
	3 A to 5 A 10 Hz to 5 kHz	0.27 % + 11 mA	
	5 A to 10 A 10 Hz to 5 kHz	0.29 % + 11 mA	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
<p>CAPACITANCE Values can be generated for the calibration of measuring Instruments</p>	<p>At 1 kHz: 1 nF 10 nF 20 nF 50 nF 100 nF 1 μF 10 μF</p>	<p>3.5 pF 31 pF 64 pF 150 pF 290 pF 4.6 nF 69 nF</p>	
<p>INDUCTANCE Values can be generated for the calibration of measuring instruments</p>	<p>At 1 kHz: 1 mH 10 mH 100 mH 1 H</p>	<p>5.9 μH 58 μH 580 μH 5.8 mH</p>	
<p>FREQUENCY Value can be generated for the calibration of measuring instruments</p>	<p>10 MHz reference</p>	<p>1.0 part in to 10^{12}</p>	<p>Frequency may also expressed time; $1/f$ for repetitive signals, in terms of seconds or other units such as RPM.</p>
<p>For generating a stimulus that can be applied to measuring instruments also for measuring a stimulus provided by the device being calibrated</p>	<p>1 Hz to 30 MHz 30 MHz to 4 GHz</p>	<p>1.5 parts in to $10^{12} + 0.60 \mu$Hz 2.0 parts in to 10^{12}</p>	
<p>Rotational speed - Optical Measurement</p>	<p>10 RPM to 99.99 RPM 100 RPM to 999.9 RPM 1000 RPM to 99999 RPM</p>	<p>2.3 RPM 2.4 RPM 3.3 RPM</p>	
<p>Generation</p>	<p>60 RPM to 3000 RPM 3000 RPM to 60000 RPM</p>	<p>0.12 RPM 1.2 RPM</p>	
<p>TEMPERATURE SIMULATION</p>			
<p>PT 100</p>	<p>-200 °C to 800 °C</p>	<p>0.065 °C</p>	
<p>Ambient</p>	<p>17 °C to 23 °C</p>	<p>0.20 °C</p>	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Temperature indicators and calibrators by electrical simulation, reference junction compensation INCLUDED			
Base Thermocouples			
Type E	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1000 °C	0.22 °C 0.21 °C 0.22 °C	
Type J	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1200 °C	0.24 °C 0.21 °C 0.22 °C	
Type K	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1372 °C	0.27 °C 0.22 °C 0.24 °C	
Type N	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1300 °C	0.35 °C 0.24 °C 0.24 °C	
Type T	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 400 °C	0.27 °C 0.23 °C 0.21 °C	
Noble thermocouples			
Type B	600 °C to 1820 °C	0.52 °C	
Type R	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.78 °C 0.57 °C 0.36 °C	
Type S	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.65 °C 0.55 °C 0.40 °C	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Temperature indicators and calibrators by electrical simulation Reference junction compensation EXCLUDED			
Base Thermocouples			
Type E	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1000 °C	0.22 °C 0.19 °C 0.20 °C	
Type J	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1200 °C	0.24 °C 0.19 °C 0.20 °C	
Type K	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1372 °C	0.26 °C 0.21 °C 0.22 °C	
Type N	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 1300 °C	0.35 °C 0.22 °C 0.22 °C	
Type T	-200 °C to -100 °C -100 °C to 0 °C 0 °C to 400 °C	0.26 °C 0.21 °C 0.20 °C	
Noble thermocouples			
Type B	600 °C to 1820 °C	0.52 °C	
Type R	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.77 °C 0.56 °C 0.35 °C	
Type S	-50 °C to 0 °C 0 °C to 400 °C 400 °C to 1767 °C	0.65 °C 0.55 °C 0.39 °C	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
EQUIPMENT FOR IEE 16 TH / 17 TH EDITION WIRING TESTING			
LOOP TESTERS			
AC Resistance at 50 Hz	Nominal applied resistances		
	0.05 Ω	4.7 m Ω	
	0.10 Ω	4.8 m Ω	
	0.21 Ω	4.9 m Ω	
	0.32 Ω	5.1 m Ω	
	0.5 Ω	5.6 m Ω	
	1 Ω	8.6 m Ω	
	5 Ω	31 m Ω	
	10 Ω	59 m Ω	
	100 Ω	580 m Ω	
	1 k Ω	5.9 Ω	
CONTINUITY TESTERS			
DC Resistance	20 m Ω	29 m Ω	
	200 m Ω to 2 Ω	29 m Ω	
	4 Ω	31 m Ω	
	6 Ω	34 m Ω	
	8 Ω	37 m Ω	
	10 Ω	41 m Ω	
	20 Ω	65 m Ω	
	100 Ω	290 m Ω	
	1 k Ω	2.9 Ω	
Continuity Current Measurement	10 mA	1.1 mA	
	100 mA	1.7 mA	
	200 mA	3.1 mA	
	300 mA	4.6 mA	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
INSULATION TESTERS			
DC Resistance	10 k Ω 20 k Ω 30 k Ω 40 k Ω 60 k Ω 100 k Ω 200 k Ω 400 k Ω 600 k Ω 1 M Ω 2 M Ω 3 M Ω 4 M Ω 5 M Ω 6 M Ω 7 M Ω 8 M Ω 9 M Ω 10 M Ω 20 M Ω 30 M Ω 40 M Ω 50 M Ω 60 M Ω 70 M Ω 80 M Ω 90 M Ω 100 M Ω 200 M Ω 400 M Ω 600 M Ω 800 M Ω 1 G Ω 10 G Ω	12 Ω 23 Ω 35 Ω 46 Ω 69 Ω 120 Ω 230 Ω 460 Ω 690 Ω 1.2 k Ω 2.3 k Ω 3.5 k Ω 4.6 k Ω 58 k Ω 69 k Ω 81 k Ω 92 k Ω 100 k Ω 120 k Ω 230 k Ω 350 k Ω 460 k Ω 580 k Ω 690 k Ω 810 k Ω 930 k Ω 1.0 M Ω 1.2 M Ω 2.8 M Ω 5.6 M Ω 8.5 M Ω 11 M Ω 14 M Ω 580 M Ω	
DC Voltage	50 V 100 V 150 V 200 V 250 V 500 V 1000 V	1.1 V 1.5 V 2.0 V 2.5 V 3.0 V 5.9 V 12 V	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
EARTH BOND TESTERS			
AC Resistance at 50 Hz	Nominal applied resistance 0.04 Ω 0.1 Ω 0.15 Ω 0.27 Ω 0.38 Ω 0.55 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1 k Ω	4.7 m Ω 4.8 m Ω 4.8 m Ω 5.0 m Ω 5.2 m Ω 5.8 m Ω 7.8 m Ω 30 m Ω 59 m Ω 580 m Ω 5.8 Ω	
AC Current at 50 Hz	100 mA 200 mA 400 mA 4 A 8 A 10 A 20 A	7.3 mA 7.9 mA 9.9 mA 100 mA 160 mA 190 mA 440 mA	
LEAKAGE TESTERS			
DC Current	2 mA 5 mA 10 mA	36 μ A 82 μ A 130 μ A	
RCD TESTERS			
RCD Trip Time	20 ms 40 ms 100 ms 200 ms 390 ms 900 ms	680 μ s 680 μ s 680 μ s 680 μ s 680 μ s 8.1 ms	
RCD Trip Current at 50 Hz	10 mA 30 mA 90 mA 100 mA 110 mA 150 mA 300 mA 1 A 2 A	620 μ A 1.7 mA 5.2 mA 5.8 mA 6.4 mA 17 mA 17 mA 58 mA 120 mA	
AC Voltage Source at 50 Hz	100 V 200 V 230 V 300 V 400 V	0.37 V 0.45 V 0.65 V 0.82 V 0.99 V	
Line Voltage Measurement	200 V to 260 V	2.4 V	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
PRESSURE			Methods consistent with EURAMET CG17. Calibration of devices with an electrical output may be undertaken.
Gas Pressure (Gauge)			
Calibration of pressure indicating instruments and gauges	-95 to -1.5 kPa 1.5 kPa to 20 kPa 20 kPa to 2.5 MPa 2.5 MPa to 11.1 MPa	67 ppm 60 ppm 44 ppm 50 ppm	Calibration using deadweight testers.
Calibration of pressure indicating instruments and gauges	-95 to 21 MPa	61 ppm + 2.0 kPa	Calibration using pressure controllers.
Gas Pressure (Absolute)			
Calibration of pressure indicating instruments and gauges	3.5 kPa to 100 kPa 3.5 kPa to 200 kPa 3.5 kPa to 800 kPa 100 kPa to 2.1 MPa 3.5 kPa to 7 MPa 100 kPa to 41.4 MPa	75 ppm + 14 Pa 46 ppm + 13 Pa 180 ppm + 24 Pa 25 ppm + 90 Pa 100 ppm + 210 Pa 74 ppm + 1.7 kPa	Calibration using pressure controllers.
Hydraulic Pressure (Gauge)			
Calibration of pressure indicating instruments and gauges	0.6 MPa to 6.0 MPa 6 MPa to 70 MPa 70 MPa to 138 MPa	80 ppm 64 ppm 100 ppm	Calibration using deadweight testers.
TEMPERATURE			
Sensors with indicators			
	-95 °C to -50 °C -50 °C to 140 °C 140 °C to 660 °C	0.10 °C 0.050 °C 0.10 °C	Calibrations performed in a Metal block
	-80 °C to 0 °C 0 °C to 100 °C	0.025 °C 0.020 °C	Calibrations performed in liquid bath
	0.01 °C	0.0050 °C	Triple point of water
Metal block calibrators and portable liquid baths	-95 °C to -660 °C	Uncertainty as for sensor and indicator	
HUMIDITY			
Relative humidity			
	At 0 °C 5 %rh 50 %rh 90 %rh	2.5 %rh 2.7 %rh 3.2 %rh	
	At 23 °C 5 %rh to 30 %rh 30 %rh to 60 %rh 60 %rh to 80 %rh 80 %rh to 95 %rh	1.3 %rh 1.5 %rh 1.7 %rh 1.9 %rh	



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
HUMIDITY (continued) Relative humidity	At 60 °C 5 %rh 50 %rh 90 %rh	2.4 %rh 2.5 %rh 2.8 %rh	
TEMPERATURE IN AIR	0 °C to 23 °C 23 °C to 60 °C	0.20 °C 0.40 °C	
END			



8239
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

PASS (Portable Appliance Safety Services) Ltd
Issue No: 011 Issue date: 08 June 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
- As an explicit function of the measurand or of a parameter (see below).
- As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
- As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %- p + (0.12 $\cdot 10^{-6} \cdot p \cdot 10^{-6}$) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means $1.5 \cdot 0.01 \cdot i$, where i is the instrument indication.