


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>8380</b> Accredited to <b>ISO/IEC 17025:2017</b>	<b>C&amp;M Scientific Ltd</b>	
	Issue No: 012    Issue date: 16 August 2021	
	<b>1 Brewster Square</b> <b>Brucefield Industrial Estate</b> <b>Livingston</b> <b>EH54 9BJ</b>	<b>Contact: Colin Ramsay</b> <b>Tel: +44 (0) 1506 463734</b> <b>E-Mail: c.ramsay@cmscientific.co.uk</b> <b>Website: www.cmscientific.co.uk</b>
<b>Calibration performed by the Organisations at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> 1 Brewster Square Brucefield Industrial Estate Livingston EH54 9BJ	<b>Local contact</b> Tel: +44 (0)1506 463734 Email: info@cmscientific.co.uk Website: www.cmscientific.co.uk	Temperature indicators Temperature controlled chambers Rotational speed of centrifuges Time Interval Electrical simulation of temperature Carbon Dioxide content	Lab

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Temperature indicators Temperature and humidity controlled chambers Rotational Speed of centrifuges Time interval Electrical simulation of temperature Carbon Dioxide content	Site



8380

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**C&M Scientific Ltd**

**Issue No: 012 Issue date: 16 August 2021**

**Calibration performed by the Organisation at the locations specified**

**Calibration and Measurement Capability (CMC)**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<p>TEMPERATURE</p> <p>Calibration by comparison</p> <p>Temperature sensors with indicators or recorders</p>	<p>-196 -90 °C to 0 °C -40 °C to 0 °C 0 °C to 125 °C 125 °C to 300 °C</p>	<p>0.30 °C 0.15 °C 0.11 °C 0.13 °C 0.31 °C</p>	<p>In liquid nitrogen In metal block bath</p>	<p>Lab and site</p>
<p>Thermal performance of centrifuges</p>	<p>0°C to 40°C</p>	<p>0.50 °C</p>		<p>Lab and Site</p>
<p>Temperature controlled, incubators, ovens, environmental chambers, refrigerators, freezers, centrifuges and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors)</p>	<p>-90 °C to +300 °C</p>	<p>0.35 °C</p>	<p>Single point probes.  Single and multipoint monitoring probes Time dependent temperature profiling</p>	<p>Lab and Site</p>
<p>RELATIVE HUMIDITY</p> <p>Calibration by comparison</p> <p>Temperature and humidity controlled, incubators, environmental chambers,</p>	<p>15 %rh to 95 %rh for the temperature range 10 °C to 70 °C</p>	<p>2.7 %rh</p>		<p>Site</p>
<p>TIME INTERVAL</p> <p>Calibration by comparison</p> <p>Timers</p>	<p>5 s to 24 hr</p>	<p>3 s</p>		<p>Lab and site</p>
<p>Rotational speed Centrifuges</p> <p>Calibration by comparison</p>	<p>500 RPM to 20000 RPM</p>	<p>1.8 RPM</p>		<p>Lab and site</p>



8380

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**C&M Scientific Ltd**  
**Issue No:** 012    **Issue date:** 16 August 2021

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL CALIBRATION OF TEMPERATURE INDICATORS AND SIMULATORS  Calibration by comparison  PRT simulation (PT100)	-200 °C to +300 °C	0.12 °C		Lab and site
Base metal thermocouples Type T only	-200 °C to +200 °C	0.75 °C		
GAS CONTENT				Lab and site
Carbon Dioxide	0% to 20%	0.40%	Gas Analyser	
Carbon Dioxide	0% to 20%	0.75%	Single point calibration in the centre of the incubator	
END				



8380

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**C&M Scientific Ltd**  
**Issue No:** 012    **Issue date:** 16 August 2021

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$