


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 8678 Accredited to ISO/IEC 17025:2005	Henderson Biomedical Ltd	
	Issue No: 003 Issue date: 12 December 2016	
	57 Kangley Bridge Road Lower Sydenham London SE26 5BA	Contact: Giles Mahon Tel: 020 8663 4610 Fax: 020 8659 7491 E-Mail: info@henderson-biomedical.co.uk Website: www.henderson-biomedical.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Worsley Bridge Rd Lower Sydenham London SE26 5AZ	Contact Giles Mahon Tel: 020 8663 4610 Fax: 020 8659 7491 Email: info@henderson-biomedical.co.uk Website: www.henderson-biomedical.co.uk	<u>Calibration:</u> Temperature Rotational speed Force Timers	Lab

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Customer's sites or premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Contact Giles Mahon Tel: 020 8663 4610 Fax: 020 8650 8408 Email: info@henderson-biomedical.co.uk Website: www.henderson-biomedical.co.uk	<u>Calibration:</u> Temperature Rotational speed Timers	Site



8678

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Henderson Biomedical Ltd
Issue No: 003 Issue date: 12 December 2016

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE Temperature controlled, incubators, ovens, refrigerators, freezers, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors)	-20 °C to +250 °C	1.9 °C		Lab & site
Sealing temperature of pouch sealers	100 °C to 220 °C	1.9 °C		Lab
CENTRIFUGE Rotational Speed	1000 rpm to 20000 rpm	4.2 rpm		Lab & site
Temperature	4 °C to 40 °C	1.9 °C		
Timers	0 s to 3600 s	1.0 s		
FORCE Sealing force of pouch sealers	50 N to 150 N	0.8 N		Lab
END				