


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>8807</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>DJB Labcare Limited</p> <p>Issue No: 002 Issue date: 01 March 2018</p>	
	<p>Unit 20 Interchange Park Newport Pagnell MK16 9QS United Kingdom</p>	<p>Contact: Ms Caitriona Hurd Tel: +44 (0) 1908 612598 E-Mail: caitriona@djblabcare.co.uk Website: https://www.djblabcare.co.uk/</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
<p>Address Unit 20 Interchange Park Newport Pagnell MK16 9QS United Kingdom</p>	<p>Local contact Ms Caitriona Hurd Tel: +44 (0) 1908 612598 E-Mail: jude@djblabcare.co.uk Website: www.djblabcare.co.uk/</p>	<p>Rotational speed Temperature Time Carbon Dioxide content</p>	<p>Lab</p>

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p>Rotational speed Temperature Time Carbon Dioxide content</p>	<p>Site</p>



8807

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines, Middlesex, UK

DJB Labcare Limited

Issue No: 002 Issue date: 01 March 2018

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Thermal performance of centrifuges	0°C to 40°C	0.90 °C		Lab and site
Incubators	0°C to 40°C	0.86 °C	Single point calibration in the centre of the incubator	Lab and site
ROTATIONAL SPEED				
Centrifuges	100 rpm to 20000 rpm	2.60 rpm		Lab and site
TIME INTERVAL				
Timers	0 s to 3600 s	0.45 s		Lab
Timers associated with centrifuges	0 s to 3600 s	0.90 s		Lab and site
Carbon Dioxide content	5%	0.50%	Single point calibration in the centre of the incubator	Lab and site
END				