

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



8807

Accredited to
ISO/IEC 17025:2017

DJB Labcare Limited

Issue No: 007 Issue date: 04 December 2023

Unit 12
Cromwell Business Centre
Newport Pagnell
Buckinghamshire
MK16 9QS

Contact: Jude Huckle
Tel: +44 (0) 1908 612598
E-Mail: jude@djblabcare.co.uk
Website: <https://www.djblabcare.co.uk/>

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 12 Cromwell Business Centre Newport Pagnell Buckinghamshire MK16 9QS United Kingdom	Local contact Jude Huckle Tel: +44 (0) 1908 612598 E-Mail: jude@djblabcare.co.uk Website: www.djblabcare.co.uk/	Rotational speed Temperature Time Carbon Dioxide content

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Rotational speed Temperature Time Carbon Dioxide content	Site



8807

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DJB Labcare Limited**Issue No:** 007 **Issue date:** 04 December 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Thermal performance of centrifuges	0°C to 40°C	0.90 °C		Lab and site
Incubators	0°C to 40°C	0.86 °C	Single point time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	Lab and site
ROTATIONAL SPEED				
Centrifuges	100 rpm to 20000 rpm	2.60 rpm	Using a reference tachometer	Lab and site
TIME INTERVAL				
Timers	0 s to 3600 s	0.45 s	Using a reference timer	Lab
Timers associated with centrifuges	0 s to 3600 s	0.90 s		Lab and site
CARBON DIOXIDE				
Carbon Dioxide content	5%	0.50%	Calibration in the centre of the incubator or from the sample port	Lab and site
END				



8807

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DJB Labcare Limited

Issue No: 007 **Issue date:** 04 December 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.
Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$