


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION 8957</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>NDT Global Services Limited</p> <p>Issue No: 006 Issue date: 21 July 2021</p>	
	<p>Opus Park Lockheed Close Preston Farm Industrial Estate Stockton-On-Tees TS18 3BP United Kingdom</p>	<p>Contact: Ian Craig Tel: +44 (0)1642 555575 E-Mail: lab@ndtgsl.co.uk Website: www.ndtgsl.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTRICAL VERIFICATION of ULTRASONIC FLAW DETECTION EQUIPMENT	As BS EN ISO 22232-1:2020 Group 2 tests and including the following calibrations and quantities:		
	Pulser Voltage Pulser Risetime Pulse duration	14 % 6.0 % 3.8 %	For instruments designed to comply with BS EN 12668-1:2010, the pulse width is determined at 10 % of the pulse height, otherwise the 50 % points are used.
	Frequency response <i>0.2 MHz to 30 MHz</i>	3.4 % at -3 dB point	For instruments designed to comply with BS EN 12668-1:2010, the centre frequency f_0 is calculated using $f_0 = \sqrt{f_u \times f_l}$, otherwise the expression $f_0 = (f_u + f_l)/2$ is used.
	Equivalent input noise	15 % of screen height	Using Method B as described in Section 9.4.3.3 of BS EN ISO 22232-1:2020.
	Calibrated attenuator, <i>0 dB to 70 dB</i> Vertical Linearity	0.26 dB to 0.90 dB 1.5 % of screen height	
CALIBRATION OF ULTRASONIC TEST BLOCKS			
Linear dimensions	0 mm to 25 mm	2.7 μ m	Using micrometer
	25 mm to 50 mm	2.9 μ m	Using micrometer
	0 mm to 200 mm	25 μ m	Using digital caliper
	0 mm to 300 mm	19 μ m	Using height gauge
Hole diameter	0.22 mm to 7.7 mm	24 μ m	Using pin gauges
	7.7 mm to 100 mm	25 μ m	Using digital caliper
Hole centre to plate edge	Hole diameter to 300 mm	29 μ m	Using height gauge
	Hole diameter 200 mm	32 μ m	Using digital caliper



8957

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NDT Global Services Limited

Issue No: 006 Issue date: 21 July 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
CALIBRATION OF ULTRASONIC TEST BLOCKS (continued)			
Hole depth	0 mm to 200 mm 0.5 mm to 7.7 mm	32 μ m 30 μ m	Using digital caliper Using pin and height gauge
Slot width	8 mm to 50 mm 0.22 mm to 30 mm	28 μ m 27 μ m	Using digital caliper Using height gauge
Slot depth	0 mm to 200 mm 0 mm to 10 mm	32 μ m 17 μ m	Using digital caliper Using height gauge
Determination of slot centre line	1 mm to 300 mm	27 μ m	
END			



8957

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NDT Global Services Limited

Issue No: 006 Issue date: 21 July 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V: $0.0025 \% + 5.0 \mu\text{V}$:

Over the range 100 mV to 1 V, the CMC is $0.0025 \% \cdot V + 5.0 \mu\text{V}$, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: $0.0036 \% + 0.12 \text{ ppm/MPa} + 4.0 \text{ Pa}$

Over the range 0.5 MPa to 140 MPa, the CMC is $0.0036 \% \cdot p + (0.12 \cdot 10^{-6} \cdot p \cdot 10^{-6}) + 4.0 \text{ Pa}$, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means $1.5 \cdot 0.01 \cdot i$, where i is the instrument indication.