


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>9149</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Mech Metrology & Power Tools Ltd</p> <p>Issue No: 004 Issue date: 17 August 2021</p>	
	<p>11 Fareham Enterprise Centre Hackett Way Fareham PO14 1TH United Kingdom</p>	<p>Contact: Mr Russell Johnston Tel: +44 (0)1329 511 777 Fax: +44 (0)1329 511 771 E-Mail: sales@mechmetrology.co.uk Website: www.mechmetrology.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
<p>RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED</p>			
LENGTH			<p>NOTES All linear measurements may be reported in inch units.</p>
Plain ring gauges (parallel) and setting standards	3 to 50 50 to 100 diameter 100 to 150	1.5 2.3 on diameter 3.1	Using length measuring machine and standards.
Plain gap gauges (parallel)	3 to 100	3.0	Using length measuring machine and/or end standards.
Length gauges, flat and spherical ended	As BS 870:2008 25 to 300	1.0 + (8.0 x length in m)	Using length measuring machine and end standards.
Screw ring gauges (parallel)	2 to 12.7	See Note	Functional test of size using check plugs.
Feeler Gauges	As BS 957:2008 0.02 to 1.00	2.0	Using length measuring machine.
<p>END</p>			



9149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Mech Metrology & Power Tools Ltd
Issue No: 004 Issue date: 17 August 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$