


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|   |  |  |
|---|--|--|
|  <p><b>9361</b></p> <p>Accredited to<br/><b>ISO/IEC 17025:2017</b></p> | <p><b>The Validation Centre (TVC) Ltd</b></p> <p>Issue No: 008 Issue date: 24 August 2021</p>                    |  |
|   | <p>Unit 15<br/>Brinell Way<br/>Harfreys Industrial Estate<br/>Great Yarmouth<br/>NR31 0LU<br/>United Kingdom</p> | <p>Contact: Mr Kim Hastings<br/>Tel: +44 (0)1493 443800<br/>Fax: +44 (0)1493 443900<br/>E-Mail: KimHastings@tvcalx.co.uk<br/>Website: www.tvcalx.co.uk</p> |
| <p>Calibration performed by the Organisation at the locations specified below</p>   |  |  |

### Laboratory location:

| Location details   | Activity  | Location code               |
|--|---|-----------------------------|
| <p><b>Address</b></p> <p>Unit 15<br/>Brinell Way<br/>Harfreys Industrial Estate<br/>Great Yarmouth<br/>NR31 0LU<br/>United Kingdom</p> <p><b>Local contact</b><br/>Mr Kim Hastings<br/>Tel: +44 (0)1493 443800<br/>Fax: +44 (0)1493 443900<br/>Email: KimHastings@tvcalx.co.uk<br/>Website: www.tvcalx.co.uk</p> | <p>Electrical verification of ultrasonic flaw detection equipment in accordance with the Group 2 tests described in:<br/>BS EN 12668-1:2010<br/>BS EN ISO 22232-1:2020</p> <p>Electrical verification of ultrasonic thickness measuring equipment in accordance with the Group 2 tests described in BS EN 15317:2013.</p> | <p>Permanent Laboratory</p> |

### Site activities performed away from the location listed above:

| Location details   | Activity  | Location code           |
|--|---|-------------------------|
| <p>Calibrations may be performed in suitable areas within the customers' premises.</p> <p>The customers' premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p> <p>Contact as above</p> | <p>Electrical verification of ultrasonic flaw detection equipment in accordance with the Group 2 tests described in:<br/>BS EN 12668-1:2010<br/>BS EN ISO 22232-1:2020</p> <p>Electrical verification of ultrasonic thickness measuring equipment in accordance with the Group 2 tests described in BS EN 15317:2013.</p> | <p>Site Calibration</p> |



9361  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**The Validation Centre (TVC) Ltd**  
**Issue No: 007 Issue date: 24 August 2021**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge   | Range   | Expanded Measurement<br>Uncertainty ( $k = 2$ )  | Remarks | Location<br>code                          |
|--|---|--|---------|---|
| <p>ELECTRICAL VERIFICATION of ULTRASONIC THICKNESS MEASURING EQUIPMENT</p> <p>As BS EN 15317:2013 Group 2 tests and including the following calibrations and quantities:</p> | <p>9.6: Battery warning and cut off voltage</p> <p>9.7: Battery warning and cut off current</p> <p>9.9: Pulse repetition frequency</p> <p>9.10: Pulse Voltage <math>V_{50}</math></p> <p>9.10: Pulse Risetime <math>t_r</math></p> <p>9.10: Pulse Duration <math>t_d</math></p> <p>9.12: Maximum and minimum thickness using gauges</p>   | <p>0.28 %</p> <p>0.24 %</p> <p>1.3 %</p> <p>3.5 %</p> <p>0.90 ns</p> <p>1.3 ns</p> <p>0.20 mm</p>  |         | Permanent Laboratory and Site Calibration |
| <p>ELECTRICAL VERIFICATION of ULTRASONIC FLAW DETECTION EQUIPMENT</p> <p>As BS EN 12668-1:2010 and including the following calibrations and quantities:</p>                  | <p>Stability after warm up (height)</p> <p>Stability after warm up (width)</p> <p>Jitter - screen height</p> <p>Jitter - screen width</p> <p>Stability against voltage variation (height)</p> <p>Stability against voltage variation (width)</p> <p>Pulser Voltage</p> <p>Pulser Risetime</p> <p>Pulser Reverberation</p> <p>Pulse duration</p> <p>Frequency response<br/>0.2 MHz to 30 MHz</p> <p>Equivalent input noise</p> <p>Calibrated attenuator</p> <p>Vertical Linearity</p> <p>Linearity of timebase</p> | <p>1.2 % of screen height</p> <p>1.2 % of screen width</p> <p>1.2 % of screen height</p> <p>1.2 % of screen width</p> <p>1.2 % of screen height</p> <p>1.2 % of screen width</p> <p>3.5 %</p> <p>0.90 ns</p> <p>2.4 % of pulser voltage</p> <p>1.3 ns</p> <p>4.1 % at -3 dB point</p> <p><math>5.4 \times 10^{-9} V/\sqrt{Hz}</math></p> <p>0.27 dB</p> <p>0.23 % of screen height</p> <p>0.15 % of screen width</p> |         |   |



9361  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**The Validation Centre (TVC) Ltd**  
**Issue No: 007 Issue date: 24 August 2021**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ )                                    | Remarks   | Location<br>code                          |
|--|--|--|---|---|
| ELECTRICAL VERIFICATION of ULTRASONIC FLAW<br>DETECTION EQUIPMENT (continued)<br><br>As BS EN ISO 22232-1:2020<br>Group 2 tests and including<br>the following calibrations and<br>quantities: | Pulser Voltage                                 | 3.5 %  | For instruments designed<br>to comply with<br>BS EN 12668-1:2010, the<br>centre frequency $f_0$ is<br>calculated using<br>$f_0 = \sqrt{(f_u \times f_l)}$ , otherwise the<br>expression<br>$f_0 = (f_u + f_l)/2$ is used. | Permanent Laboratory and Site Calibration |
|  | Pulser Risetime                                | 0.91 ns  |   |   |
|  | Pulse duration                                 | 1.3 ns   |   |   |
|  | Frequency response<br><i>0.2 MHz to 30 MHz</i> | 4.1 % at -3 dB point   |   |   |
|  | Equivalent input noise                         | 15 %   |   |   |
|  | Equivalent input noise                         | $5.4 \times 10^{-9} \text{ V}/\sqrt{\text{Hz}}$                                    |   |   |
|  | Calibrated attenuator                          | 0.27 dB  |   |   |
| Gain linearity   | 0.27 dB  | Using Method A as<br>described in Section<br>9.4.3.2 of<br>BS EN ISO 22232-1:2020. |   |   |
| Vertical Linearity   | 0.23 % of screen height                        |  | Using Method B as<br>described in Section<br>9.4.3.3 of<br>BS EN ISO 22232-1:2020.  |   |
| END  |  |  |   |   |



9361  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**The Validation Centre (TVC) Ltd**  
**Issue No:** 007    **Issue date:** 24 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$