


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> CALIBRATION <b>9474</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Next Control Systems Ltd trading as Tutela Monitoring Systems</b>	
	<b>Issue No: 003    Issue date: 20 August 2018</b>	
	<b>93 Fleet Road Fleet GU51 3PJ</b>	<b>Contact: Mr Tim Bartholomew Tel: +44 (0) 1252 406398 E-Mail: tim@nextcontrols.com Website: www.tutelamedical.com</b>
<b>Calibration performed by the Organisation at the locations specified</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> United Kingdom  <b>Local contact</b> Mr Phil Douglas Tel: +44 (0) 1252 406398 E-Mail: pdouglas@tutelamedical.com	Temperature	P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, e.g. Hospitals, Laboratories and Manufacturing & Processing Plants The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Temperature Relative Humidity	S



Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Next Control Systems Ltd trading as Tutela Monitoring Systems**

**Issue No: 003 Issue date: 20 August 2018**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>TEMPERATURE</b> Calibration of temperature indicators with sensors	-196 °C	0.15 °C	Calibration in Liquid Nitrogen	P
	- 80 °C to + 50 °C	0.16 °C	Calibration in a block calibrator	P
	-196 °C - 80 °C to + 50 °C	0.30 °C 0.20 °C	Single point measurement in customers environment	S
	-25 °C to 50 °C	0.23 °C	Calibration in a block calibrator	S
<b>HUMIDITY</b> Relative humidity instruments	10 %rh to 95 %rh <i>Temperature range 21 °C to 27 °C</i>	3.0 %rh	Single point measurement in customers environment  Uncertainty will depend on stability of the ambient conditions	S
END				