

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 9791 Accredited to ISO/IEC 17025:2017	Validair Monitoring Solutions Limited	
	Issue No: 001 Issue date: 15 March 2019	
	5 Beauchamp Business Centre Sparrowhawk Close Enigma Business Park Malvern WR14 1GL	Contact: Mike Pemberton Tel: +44 (0)1684 576452 E-Mail: mike-pemberton@validair.com Website: www.validair.com
Calibration performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 5 Beauchamp Business Centre Sparrowhawk Close Enigma Business Park Malvern WR14 1GL Local contact Mike Pemberton	Temperature	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer Local contact Mike Pemberton	Temperature	S



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Validair Monitoring Solutions Limited

Issue No: 001 Issue date: 15 March 2019

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE Temperature indicators and recorders, with temperature sensor(s)	-90 °C to -30 °C -30 °C to 0 °C 0 °C to 100 °C	0.30 °C 0.15 °C 0.10 °C	Calibration by comparison with a reference standard in a controlled environment	P & S
	-30 °C to 0 °C 0 °C to 50 °C	0.50 °C 0.50 °C	Single point calibration by comparison with a reference standard in the prevailing environment uncertainty will be dependent on the stability of the environment	P & S
END				