


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

	APEX Metrology Ltd	
	Issue No: 001	Issue date: 13 October 2017
	Unit 3 The Enterprise Centre 1 Dryden Road Loanhead EH20 9LZ United Kingdom	Contact: Mr Brian Young Tel: +44 (0)131 448 2111 E-Mail: brian@apexmetrology.com Website: http://www.apexmetrology.com
Calibrations performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 3 The Enterprise Centre 1 Dryden Road Loanhead EH20 9LZ	Local contact Mr Brian Young	Dimensional A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises	Contact: Mr Brian Young	Dimensional B



9817
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

APEX Metrology Ltd
Issue No: 001 Issue date: 13 October 2017

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
MEASURING INSTRUMENTS AND MACHINES				
Performance verification of co-ordinate measuring machines	ISO 10360-2:2009 0 mm to 1500 mm (longest diagonal using end standards)	1.1 + (1.7 x length in m)		A & B
	ISO 10360-5:2010 - single stylus probing test 10 to 50 (test sphere diameter)	0.70		
Receiver and position gauges, jigs and fixtures	0 to 853 mm x 853 mm x 859 mm Angle – 0° to 360°	9.0 + (9.2 x length in m) 0.050°		A
END				