

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|--|
|  UKAS CALIBRATION 9944 Accredited to ISO/IEC 17025:2017 | Road Safety Support Limited Issue No: 005 Issue date: 15 July 2023 | |
| | Top Floor, Claridon House London Road Stanford Le Hope Essex SS17 0JU | Contact: Steve Callaghan Tel: +44 (0)1277 634900 E-Mail: steve.callaghan@roadsafetysupport.co.uk Website: www.roadsafetysupport.co.uk |

Calibration performed by the Organisation at the locations specified

Locations covered by the organisation and their relevant activities

Site activities performed away from the location listed above:

| Location details | Activity | Location code |
|---|----------------------------------|---------------|
| Any suitable test track or public road. Proposed sites will be the subject of an initial survey for suitability and for safety reasons. | All activities within the scope. | S |



9944

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Road Safety Support Limited
Issue No: 005 Issue date: 15 July 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|-------------------|--|---|--|
| Speedmeters, including radar, laser, ANPR, sub-surface sensors (loop), traffic signal cameras, manual types and in-vehicle speedometers. Results are obtained by comparison with a reference speed measuring system. | 20 mph to 190 mph | 0.072 mph | The CMC is based on a speedometer with a resolution of 0.01 mph or 0.01 km/h. Devices with more limited resolution will attract higher uncertainties. Roadside speedometers are tested for compliance with the accuracy requirements of Traffic Light and Prohibited Lane Enforcement Camera Handbook v2.0 (19 December 2022), published by the Home Office. The following functional tests may also be reported in calibration certificates: High Registration Plate Test (ANPR systems) Multiple Vehicle Test | S S |
| END | | | | |



9944

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Road Safety Support Limited
Issue No: 005 Issue date: 15 July 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$