


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0259 Accredited to ISO/IEC 17025:2017	Health and Safety Executive for NI – Scientific Services	
	Issue No: 046 Issue date: 05 May 2022	
	HSENI Scientific Services 83 Ladas Drive Belfast Northern Ireland BT6 9FR	Contact: Morgan Naylor Tel: +44 (0)28 90 546938 E-Mail: morgan.naylor@hseni.gov.uk Website: www.hseni.gov.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address HSENI Scientific Services 83 Ladas Drive Belfast Northern Ireland BT6 9FR	Local contact Evette Boyes Tel: +44 (0)28 90 546843 E-Mail: evette.boyes@hseni.gov.uk Website: www.hseni.gov.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



0259

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Health and Safety Executive for NI – Scientific Services

Issue No: 046 Issue date: 05 May 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method N705S based on HSG 248	A, B
	Fibre counting	Documented In-House Method N33A, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method N706S based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method N48A using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
ASBESTOS CONTAINING MATERIALS	Water Absorption test: Method to differentiate between asbestos insulating boards and asbestos cement sheets	Documented In-House Method QP4 based on Appendix 3 Health and Safety Executive, Asbestos: The Analysts' Guide (HSG 248: 2021)	A
AIRBORNE INDUSTRIAL POLLUTANTS	<u>Health and Hygiene</u>		
	Determination of Inhalable/Respirable Aerosols (Weighing of filters)	In-House Procedure N629S HSE-MDHS 14/4 (June 2014)	A
	Sampling of Inhalable and Respirable aerosols	HSE-MDHS 14/4 (June 2014) In-House Procedure N629S	B



0259

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Health and Safety Executive for NI – Scientific Services

Issue No: 046 Issue date: 05 May 2022

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
AIRBORNE INDUSTRIAL POLLUTANTS (cont'd)	<u>Health and Hygiene</u> (cont'd) Assessment of <i>workplace control using measurement of</i> : <ul style="list-style-type: none">• Face velocity• Capture Velocity• Transport Velocity• Smoke Visualisation techniques	In-House Procedure N823A	B
END			