## **Schedule of Accreditation**

## **United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



## **TUV SUD Limited** Trading as TÜV SÜD National Engineering Laboratory & **National Engineering Laboratory**

Issue No: 042 Issue date: 01 March 2022

Accredited to

ISO/IEC 17025:2017

0432

**Scottish Enterprise Technology Park** 

**East Kilbride** 

**Napier Building** 

Glasgow

**G75 0QF** 

Tel: +44 (0)1355-59 + Extension below

Fax: +44 (0)1355-272999 E-Mail: sales@tuvnel.com Website: www.tuvsud.com

Testing performed by the Organisation at the locations specified below

#### Locations covered by the organisation and their relevant activities

#### **Laboratory locations:**

Location details		Activity	Location code
Address East Kilbride Glasgow Scotland G75 0QU	Local contact Mr John Galbraith (Ext: 3749) Mr Muir Porter (Ext: 3833)	Stack Emissions Testing (Support Functions: Quality System Quality Audit Administration)	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Sites requiring Stack Emissions Testing arrangements between the laboratory and the customer.	Stack Emissions Testing	В

Assessment Manager: JG2 Page 1 of 5



### **Schedule of Accreditation** issued by

# United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

# TUV SUD Limited Trading as TÜV SÜD National Engineering Laboratory & National Engineering Laboratory

Issue No: 042 Issue date: 01 March 2022

Testing performed by the Organisation at the locations specified

#### **DETAIL OF ACCREDITATION**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
TESTING OF STACK EMISSIONS TO ATMOSPHERE	Sampling with subsequent analysis by an ISO/IEC 17025 Accredited Laboratory	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and BS EN 15259:2007	
	Total Particulate Matter	BS EN 13284-1:2017 (WI/EM/962)	В
	Hydrogen Chloride	BS EN 1911:2010 (WI/EM/973)	В
	Hydrogen Fluoride	PD CEN/TS 17340:2020 (WI/EM/1271)	В
	Sulphur Dioxide	BS EN 14791:2017 (WI/EM/1275)	В
	Ammonia	BS EN ISO 21877:2019 (WI/EM/0043)	В
	Metals	BS EN 14385:2004 (WI/EM/982) EA MID 14385 (WI/EM/1272)	В
	Mercury	BS EN 13211:2001 (WI/EM/977)	В
	Speciated VOCs: Mercaptans Amines and Amides Phenols Cresols Carboxylic Acids Aldehydes Hydrogen sulphide Carbon disulphide	PD CEN/TS 13649:2014 (WI/EM/963) PD CEN/TS 13649:2014 (WI/EM/963)	В

Assessment Manager: JG2 Page 2 of 5



### **Schedule of Accreditation** issued by

# United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

# TUV SUD Limited Trading as TÜV SÜD National Engineering Laboratory & National Engineering Laboratory

Issue No: 042 Issue date: 01 March 2022

#### Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
TESTING OF STACK EMISSIONS TO ATMOSPHERE (cont'd)	Sampling with subsequent analysis by an ISO/IEC 17025 Accredited Laboratory (cont'd)	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and BS EN 15259:2007	
	Dioxins and Furans	BS EN 1948-1:2006 (WI/EM/959)	В
	Dioxin-like Polychlorinated Biphenyls (PCBs)	BS EN 1948-4:2010 (WI/EM/976)	В
	Polycyclic Aromatic Hydrocarbons (PAHs)	BS ISO 11338-1:2003 (WI/EM/981)	В
	Sampling and On-Site Analysis		
	Water Vapour	BS EN 14790:2017 (WI/EM/0009)	В
	Sampling and On-Line Analysis		
	Pressure, Temperature and Velocity (point velocity method) for:  Periodic Compliance Monitoring	BS EN 16911-1:2013 & EA MID 16911-1 (WI/EM/962using differential pressure device (Pitot tube) method) Procedure to meet requirements of PD CEN TR 17078:2017 Measurement Objective 1	В
	Carbon Dioxide*	PD CEN/TS 17405:2020 (WI/EM/971- NDIR analyser)	В
	Carbon Monoxide*	BS EN 15058:2017 (WI/EM/1273 - NDIR analyser)	В
	Oxides of Nitrogen (NOx)*	BS EN 14792:2017 (WI/EM/1273 – Chemiluminescence analyser)	В

Assessment Manager: JG2 Page 3 of 5



### **Schedule of Accreditation** issued by

# United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

# TUV SUD Limited Trading as TÜV SÜD National Engineering Laboratory & National Engineering Laboratory

Issue No: 042 Issue date: 01 March 2022

#### Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
TESTING OF STACK EMISSIONS TO ATMOSPHERE (cont'd)	Sampling and On-Line Analysis (cont'd)	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and BS EN 15259:2007	
	Oxygen*	BS EN 14789:2017 (WI/EM/1273 - Validated Zirconium cell analyser)	В
	Sulphur dioxide*	PD CEN/TS 17021:2017 (WI/EM/0058 – NDIR analyser)	В
	Total Gaseous Organic* Carbon (TOC / VOC) (0 to 1000 mgC/m³)	BS EN 12619: 2013 (WI/EM/965 - FID Analyser)	В
	Sampling with subsequent analysis by an ISO/IEC 17025 Accredited Laboratory	National, International and other recognised standards using documented In-House work instructions to meet the requirements of BS EN 15259:2007	
	Halides and Halogens Hydrogen chloride Hydrogen bromide Hydrogen fluoride Chlorine Bromine Total chlorides Total fluorides Total bromides	US EPA Method 26 (WI/EM/932) US EPA Method 26A (WI/EM/954)	В

Assessment Manager: JG2 Page 4 of 5



### Schedule of Accreditation issued by

## United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

# TUV SUD Limited Trading as TÜV SÜD National Engineering Laboratory & National Engineering Laboratory

Issue No: 042 Issue date: 01 March 2022

#### Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
STACK EMISSIONS - CONTINUOUS EMISSIONS MONITORING SYSTEMS (CEMS)	QAL 2 and the Annual Surveillance Test (AST) for CEMS	Documented in house procedure WI/EM/0059 to meet the requirements of BS EN 14181:2014, Environment Agency MID 14181 (TGN M20 Annex A) and other requirements of the Environment Agency (MCERTS) Performance Standard and BS EN 15259:2007	A, B

<sup>\* -</sup> The scale range of the analyser used for this test must be that detailed on its current MCERTS certificate or a range validated by the organisation to meet MCERTS requirements.

**END** 

Assessment Manager: JG2 Page 5 of 5