


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0676</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>ACS Physical Risk Control Ltd</b>	
	Issue No: 025    Issue date: 01 February 2022	
	14 Claremont Centre Durham Street Glasgow G41 1BS	Contact: Prof R J Willey Tel: +44 (0)141-427 5171 Fax: +44 (0)141-427 2722 E-Mail: Roger@acsrisk.com Website: www.acsrisk.com
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 14 Claremont Centre Durham Street Glasgow G41 1BS	<b>Local contact</b> Prof R J Willey  Tel: +44 (0)141-427 5171 Fax: +44 (0)141-427 2722 Email: roger@acs-env.com Website: www.acs-env.com	Health and Hygiene Head Office  Asbestos – All Support Functions  A
	•	

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	C



0676  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**ACS Physical Risk Control Ltd**  
**Issue No: 025 Issue date: 01 February 2022**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method No ACS/T5, 6, 7, 8, 9 based on HSG 248	A, C
	Fibre counting	Documented In-House Method, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	C
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method No ACS/T1, 22, 3, 5, 6, 9 based on HSG 248	C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method TPD8 Version 11 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			