


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0681 Accredited to ISO/IEC 17025:2017	Airtech Analysis Limited	
	Issue No: 031 Issue date: 07 January 2021	
	Unit 6 Blenheim Court Hurricane Way Wickford Essex SS11 8YT	Contact: Mr P Abrey Tel: +44 (0)1268 562 645 Fax: +44 (0)1268 570 198 E-Mail: info@airtech.org.uk Website: www.airtech.org.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 6 Blenheim Court Hurricane Way Wickford Essex SS11 8YT	Local contact Mr P Abrey Tel: +44 (0)1268 562 645 Fax: +44 (0)1268 570 198 E-Mail: info@airtech.org.uk	Asbestos A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial, and industrial	Asbestos	B
Mobile Laboratories	Asbestos	D



0681

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airtech Analysis Limited
Issue No: 031 Issue date: 07 January 2021

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248: February 2005 (Documented In-House Procedure)	A, B, D
	Sampling of air for fibre counting	HSG 248: February 2005 (Documented In-House Procedure)	B, D
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248: February 2005 (Documented In-House Procedure)	B, D
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248: February 2005 by Documented In-House Method using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method	B, D

END