


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>10147</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>University College London Hospitals (UCLH) NHS Foundation Trust</p> <p>Issue No: 005 Issue date: 25 January 2022</p>	
	<p>UCLH Environmental Research Laboratory EGA Wing Level minus 2 235 Euston Road London, Greater London NW1 2BU</p>	<p>Contact: Dr Shanom Ali Tel: +44 (0) 20 3447 6701 E-Mail: shanom.ali@nhs.net</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS</p> <p>Hydrotherapy pools, drinking water and hospital waters</p>	<p><u>Microbiology Tests</u></p> <p>Detection and Enumeration:</p> <p><i>Legionella</i> species including <i>L. pneumophila</i> serogroups 1-14</p> <p><i>Pseudomonas aeruginosa</i> (confirmed)</p>	<p>Documented In-house Methods</p> <p>T015 based on ISO 11731:2017 using membrane filtration with washing using GVPC. Identification using latex agglutination. Matrix A+B procedure 8,9 & 10 Media A & C.</p> <p>T016 based on The Microbiology of Drinking Water, Part 8, 2015</p>
<p>END</p>		