


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 10483 Accredited to ISO/IEC 17025:2017	Scopus Asbestos Compliance Ltd Issue No: 005 Issue date: 24 April 2023	
	Suite 5 Milner House Milner Way Ossett WF5 9JE United Kingdom	Contact: Mr S Annapureddy Tel: +44 (0) 7493 330345 E-Mail: sai@scopus-asbestos.co.uk Website: www.scopus-asbestos.co.uk
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Suite 5 Milner House Milner Way Ossett WF5 9JE United Kingdom	Local contact Mr S Annapureddy Tel: +44 (0) 7493 330345 E-Mail: sai@scopus-asbestos.co.uk Website: www.scopus-asbestos.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories – Asbestos Fibres in Air	Health and Hygiene	C



10483

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Scopus Asbestos Compliance Ltd

Issue No: 005 **Issue date:** 24 April 2023

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SAC-SOP-15 based on HSG 248	A, B, C
	Fibre counting	Documented In-House Method SAC-SOP-15, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method SAC-SOP-15, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method SAC-SOP-07 based on HSG 248	
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method SAC-SOP-07 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			