


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>10547</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Medipure Limited</p> <p>Issue No: 002 Issue date: 08 April 2021</p>	
	<p>Erw Las Colomendy Industrial Estate Denbigh LL16 5TA United Kingdom</p>	<p>Contact: James Daly Tel: +44 (0) 1745 814518 E-Mail: enquiries@medipure.co.uk Website: www.medipure.co.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS -</p> <p>Washer-disinfectant final rinse water (mains fed, steam condensate and reverse osmosis)</p>	<p><u>Microbiological Testing</u></p>	<p>Documented In -House methods</p>
	<p>Environmental Mycobacteria</p> <p><i>Pseudomonas aeruginosa</i></p>	<p>WI 038 & WI 040 (confirmation by acid fast staining) to meet the requirements of HTM 01-06 Part E 2016, WHTM01-06 and NICFPP</p> <p>WI 020 using membrane filtration and Pseud CN agar with incubation at 35°C) to meet the requirements of HTM 01-06 Part E 2016, WHTM01-06 and NICFPP</p>
	<p>Total Aerobic Colony Count</p>	<p>WI 020 using membrane filtration and Tryptone Soya Agar with incubation at 30°C) to meet the requirements of HTM 01-06 Part E 2016, WHTM01-06 and NICFPP</p>
<p>ENDOSCOPE SURROGATE DEVICES</p>	<p>Evaluation of Bacterial Contamination levels as a measure of disinfection efficacy</p>	<p>WI 004 & 005 assessment of contamination levels of <i>Pseudomonas aeruginosa</i>, <i>Staphylococcus aureus</i>, <i>Mycobacterium terrae</i> and <i>Candida albicans</i> levels on PTFE tubing before and after a Washer Disinfectant Cycle at customer premises, based on BS EN ISO 15883-4:2018 6.6.2 and Annex B, plus HTM 01-06 Part E 2016</p>
<p>END</p>		