


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 10662 Accredited to ISO/IEC 17025: 2017	TRAC Laboratories Limited Issue No: 006 Issue date: 10 May 2023	
	Suite 5, The Pavilions Bridge Hall Lane Bury BL9 7NX	Contact: Phil Cooke Tel: +44 (0)1616 418150 E-Mail: phil.cooke@trac-associates.com Website: www.trac-associates.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Suite 5, The Pavilions Bridge Hall Lane Bury BL9 7NX	Local contact Mr Phil Cooke	Health and Hygiene Asbestos	A

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Premises including domestic, commercial and industrial	Contact: Mr Phil Cooke	Health and Hygiene Asbestos	B
Mobile Laboratories	Contact: Mr Phil Cooke	Health and Hygiene Asbestos	C



10662
Accredited to
ISO/IEC 17025: 2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

TRAC Laboratories Limited
Issue No: 006 Issue date: 10 May 2023

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	A
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248: 2021 by Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining	
	Fibre counting	HSG 248: 2021 by Documented In-House Procedure using phase contrast microscopy (PCM)	
	Sampling of air for fibre counting	HSG248: 2021 (Documented In-House Procedure)	
ASBESTOS FIBRES IN AIR	4 Stage Clearance Process	HSG248: 2021 (Documented In-House Procedure) Membrane Filter Method using Phase Contrast Microscopy (PCM)	B
END			