


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING</p> <p>10740</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Fluoro-Tech Ltd</p> <p>Issue No: 006 Issue date: 21 February 2022</p>	
	<p>Unit 3 Webnor Industrial Estate Ettingshall Wolverhampton WV2 2LD</p>	<p>Contact: Mr Ian Evans Tel: +44 (0)1902 219 899 E-Mail: Ian.Evans@fluoro-tech.co.uk Website: www.fluoro-tech.co.uk</p>
<p>Testing performed at the above address only</p>		

Flexible Scope

Fluoro-Tech Ltd, is accredited for a limited flexible scope that enables them to conduct accredited testing through the modification of existing test methods and include newly revised and technically equivalent methods to activities detailed below, in accordance with their documented in-house procedure ID-FSCP-01 Information about flexible scopes of accreditation is available in UKAS document GEN 4

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>LIGHT and DENSE METALS and ALLOYS</p> <p>Part Machined and Fully Machined Fasteners (Bolts, Nuts & Washers)</p> <p>Part Machined or Fully Machined Fasteners (Bolts, Nuts & Washers)</p> <p>Wrought Bar, Billet, Forgings & Part Machined Fasteners</p>	<p><u>Non-Destructive Test</u></p> <p>Magnetic Particle: Fluorescent up to 2000 A.</p> <p>Dye Penetrant: Fluorescent Penetrant Water Washable.</p> <p>Ultrasonic Testing: Manual Contact Method.</p>	<p>ASTM E1444/E1444M ASTM E709 BS EN ISO 9934-1 BS EN 10228-1 DEF STAN 02-729 Part 2 API 6A</p> <p>ASTM E1417/E1417M ASTM E165/E165M BS EN ISO 3452-1 BS EN 10228-2 DEF-STAN 02-729 Part 4 API 6A</p> <p>ASTM A388/A388M ASTM A962/A962M BS EN 10228-3 BS EN 10228-4 DEF-STAN 02-729 Part 5 BS EN 10160 API 6A ASME V</p>



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Fluoro-Tech Ltd

Issue No: 006 **Issue date:** 21 February 2022

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
METALS, ALLOYS AND METAL PRODUCTS	<u>Mechanical Testing</u> Rockwell Hardness (HRBW and HRC)	ASTM E18
END		