


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<h3>SOCOTEC Asbestos Limited</h3> <p>Issue No: 064 Issue date: 28 March 2018</p>	
	<p>Bretby Business Park PO Box 100 Ashby Road Bretby Burton-upon-Trent Staffordshire DE15 0YZ</p>	<p>Contact: Mr Matthew Harris Tel: +44 (0) 1283 554499 Fax: +44 (0) 1283 554401 E-Mail: matthew.harris@socotec.com Website: www.socotec.co.uk</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Bretby Business Park PO Box 100 Ashby Road Bretby Burton-upon-Trent Staffordshire DE15 0YZ</p> <p>Local contact Mr Andrew Elsby Tel: +44 (0)1283 554499 Fax: +44 (0)1283 554401</p>	Health and Hygiene	A
<p>Address Askern Road Carcroft Doncaster DN6 8DG</p> <p>Local contact Mr Andrew Elsby Tel: +44 (0)1302 724455 Fax: +44 (0)1302 725240</p>	Health and Hygiene	C
<p>Address 2-8 Langlands Place Calvin South Business Park East Kilbride Glasgow G75 0YF</p> <p>Local contact Mr Kevin Duffy Tel: +44 (0)1355 225488 Fax: +44 (0)1355 249669</p>	Health and Hygiene	G
<p>Address Units C and D 2 Wilkinson Road Bankside Trade Park Love Lane Cirencester Gloucestershire GL7 1YT</p> <p>Local contact Mr David Birkin Tel: +44 (0)1285 700593</p>	Health and Hygiene	I



1089

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SOCOTEC Asbestos Limited
Issue No: 064 Issue date: 28 March 2018

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
Address Unit 20 Falcon Business Centre Ashton Road Romford RM3 8UR United Kingdom	Local contact Mr Daniel Fisher Tel: +44 (0) 1708 330760 Fax: +44 (0) 1708 330768	Health and Hygiene L
Address Unit 5 Bridgend Industrial Estate New Street Bridgend CF31 3UD	Local contact Mr David Birkin Tel: +44 (0)1594 546343 Fax: +44 (0)1594 546342	Health and Hygiene N



1089

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SOCOTEC Asbestos Limited
Issue No: 064 Issue date: 28 March 2018

Testing performed by the Organisation at the locations specified

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including: Commercial Industrial Domestic Burton upon Trent	Health and Hygiene	B
Premises including: Commercial Industrial Domestic East Kilbride	Health and Hygiene	H
Premises including: Commercial Industrial Domestic Cirencester	Health and Hygiene	J
Premises including: Commercial Industrial Domestic Romford	Health and Hygiene	K
Premises including: Commercial Industrial Domestic Bridgend	Health and Hygiene	M
Mobile Laboratories: Fibre Counting	Health and Hygiene	O



1089

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SOCOTEC Asbestos Limited

Issue No: 064 Issue date: 28 March 2018

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248:February 2005) Flexible scope - to establish temporary site laboratories to conduct the following asbestos testing activities: Fibre counting and Sampling of air for fibre counting for Asbestos fibres in Air. The following Site laboratories (denoted *) are set up in accordance with the Documented In-House Procedure: SCI – ASB-030	
	Fibre counting	HSG 248:February 2005 Documented In-House Method SCI/ASB/006 - Procedure for Airborne Fibre Counting	A,B,C,G, H, I, J,K, L, M, N, O
	Sampling of air for fibre counting	HSG 248:February 2005 Documented In-House Method SCI/ASB/005 - Procedure for Airborne Fibre Sampling	A,B,G,H, I, J,K, L, M, N, O
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance	HSG 248:February 2005 Documented In-House Method SCI/ASB/004 - Procedure for Four Stage Clearance	B, H, J, K, M
	Sampling of material for subsequent identification of asbestos	HSG 248:February 2005 Documented In-House Method SCI/ASB/001 - Procedure for Asbestos Surveying	B, H, J, K, M



1089

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SOCOTEC Asbestos Limited

Issue No: 064 Issue date: 28 March 2018

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos (cont'd)	<p><u>Health and Hygiene</u> (cont'd)</p> <p>Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248:February 2005)</p> <p>HSG 248:February 2005 by Documented In-House Method SCI/ASB/007 - Procedure for Bulk Identification</p>	A, C, G, I, L, N
ASBESTOS IN SOILS (fibre screening and identification)	<p>Asbestos Fibre Screening and Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Documented In-House Method SCI/ASB/020 using stereo-microscopy, polarised light optical microscopy and dispersion staining</p>	A
ASBESTOS IN SOILS/SEDIMENTS/ AGGREGATES (screening/identification/ quantification)	<p>Asbestos Fibre Screening/ Identification & Quantification of asbestos content of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Documented In-House Method SCI/ASB/020 using stereo-microscopy, polarised light optical microscopy and dispersion staining</p>	A, N
END			